

Metals
- 4 -
INTERFERENCE CHECK SAMPLE

Client: JACOBS Engineering Group, Inc. **SDG No.:** Q1478
Contract: JACO05 **Lab Code:** CHEM **Case No.:** Q1478 **SAS No.:** Q1478
ICS Source: EPA **Instrument ID:** P4

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Low Limit (ug/L)	High Limit (ug/L)	Analysis Date	Analysis Time	Run Number
ICSA01	Arsenic	3.58			-20	20	03/10/2025	12:05	LB134965
	Barium	1.50	6.0	25	-94	106	03/10/2025	12:05	LB134965
	Cadmium	-4.00	1.0	400	-5	7	03/10/2025	12:05	LB134965
	Chromium	59.3	52.0	114	42	62	03/10/2025	12:05	LB134965
	Lead	6.57			-12	12	03/10/2025	12:05	LB134965
	Selenium	-12.1			-20	20	03/10/2025	12:05	LB134965
	Silver	-0.43			-10	10	03/10/2025	12:05	LB134965
ICSAB01	Arsenic	110	104	106	88.4	120	03/10/2025	12:10	LB134965
	Barium	483	537	90	437	637	03/10/2025	12:10	LB134965
	Cadmium	1020	972	105	826	1120	03/10/2025	12:10	LB134965
	Chromium	581	542	107	460	624	03/10/2025	12:10	LB134965
	Lead	56.3	49.0	115	37	61	03/10/2025	12:10	LB134965
	Selenium	34.1	46.0	74	26	66	03/10/2025	12:10	LB134965
	Silver	221	201	110	170	232	03/10/2025	12:10	LB134965