

Metals
- 5b -
POST DIGEST SPIKE SUMMARY

Client: JACOBS Engineering Group, Inc. **SDG No.:** Q1478
Contract: JACO05 **Lab Code:** CHEM **Case No.:** Q1478 **SAS No.:** Q1478
Matrix: Water **Level:** LOW **Client ID:** TP-4A
Sample ID: Q1489-16 **Spiked ID:** Q1489-16A

Analyte	Units	Acceptance Limit %R	Spiked Result	C	Sample Result	C	Spike Added	% Recovery	Qual	M
Barium	ug/L	75 - 125	4850		3490		1000	136		P