

Metals

- 6 -

DUPLICATE SAMPLE SUMMARY

Client: JACOBS Engineering Group, Inc. **Level:** LOW **SDG No.:** Q1478
Contract: JACO05 **Lab Code:** CHEM **Case No.:** Q1478 **SAS No.:** Q1478
Matrix: Water **Sample ID:** Q1489-16 **Client ID:** TP-4DUP
Percent Solids for Sample: NA **Duplicate ID** Q1489-16DUP **Percent Solids for Spike Sample:** NA

Analyte	Units	Acceptance Limit	Sample Result	C	Duplicate Result	C	RPD	Qual	M
Arsenic	ug/L	20	100	U	100	U			P
Barium	ug/L	20	3490		3480		0		P
Cadmium	ug/L	20	30.0	U	30.0	U			P
Chromium	ug/L	20	50.0	U	50.0	U			P
Lead	ug/L	20	60.2		59.0	J	2		P
Mercury	ug/L	20	2.00	U	2.00	U			CV
Selenium	ug/L	20	100	U	100	U			P
Silver	ug/L	20	50.0	U	50.0	U			P

Metals

- 6 -

DUPLICATE SAMPLE SUMMARY

Client: JACOBS Engineering Group, Inc. **Level:** LOW **SDG No.:** Q1478
Contract: JACO05 **Lab Code:** CHEM **Case No.:** Q1478 **SAS No.:** Q1478
Matrix: Water **Sample ID:** Q1489-16MS **Client ID:** TP-4MSD
Percent Solids for Sample: NA **Duplicate ID** Q1489-16MSD **Percent Solids for Spike Sample:** NA

Analyte	Units	Acceptance Limit	Sample Result	C	Duplicate Result	C	RPD	Qual	M
Arsenic	ug/L	20	4420		4340		2		P
Barium	ug/L	20	4720		4790		1		P
Cadmium	ug/L	20	1010		1020		1		P
Chromium	ug/L	20	2120		2130		0		P
Lead	ug/L	20	4990		5010		0		P
Mercury	ug/L	20	41.0		38.1		7		CV
Selenium	ug/L	20	10600		10300		3		P
Silver	ug/L	20	395		390		1		P