

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: JACOBS Engineering Group, Inc.

SDG No.: Q1478

Contract: JACO05

Lab Code: CHEM

Case No.: Q1478

SAS No.: Q1478

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB166990BS							
Arsenic	ug/L	4000	4050		101	80 - 120	P
Barium	ug/L	1000	883		88	80 - 120	P
Cadmium	ug/L	1000	944		94	80 - 120	P
Chromium	ug/L	2000	1970		98	80 - 120	P
Lead	ug/L	5000	4770		95	80 - 120	P
Selenium	ug/L	10000	10100		101	80 - 120	P
Silver	ug/L	380	372		98	80 - 120	P

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: JACOBS Engineering Group, Inc.

SDG No.: Q1478

Contract: JACO05

Lab Code: CHEM

Case No.: Q1478

SAS No.: Q1478

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB167014BS Mercury	ug/L	4.0	4.19		105	80 - 120	CV