

Instrument ID: FID_B

Daily Analysis Runlog For Sequence/QC Batch ID # FB021125

Review By	yogesh	Review On	2/11/2025 1:08:25 PM
Supervise By	Ankita	Supervise On	2/12/2025 10:47:12 AM
SubDirectory	FB021125	HP Acquire Method	HP Processing Method FB021125
STD. NAME	STD REF.#		
Tune/Reschk Initial Calibration Stds	PP24110,PP24188,PP24189,PP24190,PP24191		
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	PP24111,PP24192		

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	5 GRO STD		FB031496.D	11 Feb 2025 9:32		YP/AJ	Ok
2	10 GRO STD		FB031497.D	11 Feb 2025 10:03		YP/AJ	Ok
3	20 GRO STD		FB031498.D	11 Feb 2025 11:32		YP/AJ	Ok
4	50 GRO STD		FB031499.D	11 Feb 2025 12:03		YP/AJ	Ok,M
5	100 GRO STD		FB031500.D	11 Feb 2025 12:34		YP/AJ	Ok,M
6	FB021125GROICV		FB031501.D	11 Feb 2025 13:19		YP/AJ	Ok

M : Manual Integration

Instrument ID: FID_B

Daily Analysis Runlog For Sequence/QC Batch ID # FB030325

Review By	yogesh	Review On	3/3/2025 1:53:44 PM
Supervise By	Ankita	Supervise On	3/4/2025 3:14:07 PM
SubDirectory	FB030325	HP Acquire Method	HP Processing Method FB021125

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds	PP24110,PP24188,PP24189,PP24190,PP24191
CCC Internal Standard/PEM	PP24211,PP24212,PP24213
ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	PP24111,PP24192

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	20 PPB GRO STD		FB031524.D	3 Mar 2025 10:24		YP/AJ	Ok
2	VBF0303S1		FB031525.D	3 Mar 2025 11:06		YP/AJ	Ok
3	VBF0303S2		FB031526.D	3 Mar 2025 11:34		YP/AJ	Ok
4	BSF0303S1		FB031527.D	3 Mar 2025 12:02		YP/AJ	Ok
5	Q1474-01		FB031528.D	3 Mar 2025 14:23	Vial- A, Not Purged	YP/AJ	Not Ok
6	Q1474-01		FB031529.D	3 Mar 2025 14:51	Vial- B, Not Purged	YP/AJ	Not Ok
7	Q1474-01		FB031530.D	3 Mar 2025 15:19	Vial- C	YP/AJ	Ok
8	Q1474-01MS		FB031531.D	3 Mar 2025 15:47		YP/AJ	Ok,M
9	Q1474-01MSD		FB031532.D	3 Mar 2025 16:15		YP/AJ	Ok,M
10	BSF0303S2		FB031533.D	3 Mar 2025 16:43		YP/AJ	Ok
11	20 PPB GRO STD		FB031534.D	3 Mar 2025 17:38		YP/AJ	Ok
12	Q1478-16		FB031535.D	3 Mar 2025 18:06	Vial- A, Not Purged	YP/AJ	Not Ok
13	Q1478-16		FB031536.D	3 Mar 2025 18:34	Vial- B ,Not Purged	YP/AJ	Not Ok
14	Q1478-16		FB031537.D	3 Mar 2025 19:02	Vial- C	YP/AJ	Ok
15	BSF0303S3		FB031538.D	3 Mar 2025 19:29		YP/AJ	Ok
16	20 PPB GRO STD		FB031539.D	3 Mar 2025 19:57		YP/AJ	Ok

M : Manual Integration

Instrument ID: FID_B

Daily Analysis Runlog For Sequence/QC Batch ID # FB030425

Review By	yogesh	Review On	3/4/2025 11:58:45 AM	
Supervise By	Ankita	Supervise On	3/5/2025 9:18:23 AM	
SubDirectory	FB030425	HP Acquire Method	HP Processing Method	FB021125
STD. NAME		STD REF.#		
Tune/Reschk Initial Calibration Stds		PP24110,PP24188,PP24189,PP24190,PP24191		
CCC		PP24214,PP24215		
Internal Standard/PEM				
ICV/I.BLK		PP24111,PP24192		
Surrogate Standard				
MS/MSD Standard				
LCS Standard				

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	20 PPB GRO STD		FB031540.D	4 Mar 2025 9:17		YP/AJ	Ok
2	VBF0304W1		FB031541.D	4 Mar 2025 11:14		YP/AJ	Ok
3	BSF0304W1		FB031542.D	4 Mar 2025 11:42		YP/AJ	Ok
4	BSF0304W2		FB031543.D	4 Mar 2025 12:10		YP/AJ	Ok
5	Q1478-02		FB031544.D	4 Mar 2025 12:38	Surr Fail, need 10x dilution	YP/AJ	Dilution
6	Q1478-04		FB031545.D	4 Mar 2025 13:06	Surr Fail	YP/AJ	ReRun
7	Q1478-06		FB031546.D	4 Mar 2025 13:34	Surr Fail	YP/AJ	ReRun
8	Q1478-08		FB031547.D	4 Mar 2025 14:02	Surr Fail	YP/AJ	ReRun
9	BSF0304W3		FB031548.D	4 Mar 2025 14:29		YP/AJ	Ok
10	20 PPB GRO STD		FB031549.D	4 Mar 2025 14:57		YP/AJ	Ok
11	Q1478-04RE		FB031550.D	4 Mar 2025 16:07	Surr Fail	YP/AJ	Confirms
12	Q1478-06RE		FB031551.D	4 Mar 2025 16:35	Surr Fail	YP/AJ	Confirms
13	Q1478-08RE		FB031552.D	4 Mar 2025 17:03	Surr Fail	YP/AJ	Confirms
14	Q1478-02		FB031553.D	4 Mar 2025 17:30	Surr Fail	YP/AJ	Ok
15	BSF0304W4		FB031554.D	4 Mar 2025 17:59		YP/AJ	Ok
16	20 PPB GRO STD		FB031555.D	4 Mar 2025 18:27		YP/AJ	Ok

M : Manual Integration