

WATER GASOLINE RANGE ORGANICS LABORATORY CONTROL SPIKE/LABORATORY CONTROL SPIKE DUPLICATION

Lab Name: Chemtech **Client:** JACOBS Engineering Group, Inc.
Lab Code: CHEM **Cas No:** Q1478 **SAS No :** Q1478 **SDG No:** Q1478
Matrix Spike - EPA Sample No : BSF0304W2 **Datafile:** FB031543.D

COMPOUND	SPIKE ADDED ug/L	CONCENTRATION ug/L	LCS/LCSD CONCENTRATION ug/L	% REC	QC LIMITS
GRO	180	0	176	98	50-150

LCS/LCSD % Recovery RPD : 0