

CALIBRATION VERIFICATION SUMMARY

Contract: JACO05

Lab Code: CHEM Case No.: Q1478 SAS No.: Q1478 SDG NO.: Q1478

Continuing Calib Date: 03/04/2025 Initial Calibration Date(s): 02/24/2025 02/24/2025

Continuing Calib Time: 22:59 Initial Calibration Time(s): 14:59 22:17

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Aroclor-1254-1	(1)	6.53	6.53	6.43	6.63	0.00
Aroclor-1254-2	(2)	6.74	6.75	6.65	6.85	0.01
Aroclor-1254-3	(3)	7.11	7.11	7.01	7.21	0.00
Aroclor-1254-4	(4)	7.39	7.39	7.29	7.49	0.00
Aroclor-1254-5	(5)	7.80	7.81	7.71	7.91	0.01
Tetrachloro-m-xylene		4.53	4.53	4.43	4.63	0.00
Decachlorobiphenyl		10.25	10.26	10.16	10.36	0.01

CALIBRATION VERIFICATION SUMMARY

Contract: JACO05

Lab Code: CHEM Case No.: Q1478 SAS No.: Q1478 SDG NO.: Q1478

Continuing Calib Date: 03/04/2025 Initial Calibration Date(s): 02/24/2025 02/24/2025

Continuing Calib Time: 22:59 Initial Calibration Time(s): 14:59 22:17

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Aroclor-1254-1	(1)	5.72	5.73	5.63	5.83	0.01
Aroclor-1254-2	(2)	5.87	5.88	5.78	5.98	0.01
Aroclor-1254-3	(3)	6.28	6.28	6.18	6.38	0.00
Aroclor-1254-4	(4)	6.50	6.51	6.41	6.61	0.01
Aroclor-1254-5	(5)	6.92	6.93	6.83	7.03	0.01
Tetrachloro-m-xylene		3.83	3.83	3.73	3.93	0.00
Decachlorobiphenyl		8.88	8.89	8.79	8.99	0.01