

CALIBRATION VERIFICATION SUMMARY

Contract: JACO05

Lab Code: CHEM Case No.: Q1478 SAS No.: Q1478 SDG NO.: Q1478

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 02/24/2025 02/24/2025

Client Sample No.: CCAL01 Date Analyzed: 03/03/2025

Lab Sample No.: AR1660CCC500 Data File : PP070186.D Time Analyzed: 16:32

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.676	5.581	5.781	503.890	500.000	0.8
Aroclor-1016-2	5.698	5.602	5.802	528.440	500.000	5.7
Aroclor-1016-3	5.760	5.665	5.865	512.230	500.000	2.4
Aroclor-1016-4	5.858	5.762	5.962	527.880	500.000	5.6
Aroclor-1016-5	6.150	6.056	6.256	520.270	500.000	4.1
Aroclor-1260-1	7.270	7.175	7.375	526.060	500.000	5.2
Aroclor-1260-2	7.524	7.429	7.629	506.890	500.000	1.4
Aroclor-1260-3	7.882	7.788	7.988	531.650	500.000	6.3
Aroclor-1260-4	8.106	8.012	8.212	515.390	500.000	3.1
Aroclor-1260-5	8.427	8.333	8.533	525.080	500.000	5.0
Decachlorobiphenyl	10.248	10.156	10.356	46.720	50.000	-6.6
Tetrachloro-m-xylene	4.524	4.427	4.627	52.130	50.000	4.3

CALIBRATION VERIFICATION SUMMARY

Contract: JACO05

Lab Code: CHEM Case No.: Q1478 SAS No.: Q1478 SDG NO.: Q1478

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 02/24/2025 02/24/2025

Client Sample No.: CCAL01 Date Analyzed: 03/03/2025

Lab Sample No.: AR1660CCC500 Data File : PP070186.D Time Analyzed: 16:32

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.916	4.820	5.020	508.150	500.000	1.6
Aroclor-1016-2	4.935	4.839	5.039	505.540	500.000	1.1
Aroclor-1016-3	5.113	5.017	5.217	523.260	500.000	4.7
Aroclor-1016-4	5.154	5.058	5.258	520.820	500.000	4.2
Aroclor-1016-5	5.369	5.273	5.473	533.230	500.000	6.6
Aroclor-1260-1	6.407	6.312	6.512	491.760	500.000	-1.6
Aroclor-1260-2	6.594	6.500	6.700	494.530	500.000	-1.1
Aroclor-1260-3	6.749	6.654	6.854	463.410	500.000	-7.3
Aroclor-1260-4	7.220	7.126	7.326	530.120	500.000	6.0
Aroclor-1260-5	7.461	7.367	7.567	553.030	500.000	10.6
Decachlorobiphenyl	8.880	8.788	8.988	44.440	50.000	-11.1
Tetrachloro-m-xylene	3.828	3.730	3.930	48.550	50.000	-2.9