

CALIBRATION VERIFICATION SUMMARY

Contract: JACO05

Lab Code: CHEM Case No.: Q1478 SAS No.: Q1478 SDG NO.: Q1478

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 02/24/2025 02/24/2025

Client Sample No.: CCAL02 Date Analyzed: 03/03/2025

Lab Sample No.: AR1660CCC500 Data File : PP070201.D Time Analyzed: 21:08

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.675	5.581	5.781	475.920	500.000	-4.8
Aroclor-1016-2	5.697	5.602	5.802	496.910	500.000	-0.6
Aroclor-1016-3	5.759	5.665	5.865	484.100	500.000	-3.2
Aroclor-1016-4	5.856	5.762	5.962	492.010	500.000	-1.6
Aroclor-1016-5	6.149	6.056	6.256	472.290	500.000	-5.5
Aroclor-1260-1	7.268	7.175	7.375	485.460	500.000	-2.9
Aroclor-1260-2	7.523	7.429	7.629	455.890	500.000	-8.8
Aroclor-1260-3	7.881	7.788	7.988	480.850	500.000	-3.8
Aroclor-1260-4	8.105	8.012	8.212	474.730	500.000	-5.1
Aroclor-1260-5	8.425	8.333	8.533	486.790	500.000	-2.6
Decachlorobiphenyl	10.246	10.156	10.356	43.810	50.000	-12.4
Tetrachloro-m-xylene	4.523	4.427	4.627	51.210	50.000	2.4

CALIBRATION VERIFICATION SUMMARY

Contract: JACO05

Lab Code: CHEM Case No.: Q1478 SAS No.: Q1478 SDG NO.: Q1478

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 02/24/2025 02/24/2025

Client Sample No.: CCAL02 Date Analyzed: 03/03/2025

Lab Sample No.: AR1660CCC500 Data File : PP070201.D Time Analyzed: 21:08

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	4.916	4.820	5.020	488.870	500.000	-2.2
Aroclor-1016-2	4.935	4.839	5.039	492.390	500.000	-1.5
Aroclor-1016-3	5.112	5.017	5.217	499.400	500.000	-0.1
Aroclor-1016-4	5.154	5.058	5.258	498.070	500.000	-0.4
Aroclor-1016-5	5.369	5.273	5.473	542.900	500.000	8.6
Aroclor-1260-1	6.406	6.312	6.512	457.960	500.000	-8.4
Aroclor-1260-2	6.594	6.500	6.700	459.580	500.000	-8.1
Aroclor-1260-3	6.748	6.654	6.854	435.380	500.000	-12.9
Aroclor-1260-4	7.220	7.126	7.326	479.080	500.000	-4.2
Aroclor-1260-5	7.461	7.367	7.567	490.430	500.000	-1.9
Decachlorobiphenyl	8.880	8.788	8.988	41.410	50.000	-17.2
Tetrachloro-m-xylene	3.828	3.730	3.930	47.150	50.000	-5.7