

CALIBRATION VERIFICATION SUMMARY

Contract: JACO05

Lab Code: CHEM Case No.: Q1478 SAS No.: Q1478 SDG NO.: Q1478

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 02/24/2025 02/24/2025

Client Sample No.: CCAL03 Date Analyzed: 03/04/2025

Lab Sample No.: AR1660CCC500 Data File : PP070222.D Time Analyzed: 17:34

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	5.679	5.581	5.781	510.200	500.000	2.0
Aroclor-1016-2	5.701	5.602	5.802	525.990	500.000	5.2
Aroclor-1016-3	5.763	5.665	5.865	521.070	500.000	4.2
Aroclor-1016-4	5.861	5.762	5.962	523.320	500.000	4.7
Aroclor-1016-5	6.154	6.056	6.256	515.090	500.000	3.0
Aroclor-1260-1	7.273	7.175	7.375	522.590	500.000	4.5
Aroclor-1260-2	7.526	7.429	7.629	498.280	500.000	-0.3
Aroclor-1260-3	7.885	7.788	7.988	517.510	500.000	3.5
Aroclor-1260-4	8.109	8.012	8.212	496.430	500.000	-0.7
Aroclor-1260-5	8.430	8.333	8.533	512.380	500.000	2.5
Decachlorobiphenyl	10.250	10.156	10.356	46.320	50.000	-7.4
Tetrachloro-m-xylene	4.528	4.427	4.627	53.810	50.000	7.6

CALIBRATION VERIFICATION SUMMARY

Contract: JACO05

Lab Code: CHEM Case No.: Q1478 SAS No.: Q1478 SDG NO.: Q1478

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 02/24/2025 02/24/2025

Client Sample No.: CCAL03 Date Analyzed: 03/04/2025

Lab Sample No.: AR1660CCC500 Data File : PP070222.D Time Analyzed: 17:34

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.917	4.820	5.020	528.540	500.000	5.7
Aroclor-1016-2	4.936	4.839	5.039	529.730	500.000	5.9
Aroclor-1016-3	5.113	5.017	5.217	547.360	500.000	9.5
Aroclor-1016-4	5.155	5.058	5.258	534.090	500.000	6.8
Aroclor-1016-5	5.370	5.273	5.473	573.290	500.000	14.7
Aroclor-1260-1	6.407	6.312	6.512	519.270	500.000	3.9
Aroclor-1260-2	6.595	6.500	6.700	506.910	500.000	1.4
Aroclor-1260-3	6.749	6.654	6.854	475.410	500.000	-4.9
Aroclor-1260-4	7.222	7.126	7.326	506.720	500.000	1.3
Aroclor-1260-5	7.463	7.367	7.567	511.490	500.000	2.3
Decachlorobiphenyl	8.882	8.788	8.988	45.920	50.000	-8.2
Tetrachloro-m-xylene	3.829	3.730	3.930	50.000	50.000	0.0