

CALIBRATION VERIFICATION SUMMARY

Contract: JACO05

Lab Code: CHEM Case No.: Q1478 SAS No.: Q1478 SDG NO.: Q1478

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 02/24/2025 02/24/2025

Client Sample No.: CCAL04 Date Analyzed: 03/04/2025

Lab Sample No.: AR1660CCC500 Data File : PP070237.D Time Analyzed: 22:10

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.679	5.581	5.781	520.730	500.000	4.1
Aroclor-1016-2	5.700	5.602	5.802	536.140	500.000	7.2
Aroclor-1016-3	5.763	5.665	5.865	535.540	500.000	7.1
Aroclor-1016-4	5.860	5.762	5.962	542.220	500.000	8.4
Aroclor-1016-5	6.154	6.056	6.256	541.420	500.000	8.3
Aroclor-1260-1	7.272	7.175	7.375	552.650	500.000	10.5
Aroclor-1260-2	7.526	7.429	7.629	517.020	500.000	3.4
Aroclor-1260-3	7.885	7.788	7.988	538.450	500.000	7.7
Aroclor-1260-4	8.108	8.012	8.212	522.710	500.000	4.5
Aroclor-1260-5	8.429	8.333	8.533	530.240	500.000	6.0
Decachlorobiphenyl	10.250	10.156	10.356	48.080	50.000	-3.8
Tetrachloro-m-xylene	4.526	4.427	4.627	53.330	50.000	6.7

CALIBRATION VERIFICATION SUMMARY

Contract: JACO05

Lab Code: CHEM Case No.: Q1478 SAS No.: Q1478 SDG NO.: Q1478

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 02/24/2025 02/24/2025

Client Sample No.: CCAL04 Date Analyzed: 03/04/2025

Lab Sample No.: AR1660CCC500 Data File : PP070237.D Time Analyzed: 22:10

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.916	4.820	5.020	521.620	500.000	4.3
Aroclor-1016-2	4.935	4.839	5.039	524.820	500.000	5.0
Aroclor-1016-3	5.112	5.017	5.217	546.690	500.000	9.3
Aroclor-1016-4	5.154	5.058	5.258	530.420	500.000	6.1
Aroclor-1016-5	5.369	5.273	5.473	579.710	500.000	15.9
Aroclor-1260-1	6.406	6.312	6.512	515.410	500.000	3.1
Aroclor-1260-2	6.594	6.500	6.700	525.440	500.000	5.1
Aroclor-1260-3	6.748	6.654	6.854	488.910	500.000	-2.2
Aroclor-1260-4	7.220	7.126	7.326	532.410	500.000	6.5
Aroclor-1260-5	7.461	7.367	7.567	563.370	500.000	12.7
Decachlorobiphenyl	8.880	8.788	8.988	49.000	50.000	-2.0
Tetrachloro-m-xylene	3.828	3.730	3.930	50.860	50.000	1.7