

Instrument ID: ECD_P

Daily Analysis Runlog For Sequence/QC Batch ID # PP022425

Review By	yogesh	Review On	2/25/2025 9:00:36 AM
Supervise By	Ankita	Supervise On	2/25/2025 9:48:29 AM
SubDirectory	PP022425	HP Acquire Method	HP Processing Method PP022425

STD. NAME	STD REF.#
Tune/Reschk	
Initial Calibration Stds	PP23735,PP23736,PP23737,PP23738,PP23739,PP23740,PP23741,PP23742,PP23743,PP23744,PP23745,PP23747,PP23748,PP23749,PP23750,PP23751,PP23752,PP23753,PP23754,PP23755,PP23756,PP23757,PP23758,PP23759,PP23760,PP23761,PP23762,PP23763,PP23764,PP23765,PP23766,PP23767,PP23768,PP23769,PP23770,PP23771,PP23772,PP23773,PP23774,PP23775
CCC	PP23737,PP23742,PP23749,PP23754,PP23758,PP23763,PP23768,PP23773
Internal Standard/PEM	
ICV/I.BLK	PP23778,PP23780,PP23783,PP23784,PP23786,PP23788,PP23790,PP23947
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleId	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	HEXANE	HEXANE	PP069994.D	24 Feb 2025 14:27		YP\AJ	Ok
2	I.BLK	I.BLK	PP069995.D	24 Feb 2025 14:43		YP\AJ	Ok
3	AR1660ICC1000	AR1660ICC1000	PP069996.D	24 Feb 2025 14:59		YP\AJ	Ok
4	AR1660ICC750	AR1660ICC750	PP069997.D	24 Feb 2025 15:15		YP\AJ	Ok
5	AR1660ICC500	AR1660ICC500	PP069998.D	24 Feb 2025 15:32		YP\AJ	Ok
6	AR1660ICC250	AR1660ICC250	PP069999.D	24 Feb 2025 15:48		YP\AJ	Ok
7	AR1660ICC050	AR1660ICC050	PP070000.D	24 Feb 2025 16:04		YP\AJ	Ok,M
8	AR1221ICC500	AR1221ICC500	PP070001.D	24 Feb 2025 16:20		YP\AJ	Ok,M
9	AR1232ICC500	AR1232ICC500	PP070002.D	24 Feb 2025 16:37		YP\AJ	Ok
10	AR1242ICC1000	AR1242ICC1000	PP070003.D	24 Feb 2025 16:53		YP\AJ	Ok
11	AR1242ICC750	AR1242ICC750	PP070004.D	24 Feb 2025 17:09		YP\AJ	Ok,M
12	AR1242ICC500	AR1242ICC500	PP070005.D	24 Feb 2025 17:25		YP\AJ	Ok
13	AR1242ICC250	AR1242ICC250	PP070006.D	24 Feb 2025 17:42		YP\AJ	Ok,M
14	AR1242ICC050	AR1242ICC050	PP070007.D	24 Feb 2025 17:58		YP\AJ	Ok,M
15	AR1248ICC1000	AR1248ICC1000	PP070008.D	24 Feb 2025 18:14		YP\AJ	Ok
16	AR1248ICC750	AR1248ICC750	PP070009.D	24 Feb 2025 18:30		YP\AJ	Ok
17	AR1248ICC500	AR1248ICC500	PP070010.D	24 Feb 2025 18:46		YP\AJ	Ok
18	AR1248ICC250	AR1248ICC250	PP070011.D	24 Feb 2025 19:03		YP\AJ	Ok

Instrument ID: ECD_P

Daily Analysis Runlog For Sequence/QC Batch ID # PP022425

Review By	yogesh	Review On	2/25/2025 9:00:36 AM
Supervise By	Ankita	Supervise On	2/25/2025 9:48:29 AM
SubDirectory	PP022425	HP Acquire Method	HP Processing Method PP022425
STD. NAME	STD REF.#		
Tune/Reschk			
Initial Calibration Stds	PP23735,PP23736,PP23737,PP23738,PP23739,PP23740,PP23741,PP23742,PP23743,PP23744,PP23745,PP23747,PP23748,PP23749,PP23750,P P23751,PP23752,PP23753,PP23754,PP23755,PP23756,PP23757,PP23758,PP23759,PP23760,PP23761,PP23762,PP23763,PP23764,PP23765,PP 23766,PP23767,PP23768,PP23769,PP23770,PP23771,PP23772,PP23773,PP23774,PP23775		
CCC	PP23737,PP23742,PP23749,PP23754,PP23758,PP23763,PP23768,PP23773		
Internal Standard/PEM			
ICV/I.BLK	PP23778,PP23780,PP23783,PP23784,PP23786,PP23788,PP23790,PP23947		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

19	AR1248ICC050	AR1248ICC050	PP070012.D	24 Feb 2025 19:19		YPIAJ	Ok,M
20	AR1254ICC1000	AR1254ICC1000	PP070013.D	24 Feb 2025 19:35		YPIAJ	Ok,M
21	AR1254ICC750	AR1254ICC750	PP070014.D	24 Feb 2025 19:51		YPIAJ	Ok,M
22	AR1254ICC500	AR1254ICC500	PP070015.D	24 Feb 2025 20:08		YPIAJ	Ok
23	AR1254ICC250	AR1254ICC250	PP070016.D	24 Feb 2025 20:24		YPIAJ	Ok
24	AR1254ICC050	AR1254ICC050	PP070017.D	24 Feb 2025 20:40		YPIAJ	Ok,M
25	AR1262ICC500	AR1262ICC500	PP070018.D	24 Feb 2025 20:56		YPIAJ	Ok
26	AR1268ICC1000	AR1268ICC1000	PP070019.D	24 Feb 2025 21:12		YPIAJ	Ok
27	AR1268ICC750	AR1268ICC750	PP070020.D	24 Feb 2025 21:29		YPIAJ	Ok,M
28	AR1268ICC500	AR1268ICC500	PP070021.D	24 Feb 2025 21:45		YPIAJ	Ok
29	AR1268ICC250	AR1268ICC250	PP070022.D	24 Feb 2025 22:01		YPIAJ	Ok,M
30	AR1268ICC050	AR1268ICC050	PP070023.D	24 Feb 2025 22:17		YPIAJ	Ok,M
31	PP022425ICV500	ICVPP022425	PP070024.D	24 Feb 2025 22:34		YPIAJ	Ok
32	AR1242ICV500	ICVPP022425AR1242	PP070025.D	24 Feb 2025 22:50		YPIAJ	Ok
33	AR1248ICV500	ICVPP022425AR1248	PP070026.D	24 Feb 2025 23:06		YPIAJ	Ok
34	AR1254ICV500	ICVPP022425AR1254	PP070027.D	24 Feb 2025 23:22		YPIAJ	Ok
35	AR1268ICV500	ICVPP022425AR1268	PP070028.D	24 Feb 2025 23:38		YPIAJ	Ok

M : Manual Integration

Instrument ID: ECD_P

Daily Analysis Runlog For Sequence/QC Batch ID # PP030325

Review By	yogesh	Review On	3/3/2025 11:40:00 AM
Supervise By	Ankita	Supervise On	3/4/2025 3:13:46 PM
SubDirectory	PP030325	HP Acquire Method	HP Processing Method PP022425

STD. NAME	STD REF.#
Tune/Reschk	
Initial Calibration Stds	PP23735,PP23736,PP23737,PP23738,PP23739,PP23740,PP23741,PP23742,PP23743,PP23744,PP23745,PP23747,PP23748,PP23749,PP23750,PP23751,PP23752,PP23753,PP23754,PP23755,PP23756,PP23757,PP23758,PP23759,PP23760,PP23761,PP23762,PP23763,PP23764,PP23765,PP23766,PP23767,PP23768,PP23769,PP23770,PP23771,PP23772,PP23773,PP23774,PP23775
CCC	PP23737,PP23742,PP23749,PP23754,PP23758,PP23763,PP23768,PP23773
Internal Standard/PEM	
ICV/I.BLK	PP23778,PP23780,PP23783,PP23784,PP23786,PP23788,PP23790,PP23947
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleId	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	HEXANE	HEXANE	PP070170.D	03 Mar 2025 09:30		YP\AJ	Ok
2	AR1660CCC500	AR1660CCC500	PP070171.D	03 Mar 2025 09:46		YP\AJ	Ok
3	AR1242CCC500	AR1242CCC500	PP070172.D	03 Mar 2025 10:03		YP\AJ	Ok
4	AR1248CCC500	AR1248CCC500	PP070173.D	03 Mar 2025 10:19		YP\AJ	Ok
5	AR1254CCC500	AR1254CCC500	PP070174.D	03 Mar 2025 10:35		YP\AJ	Ok
6	I.BLK	I.BLK	PP070175.D	03 Mar 2025 10:52		YP\AJ	Ok
7	PB166931BL	PB166931BL	PP070176.D	03 Mar 2025 11:08		YP\AJ	Ok
8	Q1457-01	HR-0-02272025	PP070177.D	03 Mar 2025 11:24		YP\AJ	Ok
9	AR1660CCC500	AR1660CCC500	PP070178.D	03 Mar 2025 12:24		YP\AJ	Ok
10	AR1242CCC500	AR1242CCC500	PP070179.D	03 Mar 2025 12:40		YP\AJ	Ok,M
11	AR1248CCC500	AR1248CCC500	PP070180.D	03 Mar 2025 12:56		YP\AJ	Ok
12	AR1254CCC500	AR1254CCC500	PP070181.D	03 Mar 2025 13:13		YP\AJ	Ok
13	I.BLK	I.BLK	PP070182.D	03 Mar 2025 13:29		YP\AJ	Ok
14	PB166954BS	PB166954BS	PP070183.D	03 Mar 2025 15:06		YP\AJ	Ok
15	PB166954BL	PB166954BL	PP070184.D	03 Mar 2025 15:22		YP\AJ	Ok
16	Q1470-01	1359-GAS	PP070185.D	03 Mar 2025 15:38	TCMX high in 1st column	YP\AJ	Ok,M
17	AR1660CCC500	AR1660CCC500	PP070186.D	03 Mar 2025 16:32		YP\AJ	Ok
18	AR1242CCC500	AR1242CCC500	PP070187.D	03 Mar 2025 16:48		YP\AJ	Ok,M

Instrument ID: ECD_P

Daily Analysis Runlog For Sequence/QC Batch ID # PP030325

Review By	yogesh	Review On	3/3/2025 11:40:00 AM
Supervise By	Ankita	Supervise On	3/4/2025 3:13:46 PM
SubDirectory	PP030325	HP Acquire Method	HP Processing Method PP022425
STD. NAME	STD REF.#		
Tune/Reschk			
Initial Calibration Stds	PP23735,PP23736,PP23737,PP23738,PP23739,PP23740,PP23741,PP23742,PP23743,PP23744,PP23745,PP23747,PP23748,PP23749,PP23750,P P23751,PP23752,PP23753,PP23754,PP23755,PP23756,PP23757,PP23758,PP23759,PP23760,PP23761,PP23762,PP23763,PP23764,PP23765,PP 23766,PP23767,PP23768,PP23769,PP23770,PP23771,PP23772,PP23773,PP23774,PP23775		
CCC	PP23737,PP23742,PP23749,PP23754,PP23758,PP23763,PP23768,PP23773		
Internal Standard/PEM			
ICV/I.BLK	PP23778,PP23780,PP23783,PP23784,PP23786,PP23788,PP23790,PP23947		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

19	AR1248CCC500	AR1248CCC500	PP070188.D	03 Mar 2025 17:05		YPIAJ	Ok,M
20	AR1254CCC500	AR1254CCC500	PP070189.D	03 Mar 2025 17:21		YPIAJ	Ok,M
21	I.BLK	I.BLK	PP070190.D	03 Mar 2025 17:37		YPIAJ	Ok
22	PB166955BL	PB166955BL	PP070191.D	03 Mar 2025 17:53		YPIAJ	Ok
23	PB166955BS	PB166955BS	PP070192.D	03 Mar 2025 18:10		YPIAJ	Ok
24	Q1471-01	22225	PP070193.D	03 Mar 2025 18:26		YPIAJ	Ok
25	Q1474-01	BU-03-02282025	PP070194.D	03 Mar 2025 18:42		YPIAJ	Ok
26	Q1474-01MS	BU-03-02282025MS	PP070195.D	03 Mar 2025 18:58		YPIAJ	Ok,M
27	Q1474-01MSD	BU-03-02282025MSD	PP070196.D	03 Mar 2025 19:14		YPIAJ	Ok,M
28	Q1475-01	TR-04-02282025	PP070197.D	03 Mar 2025 19:31		YPIAJ	Ok,M
29	Q1476-02	TRE-25-0016	PP070198.D	03 Mar 2025 19:47		YPIAJ	Ok,M
30	Q1478-14	IDW-SO-COMP-02282	PP070199.D	03 Mar 2025 20:03		YPIAJ	Ok
31	Q1476-03	TRE-25-0017	PP070200.D	03 Mar 2025 20:19	DCB High in both column,TCMX high in 2nd column	YPIAJ	Not Ok
32	AR1660CCC500	AR1660CCC500	PP070201.D	03 Mar 2025 21:08		YPIAJ	Ok
33	AR1242CCC500	AR1242CCC500	PP070202.D	03 Mar 2025 21:24		YPIAJ	Ok,M
34	AR1248CCC500	AR1248CCC500	PP070203.D	03 Mar 2025 21:40		YPIAJ	Ok,M
35	AR1254CCC500	AR1254CCC500	PP070204.D	03 Mar 2025 21:56		YPIAJ	Ok,M
36	I.BLK	I.BLK	PP070205.D	03 Mar 2025 22:13		YPIAJ	Ok

M : Manual Integration

Instrument ID: ECD_P

Daily Analysis Runlog For Sequence/QC Batch ID # PP030425

Review By	yogesh	Review On	3/4/2025 3:37:43 PM
Supervise By	mohammad	Supervise On	3/7/2025 1:26:19 AM
SubDirectory	PP030425	HP Acquire Method	HP Processing Method PP022425

STD. NAME	STD REF.#
Tune/Reschk	
Initial Calibration Stds	PP23735,PP23736,PP23737,PP23738,PP23739,PP23740,PP23741,PP23742,PP23743,PP23744,PP23745,PP23747,PP23748,PP23749,PP23750,PP23751,PP23752,PP23753,PP23754,PP23755,PP23756,PP23757,PP23758,PP23759,PP23760,PP23761,PP23762,PP23763,PP23764,PP23765,PP23766,PP23767,PP23768,PP23769,PP23770,PP23771,PP23772,PP23773,PP23774,PP23775
CCC	PP23737,PP23742,PP23749,PP23754,PP23758,PP23763,PP23768,PP23773
Internal Standard/PEM	
ICV/I.BLK	PP23778,PP23780,PP23783,PP23784,PP23786,PP23788,PP23790,PP23947
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	HEXANE	HEXANE	PP070206.D	04 Mar 2025 08:40		YPIAJ	Ok
2	AR1660CCC500	AR1660CCC500	PP070207.D	04 Mar 2025 08:56		YPIAJ	Ok
3	AR1242CCC500	AR1242CCC500	PP070208.D	04 Mar 2025 10:50		YPIAJ	Ok
4	AR1248CCC500	AR1248CCC500	PP070209.D	04 Mar 2025 11:07		YPIAJ	Ok
5	AR1254CCC500	AR1254CCC500	PP070210.D	04 Mar 2025 11:23	4th peak of 2nd column having F Flag	YPIAJ	Not Ok
6	I.BLK	I.BLK	PP070211.D	04 Mar 2025 11:39		YPIAJ	Ok
7	Q1476-03RE	TRE-25-0017RE	PP070212.D	04 Mar 2025 11:57	DCB High in 1st column,TCMX high in 2nd column	YPIAJ	Not Ok
8	Q1476-03	TRE-25-0017	PP070213.D	04 Mar 2025 12:24	DCB High in 1st column	YPIAJ	Ok,M
9	PB166971BL	PB166971BL	PP070214.D	04 Mar 2025 14:51		YPIAJ	Ok
10	PB166971BS	PB166971BS	PP070215.D	04 Mar 2025 15:08		YPIAJ	Ok,M
11	Q1480-01	TP-1	PP070216.D	04 Mar 2025 15:24		YPIAJ	Ok,M
12	Q1480-01MS	TP-1MS	PP070217.D	04 Mar 2025 15:40		YPIAJ	Ok,M
13	Q1480-01MSD	TP-1MSD	PP070218.D	04 Mar 2025 15:57		YPIAJ	Ok,M
14	Q1480-09	TP-2	PP070219.D	04 Mar 2025 16:13		YPIAJ	Ok,M
15	Q1480-17	TP-3	PP070220.D	04 Mar 2025 16:29		YPIAJ	Ok,M
16	Q1480-25	TP-4	PP070221.D	04 Mar 2025 16:45		YPIAJ	Ok
17	AR1660CCC500	AR1660CCC500	PP070222.D	04 Mar 2025 17:34		YPIAJ	Ok

Instrument ID: ECD_P

Daily Analysis Runlog For Sequence/QC Batch ID # PP030425

Review By	yogesh	Review On	3/4/2025 3:37:43 PM
Supervise By	mohammad	Supervise On	3/7/2025 1:26:19 AM
SubDirectory	PP030425	HP Acquire Method	HP Processing Method PP022425
STD. NAME	STD REF.#		
Tune/Reschk			
Initial Calibration Stds	PP23735,PP23736,PP23737,PP23738,PP23739,PP23740,PP23741,PP23742,PP23743,PP23744,PP23745,PP23747,PP23748,PP23749,PP23750,P P23751,PP23752,PP23753,PP23754,PP23755,PP23756,PP23757,PP23758,PP23759,PP23760,PP23761,PP23762,PP23763,PP23764,PP23765,PP 23766,PP23767,PP23768,PP23769,PP23770,PP23771,PP23772,PP23773,PP23774,PP23775		
CCC	PP23737,PP23742,PP23749,PP23754,PP23758,PP23763,PP23768,PP23773		
Internal Standard/PEM			
ICV/I.BLK	PP23778,PP23780,PP23783,PP23784,PP23786,PP23788,PP23790,PP23947		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

18	AR1242CCC500	AR1242CCC500	PP070223.D	04 Mar 2025 17:50		YPIAJ	Ok,M
19	AR1248CCC500	AR1248CCC500	PP070224.D	04 Mar 2025 18:07		YPIAJ	Ok,M
20	AR1254CCC500	AR1254CCC500	PP070225.D	04 Mar 2025 18:23		YPIAJ	Ok,M
21	I.BLK	I.BLK	PP070226.D	04 Mar 2025 18:39		YPIAJ	Ok
22	PB166964BL	PB166964BL	PP070227.D	04 Mar 2025 18:55		YPIAJ	Ok
23	PB166964BS	PB166964BS	PP070228.D	04 Mar 2025 19:11		YPIAJ	Ok
24	PB166964BSD	PB166964BSD	PP070229.D	04 Mar 2025 19:28		YPIAJ	Ok
25	Q1473-01	BUR-25-0011	PP070230.D	04 Mar 2025 19:44		YPIAJ	Ok
26	Q1473-02	BUR-1293	PP070231.D	04 Mar 2025 20:00	DCB not detected	YPIAJ	ReRun
27	Q1476-04	TRE-25-0023	PP070232.D	04 Mar 2025 20:16		YPIAJ	Ok,M
28	Q1478-01	IDW-AQ-MW-19B-COM	PP070233.D	04 Mar 2025 20:33		YPIAJ	Ok
29	Q1478-03	IDW-AQ-IW-01-COMP	PP070234.D	04 Mar 2025 20:49		YPIAJ	Ok
30	Q1478-07	IDW-AQ-IW-03-COMP	PP070235.D	04 Mar 2025 21:05		YPIAJ	Ok,M
31	Q1480-33	TP-3-WATER-SAMPLE	PP070236.D	04 Mar 2025 21:21		YPIAJ	Ok
32	AR1660CCC500	AR1660CCC500	PP070237.D	04 Mar 2025 22:10		YPIAJ	Ok
33	AR1242CCC500	AR1242CCC500	PP070238.D	04 Mar 2025 22:26		YPIAJ	Ok,M
34	AR1248CCC500	AR1248CCC500	PP070239.D	04 Mar 2025 22:42		YPIAJ	Ok,M
35	AR1254CCC500	AR1254CCC500	PP070240.D	04 Mar 2025 22:59	4th peak of 2nd column having F Flag	YPIAJ	Not Ok
36	I.BLK	I.BLK	PP070241.D	04 Mar 2025 23:15		YPIAJ	Ok

Instrument ID: ECD_P

Daily Analysis Runlog For Sequence/QC Batch ID # PP030425

Review By	yogesh	Review On	3/4/2025 3:37:43 PM
Supervise By	mohammad	Supervise On	3/7/2025 1:26:19 AM
SubDirectory	PP030425	HP Acquire Method	HP Processing Method PP022425
STD. NAME	STD REF.#		
Tune/Reschk	PP23735,PP23736,PP23737,PP23738,PP23739,PP23740,PP23741,PP23742,PP23743,PP23744,PP23745,PP23747,PP23748,PP23749,PP23750,P P23751,PP23752,PP23753,PP23754,PP23755,PP23756,PP23757,PP23758,PP23759,PP23760,PP23761,PP23762,PP23763,PP23764,PP23765,PP 23766,PP23767,PP23768,PP23769,PP23770,PP23771,PP23772,PP23773,PP23774,PP23775 PP23737,PP23742,PP23749,PP23754,PP23758,PP23763,PP23768,PP23773 PP23778,PP23780,PP23783,PP23784,PP23786,PP23788,PP23790,PP23947		
Initial Calibration Stds			
CCC			
Internal Standard/PEM			
ICV/I.BLK			
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

37	Q1478-05	IDW-AQ-IW-02-COMP	PP070242.D	04 Mar 2025 23:31		YP\AJ	Ok
38	Q1476-01	TRE-25-0014	PP070243.D	04 Mar 2025 23:47	All surrogate fail, AR1242 4 peak report	YP\AJ	Not Ok
39	Q1476-05	TRE-25-0024	PP070244.D	05 Mar 2025 00:04	TCMX high in 1st column	YP\AJ	Ok,M
40	AR1660CCC500	AR1660CCC500	PP070245.D	05 Mar 2025 00:52		YP\AJ	Ok
41	AR1242CCC500	AR1242CCC500	PP070246.D	05 Mar 2025 01:08		YP\AJ	Ok,M
42	AR1248CCC500	AR1248CCC500	PP070247.D	05 Mar 2025 01:25		YP\AJ	Ok,M
43	AR1254CCC500	AR1254CCC500	PP070248.D	05 Mar 2025 01:41	4th peak of 2nd column having F Flag	YP\AJ	Not Ok
44	I.BLK	I.BLK	PP070249.D	05 Mar 2025 01:57		YP\AJ	Ok

M : Manual Integration