

Matrix Spike/Matrix Spike Duplicate Summary

SW-846

SDG No.: Q1478

Client: JACOBS Engineering Group, Inc.

Analytical Method: SW8270E

Parameter	Spike	Sample Result	Result	Units	Rec	Rec Qual	RPD	RPD Qual	Low	Limits High	RPD
Lab Sample ID: Q1478-13MS		Client Sample ID: IDW-SO-COMP-022825MS		DataFile: BF141887.D							
Pyridine	500	0	250	ug/L	50				20 (10)	160 (109)	
1,4-Dichlorobenzene	500	0	350	ug/L	70				70 (55)	130 (125)	
2-Methylphenol	500	0	440	ug/L	88				70 (60)	130 (131)	
3+4-Methylphenols	500	0	460	ug/L	92				20 (54)	160 (136)	
Hexachloroethane	500	0	340	ug/L	68				20 (19)	160 (146)	
Nitrobenzene	500	0	460	ug/L	92				70 (62)	130 (112)	
Hexachlorobutadiene	500	0	410	ug/L	82				70 (52)	130 (125)	
2,4,6-Trichlorophenol	500	0	490	ug/L	98				70 (78)	130 (112)	
2,4,5-Trichlorophenol	500	0	530	ug/L	106				70 (71)	130 (111)	
2,4-Dinitrotoluene	500	0	620	ug/L	124				70 (74)	130 (137)	
Hexachlorobenzene	500	0	500	ug/L	100				70 (72)	130 (115)	
Pentachlorophenol	1000	0	1100	ug/L	110				20 (52)	160 (162)	