

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q1478

Client: JACOBS Engineering Group, Inc.

Analytical Method: 8270E

DataFile: BF141884.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	RPD	Limits		
								Qual	Low	High	RPD
PB167016BS	Pyridine	50	32.7	ug/L	65				20 (29)	160 (97)	
	1,4-Dichlorobenzene	50	39.2	ug/L	78				70 (76)	130 (103)	
	2-Methylphenol	50	38.5	ug/L	77				70 (69)	130 (109)	
	3+4-Methylphenols	50	39.2	ug/L	78				20 (67)	160 (106)	
	Hexachloroethane	50	38.9	ug/L	78				20 (76)	160 (118)	
	Nitrobenzene	50	42.0	ug/L	84				70 (58)	130 (106)	
	Hexachlorobutadiene	50	41.5	ug/L	83				70 (69)	130 (101)	
	2,4,6-Trichlorophenol	50	41.6	ug/L	83				70 (61)	130 (110)	
	2,4,5-Trichlorophenol	50	43.8	ug/L	88				70 (70)	130 (106)	
	2,4-Dinitrotoluene	50	51.9	ug/L	104				70 (60)	130 (115)	
	Hexachlorobenzene	50	43.6	ug/L	87				70 (73)	130 (106)	
	Pentachlorophenol	100	91.7	ug/L	92				20 (47)	160 (114)	