

5B

SEMIVOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
DECAFLUOROTRIPHENYLPHOSPHINE (DFTPP)

Lab Name: CHEMTECH

Contract: JAC005

Lab Code: CHEM

SAS No.: Q1478

SDG NO.: Q1478

Lab File ID: BF141792.D

DFTPP Injection Date: 02/27/2025

Instrument ID: BNA_F

DFTPP Injection Time: 14:47

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
51	10.0 - 80.0% of mass 198	34.9
68	Less than 2.0% of mass 69	0.6 (1.7) 1
69	Mass 69 relative abundance	34.9
70	Less than 2.0% of mass 69	0.2 (0.4) 1
127	10.0 - 80.0% of mass 198	46.5
197	Less than 2.0% of mass 198	0.0
198	Base Peak, 100% relative abundance	100
199	5.0 to 9.0% of mass 198	6.7
275	10.0 - 60.0% of mass 198	27.6
365	Greater than 1% of mass 198	3.7
441	Present, but less than mass 443	14.6
442	Greater than 50% of mass 198	100
443	15.0 - 24.0% of mass 442	18.1 (19.4) 2

1-Value is % mass 69

2-Value is % mass 442

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
SSTDICC2.5	SSTDICC2.5	BF141793.D	02/27/2025	15:17
SSTDICC005	SSTDICC005	BF141794.D	02/27/2025	15:46
SSTDICC010	SSTDICC010	BF141795.D	02/27/2025	16:16
SSTDICC020	SSTDICC020	BF141796.D	02/27/2025	16:46
SSTDICCC040	SSTDICCC040	BF141797.D	02/27/2025	17:16
SSTDICC050	SSTDICC050	BF141798.D	02/27/2025	17:46
SSTDICC060	SSTDICC060	BF141799.D	02/27/2025	18:15
SSTDICC080	SSTDICC080	BF141800.D	02/27/2025	18:45