

Matrix Spike/Matrix Spike Duplicate Summary

SW-846

SDG No.: Q1478

Client: JACOBS Engineering Group, Inc.

Analytical Method: SW8260-Low

										Limits	
Parameter	Spike	Sample Result	Result	Units	Rec	Qual	RPD	RPD Qual	Low	High	RPD
Lab Sample ID :	Q1501-05MS	Client Sample ID :	BR-05-465-030525MS					Datafile :	VX045163.D		
Vinyl chloride	50	0	47.4	ug/L	95				70 (69)	130 (125)	
1,1-Dichloroethene	50	0	70.0	ug/L	140	*			70 (53)	130 (162)	
2-Butanone	250	5.70	290	ug/L	114				40 (67)	160 (137)	
Carbon Tetrachloride	50	0	50.1	ug/L	100				70 (66)	130 (133)	
Chloroform	50	0	50.2	ug/L	100				70 (83)	130 (119)	
Benzene	50	0	51.7	ug/L	103				70 (81)	130 (128)	
1,2-Dichloroethane	50	0	51.6	ug/L	103				70 (76)	130 (120)	
Trichloroethene	50	0	94.9	ug/L	190	*			70 (28)	130 (175)	
Tetrachloroethene	50	0	88.9	ug/L	178	*			70 (48)	130 (153)	
Chlorobenzene	50	0	51.6	ug/L	103				70 (85)	130 (114)	