

Matrix Spike/Matrix Spike Duplicate Summary

SW-846

SDG No.: Q1478

Client: JACOBS Engineering Group, Inc.

Analytical Method: SW8260-Low

Limits											
Parameter	Spike	Sample	Result	Units	Rec			RPD	Low	High	RPD
		Result			Qual	RPD	Qual				
Lab Sample ID :	Q1501-06MSD	Client Sample ID :	BR-05-465-030525MSD						Datafile :	VX045164.D	
Vinyl chloride	50	0	43.3	ug/L	87		9		70 (69)	130 (125)	20 (20)
1,1-Dichloroethene	50	0	73.0	ug/L	146	*	4		70 (53)	130 (162)	20 (20)
2-Butanone	250	5.70	280	ug/L	110		4		40 (67)	160 (137)	20 (20)
Carbon Tetrachloride	50	0	47.3	ug/L	95		6		70 (66)	130 (133)	20 (20)
Chloroform	50	0	48.6	ug/L	97		3		70 (83)	130 (119)	20 (20)
Benzene	50	0	49.0	ug/L	98		5		70 (81)	130 (128)	20 (20)
1,2-Dichloroethane	50	0	49.8	ug/L	100		4		70 (76)	130 (120)	20 (20)
Trichloroethene	50	0	91.5	ug/L	183	*	4		70 (28)	130 (175)	20 (20)
Tetrachloroethene	50	0	85.3	ug/L	171	*	4		70 (48)	130 (153)	20 (20)
Chlorobenzene	50	0	49.8	ug/L	100		4		70 (85)	130 (114)	20 (20)