

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q1478

Client: JACOBS Engineering Group, Inc.

Analytical Method: SW8260-Low

Datafile : VX045158.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VX0306WBS01	Vinyl chloride	20	19.4	ug/L	97			70 (65)	130 (117)	
	1,1-Dichloroethene	20	21.3	ug/L	106			70 (74)	130 (110)	
	2-Butanone	100	100	ug/L	100			40 (65)	160 (122)	
	Carbon Tetrachloride	20	20.5	ug/L	103			70 (77)	130 (113)	
	Chloroform	20	21.2	ug/L	106			70 (79)	130 (113)	
	Benzene	20	21.5	ug/L	108			70 (82)	130 (109)	
	1,2-Dichloroethane	20	21.4	ug/L	107			70 (80)	130 (115)	
	Trichloroethene	20	20.8	ug/L	104			70 (77)	130 (113)	
	Tetrachloroethene	20	22.2	ug/L	111			70 (67)	130 (123)	
	Chlorobenzene	20	21.6	ug/L	108			70 (82)	130 (109)	