

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX030325\
 Data File : VX045102.D
 Acq On : 03 Mar 2025 09:39
 Operator : JC/MD
 Sample : VX0303WBL01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0303WBL01

Quant Time: Mar 04 05:18:31 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022825W.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 28 06:45:16 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	88708	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	176545	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	160387	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	70988	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.946	65	71659	50.785	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 101.560%	
35) Dibromofluoromethane	5.379	113	60357	51.128	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 102.260%	
50) Toluene-d8	8.647	98	219822	51.363	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 102.720%	
62) 4-Bromofluorobenzene	11.079	95	74529	52.552	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	= 105.100%	


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX030325\
Data File : VX045102.D
Acq On : 03 Mar 2025 09:39
Operator : JC/MD
Sample : VX0303WBL01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0303WBL01

Quant Time: Mar 04 05:18:31 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022825W.M
Quant Title : SW846 8260
QLast Update : Fri Feb 28 06:45:16 2025
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 5.544 min Scan# 71
Delta R.T. -0.006 min
Lab File: VX045102.D
Acq: 03 Mar 2025 09:39




Instrument :
MSVOA_X
ClientSampleId :
VX0303WBL01


Tgt Ion:168 Resp: 88708
Ion Ratio Lower Upper
168 100
99 62.5 48.2 72.4



#33
1,2-Dichloroethane-d4
Concen: 50.785 ug/l
RT: 5.946 min Scan# 797
Delta R.T. -0.006 min
Lab File: VX045102.D
Acq: 03 Mar 2025 09:39

Tgt Ion: 65 Resp: 71659
Ion Ratio Lower Upper
65 100
67 51.9 0.0 106.2









#34
1,4-Difluorobenzene
Concen: 50.000 ug/l
RT: 6.757 min Scan# 91
Delta R.T. -0.000 min
Lab File: VX045102.D
Acq: 03 Mar 2025 09:39




Instrument :
MSVOA_X
ClientSampleId :
VX0303WBL01


Tgt Ion:114 Resp: 176545
Ion Ratio Lower Upper
114 100
63 21.1 0.0 41.8
88 15.6 0.0 32.8



#35
Dibromofluoromethane
Concen: 51.128 ug/l
RT: 5.379 min Scan# 704
Delta R.T. -0.006 min
Lab File: VX045102.D
Acq: 03 Mar 2025 09:39

Tgt Ion:113 Resp: 60357
Ion Ratio Lower Upper
113 100
111 102.3 81.8 122.6
192 17.8 14.3 21.5









#50
Toluene-d8
Concen: 51.363 ug/l
RT: 8.647 min Scan# 11
Delta R.T. -0.000 min
Lab File: VX045102.D
Acq: 03 Mar 2025 09:39




Instrument :
MSVOA_X
ClientSampleId :
VX0303WBL01


Tgt Ion: 98 Resp: 219822
Ion Ratio Lower Upper
98 100
100 65.3 52.0 78.0



#62
4-Bromofluorobenzene
Concen: 52.552 ug/l
RT: 11.079 min Scan# 1639
Delta R.T. -0.000 min
Lab File: VX045102.D
Acq: 03 Mar 2025 09:39

Tgt Ion: 95 Resp: 74529
Ion Ratio Lower Upper
95 100
174 75.6 0.0 148.2
176 72.7 0.0 141.4









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 10.049 min Scan# 1470
Delta R.T. -0.000 min
Lab File: VX045102.D
Acq: 03 Mar 2025 09:39

Instrument :
MSVOA_X
ClientSampleId :
VX0303WBL01

Tgt Ion:117 Resp: 160387
Ion Ratio Lower Upper
117 100
82 57.6 46.3 69.5
119 31.6 25.7 38.5



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 12.018 min Scan# 1793
Delta R.T. -0.000 min
Lab File: VX045102.D
Acq: 03 Mar 2025 09:39

Tgt Ion:152 Resp: 70988
Ion Ratio Lower Upper
152 100
115 60.4 44.2 132.6
150 155.9 0.0 349.0

