

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: JACO05
 Lab Code: CHEM Case No.: Q1478 SAS No.: Q1478 SDG NO.: Q1478
 Lab File ID: VX045100.D Date Analyzed: 03/03/2025
 Instrument ID: MSVOA_X Time Analyzed: 08:47
 GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	110336	5.54	188916	6.75	164466	10.05
UPPER LIMIT	220672	6.038	377832	7.251	328932	10.549
LOWER LIMIT	55168	5.038	94458	6.251	82233	9.549
EPA SAMPLE NO.						
IDW-AQ-DRUM-610-022825	84810	5.54	169281	6.76	138779	10.05
IDW-AQ-DRUM-616-022825	77986	5.54	153844	6.76	143555	10.05
IDW-AQ-DRUM-614-022825	75051	5.54	147850	6.75	135587	10.05
VX0303WBL01	88708	5.54	176545	6.76	160387	10.05
VX0303WBS01	102460	5.54	180548	6.75	158197	10.05

IS1 = Pentafluorobenzene
 IS2 = 1,4-Difluorobenzene
 IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.