

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: JACO05
 Lab Code: CHEM Case No.: Q1478 SAS No.: Q1478 SDG NO.: Q1478
 Lab File ID: VX045127.D Date Analyzed: 03/05/2025
 Instrument ID: MSVOA_X Time Analyzed: 09:12
 GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	112291	5.54	190470	6.75	164207	10.05
UPPER LIMIT	224582	6.044	380940	7.251	328414	10.549
LOWER LIMIT	56145.5	5.044	95235	6.251	82103.5	9.549
EPA SAMPLE NO.						
IDW-AQ-DRUM-610-022825DL	75743	5.54	152545	6.76	141767	10.05
IDW-AQ-DRUM-616-022825DL	75457	5.55	145945	6.76	124906	10.06
IDW-AQ-DRUM-614-022825DL	73767	5.55	147372	6.76	131471	10.05
IDW-AQ-DRUM-612-022825	74500	5.55	149821	6.76	139770	10.05
VX0305WBL01	68143	5.54	136785	6.76	125445	10.05
VX0305WBS01	98769	5.54	177910	6.76	153160	10.06
VX0305WBSD01	104383	5.55	188951	6.76	167692	10.05

IS1 = Pentafluorobenzene

IS2 = 1,4-Difluorobenzene

IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.