

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: JACO05
 Lab Code: CHEM Case No.: Q1478 SAS No.: Q1478 SDG NO.: Q1478
 Lab File ID: VX045100.D Date Analyzed: 03/03/2025
 Instrument ID: MSVOA_X Time Analyzed: 08:47
 GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS4 AREA #	RT #				
12 HOUR STD	76846	12.018				
UPPER LIMIT	153692	12.518				
LOWER LIMIT	38423	11.518				
EPA SAMPLE NO.						
IDW-AQ-DRUM-610-022825	68503	12.02				
IDW-AQ-DRUM-616-022825	60766	12.02				
IDW-AQ-DRUM-614-022825	59630	12.02				
VX0303WBL01	70988	12.02				
VX0303WBS01	74771	12.02				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.