

VOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
BROMOFLUOROBENZENE (BFB)

Lab Name: CHEMTECH Contract: JACO05
Lab Code: CHEM Case No.: Q1478 SAS No.: Q1478 SDG NO.: Q1478
Lab File ID: VX045126.D BFB Injection Date: 03/05/2025
Instrument ID: MSVOA_X BFB Injection Time: 08:40
GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: Y/N N

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
50	15.0 - 40.0% of mass 95	21
75	30.0 - 60.0% of mass 95	53.5
95	Base Peak, 100% relative abundance	100
96	5.0 - 9.0% of mass 95	7.1
173	Less than 2.0% of mass 174	0.8 (1.1) 1
174	50.0 - 100.0% of mass 95	77
175	5.0 - 9.0% of mass 174	5.1 (6.7) 1
176	95.0 - 101.0% of mass 174	73.3 (95.1) 1
177	5.0 - 9.0% of mass 176	5 (6.9) 2

1-Value is % mass 174

2-Value is % mass 176

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
VSTDCCC050	VSTDCCC050	VX045127.D	03/05/2025	09:12
VX0305WBL01	VX0305WBL01	VX045129.D	03/05/2025	10:03
VX0305WBS01	VX0305WBS01	VX045130.D	03/05/2025	10:26
VX0305WBSD01	VX0305WBSD01	VX045135.D	03/05/2025	12:26
IDW-AQ-DRUM-616-022825DL	Q1478-04DL	VX045136.D	03/05/2025	12:50
IDW-AQ-DRUM-614-022825DL	Q1478-06DL	VX045137.D	03/05/2025	13:13
IDW-AQ-DRUM-610-022825DL	Q1478-02DL	VX045144.D	03/05/2025	15:56
IDW-AQ-DRUM-612-022825	Q1478-08	VX045145.D	03/05/2025	16:19