

Instrument ID: MSVOA_X

Daily Analysis Runlog For Sequence/QC Batch ID # VX030525

Review By	John Carlone	Review On	3/6/2025 8:28:41 AM
Supervise By	Maresh Dadoda	Supervise On	3/6/2025 2:56:36 PM
SubDirectory	VX030525	HP Acquire Method	HP Processing Method 82X022825W.M

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds	VP133217
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP133218,VP133219

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VX045126.D	05 Mar 2025 08:40		JC/MD	Ok
2	VSTDCCC050	VSTDCCC050	VX045127.D	05 Mar 2025 09:12	pH#Lot#V12668	JC/MD	Ok,M
3	VX0305MBL01	VX0305MBL01	VX045128.D	05 Mar 2025 09:39		JC/MD	Ok
4	VX0305WBL01	VX0305WBL01	VX045129.D	05 Mar 2025 10:03		JC/MD	Ok
5	VX0305WBS01	VX0305WBS01	VX045130.D	05 Mar 2025 10:26		JC/MD	Ok,M
6	VX0305MBS01	VX0305MBS01	VX045131.D	05 Mar 2025 10:53		JC/MD	Ok,M
7	Q1480-17	TP-3	VX045132.D	05 Mar 2025 11:16		JC/MD	Ok
8	Q1448-04ME	P3ME	VX045133.D	05 Mar 2025 11:40		JC/MD	Ok
9	IBLK	IBLK	VX045134.D	05 Mar 2025 12:03		JC/MD	Ok
10	VX0305WBSD01	VX0305WBSD01	VX045135.D	05 Mar 2025 12:26		JC/MD	Ok,M
11	Q1478-04DL	IDW-AQ-DRUM-616-02	VX045136.D	05 Mar 2025 12:50	vial B pH<2	JC/MD	Ok
12	Q1478-06DL	IDW-AQ-DRUM-614-02	VX045137.D	05 Mar 2025 13:13	vial B pH<2	JC/MD	Ok
13	Q1478-02DL	IDW-AQ-DRUM-610-02	VX045138.D	05 Mar 2025 13:36	Not req.	JC/MD	Not Ok
14	Q1478-08	IDW-AQ-DRUM-612-02	VX045139.D	05 Mar 2025 13:59	Need lower dilution	JC/MD	Not Ok
15	VX0305MBSD01	VX0305MBSD01	VX045140.D	05 Mar 2025 14:23		JC/MD	Ok,M
16	Q1448-04MEDL	P3MEDL	VX045141.D	05 Mar 2025 14:46	Not req.	JC/MD	Not Ok
17	IBLK	IBLK	VX045142.D	05 Mar 2025 15:09		JC/MD	Ok
18	Q1478-08	IDW-AQ-DRUM-612-02	VX045143.D	05 Mar 2025 15:32	Need lower dilution	JC/MD	Not Ok

Instrument ID: MSVOA_X

Daily Analysis Runlog For Sequence/QC Batch ID # VX030525

Review By	John Carlone	Review On	3/6/2025 8:28:41 AM
Supervise By	Maresh Dadoda	Supervise On	3/6/2025 2:56:36 PM
SubDirectory	VX030525	HP Acquire Method	HP Processing Method 82X022825W.M
STD. NAME	STD REF.#		
Tune/Reschk Initial Calibration Stds	VP133217		
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP133218,VP133219		

19	Q1478-02DL	IDW-AQ-DRUM-610-02	VX045144.D	05 Mar 2025 15:56	vial B pH<2	JC/MD	Ok
20	Q1478-08	IDW-AQ-DRUM-612-02	VX045145.D	05 Mar 2025 16:19	vial A pH<2	JC/MD	Ok
21	Q1501-03	EB01-030525	VX045146.D	05 Mar 2025 16:42	vial A pH<2 EB	JC/MD	Ok
22	Q1501-07	TB01-030525	VX045147.D	05 Mar 2025 17:05	vial A pH<2 TB	JC/MD	Ok
23	Q1501-04	BR-05-465-030525	VX045148.D	05 Mar 2025 17:28	Need Straight Run	JC/MD	Not Ok
24	Q1488-15	TB03042025	VX045149.D	05 Mar 2025 17:52	vial A pH<2 TB	JC/MD	Ok
25	Q1488-13	ENV-102-GW01	VX045150.D	05 Mar 2025 18:15	vial A pH<2	JC/MD	Ok
26	Q1488-14	ENV-104-GW01	VX045151.D	05 Mar 2025 18:38	vial A pH<2 Need 4X	JC/MD	Dilution
27	Q1494-01	PURGE-WATER	VX045152.D	05 Mar 2025 19:01	vial A pH<2 E flag in previous sample	JC/MD	ReRun
28	VSTDCCC050	VSTDCCC050EC	VX045153.D	05 Mar 2025 19:24		JC/MD	Ok,M

M : Manual Integration