

Instrument ID: MSVOA_Y

Daily Analysis Runlog For Sequence/QC Batch ID # VY030425

Review By	Mahesh Dadoda	Review On	3/5/2025 12:14:39 PM		
Supervise By	Semsettin Yesilyurt	Supervise On	3/5/2025 12:18:08 PM		
SubDirectory	VY030425	HP Acquire Method	MSVOA_Y	HP Processing Method	82y030425s.m

STD. NAME	STD REF.#
Tune/Reschk	VP133216
Initial Calibration Stds	VP133207,VP133208,VP133209,VP133210,VP133211,VP133212
CCC	VP133214,VP133215
Internal Standard/PEM	VP131783
ICV/I.BLK	VP133213
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VY021395.D	04 Mar 2025 08:52		SY/MD	Ok
2	VSTDICC005	VSTDICC005	VY021396.D	04 Mar 2025 09:23	not used	SY/MD	Not Ok
3	VSTDICC010	VSTDICC010	VY021397.D	04 Mar 2025 09:46		SY/MD	Ok,M
4	VSTDICC020	VSTDICC020	VY021398.D	04 Mar 2025 10:09	Comp.#11 is on Linear Regression	SY/MD	Ok,M
5	VSTDICCC050	VSTDICCC050	VY021399.D	04 Mar 2025 10:30	Comp.#95 is on Quadratic Regression	SY/MD	Ok,M
6	VSTDICC100	VSTDICC100	VY021400.D	04 Mar 2025 11:06		SY/MD	Ok,M
7	VSTDICC150	VSTDICC150	VY021401.D	04 Mar 2025 11:29	Method fail for comp.#13	SY/MD	Ok,M
8	VIBLK	VIBLK	VY021402.D	04 Mar 2025 11:52		SY/MD	Ok
9	VSTDICC005	VSTDICC005	VY021403.D	04 Mar 2025 12:15		SY/MD	Ok,M
10	VSTDICV050	ICVVY030425	VY021404.D	04 Mar 2025 13:12		SY/MD	Ok,M

M : Manual Integration

Instrument ID: MSVOA_Y

Daily Analysis Runlog For Sequence/QC Batch ID # VY030625

Review By	Maresh Dadoda	Review On	3/7/2025 1:32:16 PM
Supervise By	Semsettin Yesilyurt	Supervise On	3/7/2025 1:32:53 PM
SubDirectory	VY030625	HP Acquire Method	HP Processing Method 82y030425s.m

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds	VP133223
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP133224,VP133225 VP131783

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VY021425.D	06 Mar 2025 09:39		SY/MD	Ok
2	VSTDCCC050	VSTDCCC050	VY021426.D	06 Mar 2025 10:10		SY/MD	Ok,M
3	VY0306SBL01	VY0306SBL01	VY021427.D	06 Mar 2025 11:59		SY/MD	Ok
4	VY0306SBS01	VY0306SBS01	VY021428.D	06 Mar 2025 12:33		SY/MD	Ok,M
5	VY0306SBSD01	VY0306SBSD01	VY021429.D	06 Mar 2025 12:55		SY/MD	Ok,M
6	Q1489-05	TP-2	VY021430.D	06 Mar 2025 13:28	vial-B	SY/MD	Ok
7	Q1489-01	TP-1	VY021431.D	06 Mar 2025 13:52	vial-B Not purge	SY/MD	Not Ok
8	Q1421-09	P001-CLAY-CF02-01	VY021432.D	06 Mar 2025 14:15	vial-B	SY/MD	Ok
9	Q1479-01	P5	VY021433.D	06 Mar 2025 14:39	vial-A	SY/MD	Ok
10	Q1479-02	DSP2	VY021434.D	06 Mar 2025 15:02	vial-A Internal Standard Fail	SY/MD	ReRun
11	Q1488-01	ENV-101-SB01	VY021435.D	06 Mar 2025 15:25	vial-A	SY/MD	Ok
12	Q1488-03	ENV-101-SB02	VY021436.D	06 Mar 2025 15:49	vial-A	SY/MD	Ok
13	Q1488-05	ENV-102-SB01	VY021437.D	06 Mar 2025 16:12	vial-A	SY/MD	Ok
14	Q1488-07	ENV-102-SB02	VY021438.D	06 Mar 2025 16:36	vial-A Not purge	SY/MD	Not Ok
15	Q1488-09	ENV-104-SB01	VY021439.D	06 Mar 2025 16:59	vial-A Internal Standard Fail	SY/MD	ReRun
16	Q1488-11	ENV-104-SB02	VY021440.D	06 Mar 2025 17:22	vial-A Not purge	SY/MD	Not Ok
17	Q1463-01	T1	VY021441.D	06 Mar 2025 17:46	vial-B	SY/MD	Ok
18	Q1463-05	T5	VY021442.D	06 Mar 2025 18:09	vial-B	SY/MD	Ok

Instrument ID: MSVOA_Y

Daily Analysis Runlog For Sequence/QC Batch ID # VY030625

Review By	Maresh Dadoda	Review On	3/7/2025 1:32:16 PM
Supervise By	Semsettin Yesilyurt	Supervise On	3/7/2025 1:32:53 PM
SubDirectory	VY030625	HP Acquire Method	HP Processing Method 82y030425s.m
STD. NAME	STD REF.#		
Tune/Reschk Initial Calibration Stds	VP133223		
CCC	VP133224,VP133225		
Internal Standard/PEM	VP131783		
ICV/I.BLK			
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

19	Q1463-04	T4	VY021443.D	06 Mar 2025 18:33	Vial-B	SY/MD	Ok
20	Q1463-06	T6	VY021444.D	06 Mar 2025 18:56	Vial-A	SY/MD	Ok
21	Q1463-07	T7	VY021445.D	06 Mar 2025 19:20	Vial-A	SY/MD	Ok
22	Q1463-08	T8	VY021446.D	06 Mar 2025 19:43	Vial-A	SY/MD	Ok
23	Q1463-09	T9	VY021447.D	06 Mar 2025 20:06	Vial-A	SY/MD	Ok
24	Q1463-10	T10	VY021448.D	06 Mar 2025 20:30	Vial-A Internal Standard Fail; Surrogate Fail	SY/MD	ReRun
25	VSTDCCC050	VSTDCCC050EC	VY021449.D	06 Mar 2025 21:15		SY/MD	Ok,M

M : Manual Integration

Instrument ID: MSVOA_Y

Daily Analysis Runlog For Sequence/QC Batch ID # VY030725

Review By	Maresh Dadoda	Review On	3/11/2025 2:51:48 PM
Supervise By	Semsettin Yesilyurt	Supervise On	3/11/2025 2:52:31 PM
SubDirectory	VY030725	HP Acquire Method	MSVOA_Y HP Processing Method 82y030425s.m

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds	VP133232
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP133233,VP133234 VP131783

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VY021450.D	07 Mar 2025 09:27		SY/MD	Ok
2	VSTDCCC050	VSTDCCC050	VY021451.D	07 Mar 2025 09:59		SY/MD	Ok,M
3	VY0307SBL01	VY0307SBL01	VY021452.D	07 Mar 2025 10:35		SY/MD	Ok
4	VY0307SBS01	VY0307SBS01	VY021453.D	07 Mar 2025 12:03		SY/MD	Ok,M
5	VY0307SBSD01	VY0307SBSD01	VY021454.D	07 Mar 2025 12:26	BSD Failed Low for com.#73,87,88	SY/MD	Ok,M
6	Q1463-03	T3	VY021455.D	07 Mar 2025 12:55	vial-B	SY/MD	Ok,M
7	Q1479-02	DSP2	VY021456.D	07 Mar 2025 13:18	vial-B	SY/MD	Ok,M
8	Q1463-02	T2	VY021457.D	07 Mar 2025 13:42	vial-A Internal standard fail	SY/MD	ReRun
9	Q1463-10	T10	VY021458.D	07 Mar 2025 14:05	vial-B	SY/MD	Ok
10	Q1488-07	ENV-102-SB02	VY021459.D	07 Mar 2025 14:28	vial-B	SY/MD	Ok,M
11	Q1488-09	ENV-104-SB01	VY021460.D	07 Mar 2025 14:52	Vial-B	SY/MD	Ok
12	Q1488-11	ENV-104-SB02	VY021461.D	07 Mar 2025 15:15	Vial-B Internal standard fail	SY/MD	Ok
13	Q1514-01	ENV-105-SB01	VY021462.D	07 Mar 2025 15:39	Vial-A BSD failed Low	SY/MD	ReRun
14	Q1514-03	ENV-105-SB02	VY021463.D	07 Mar 2025 16:02	Vial-A BSD failed Low	SY/MD	ReRun
15	Q1514-05	ENV-103-SB01	VY021464.D	07 Mar 2025 16:25	Vial-A BSD failed Low	SY/MD	ReRun
16	Q1515-02	AU-06-030625-E2	VY021465.D	07 Mar 2025 16:49	Vial-A Internal standard fail; Surrogate fail	SY/MD	ReRun
17	Q1507-01	50-MIDDLESEX-AVE	VY021466.D	07 Mar 2025 17:12	Vial-A Not purge	SY/MD	Not Ok

Instrument ID: MSVOA_Y

Daily Analysis Runlog For Sequence/QC Batch ID # VY030725

Review By	Mahesh Dadoda	Review On	3/11/2025 2:51:48 PM		
Supervise By	Semsettin Yesilyurt	Supervise On	3/11/2025 2:52:31 PM		
SubDirectory	VY030725	HP Acquire Method	MSVOA_Y	HP Processing Method	82y030425s.m
STD. NAME		STD REF.#			
Tune/Reschk Initial Calibration Stds		VP133232			
CCC		VP133233,VP133234			
Internal Standard/PEM		VP131783			
ICV/I.BLK					
Surrogate Standard					
MS/MSD Standard					
LCS Standard					

18	Q1508-01	RBR251372	VY021467.D	07 Mar 2025 17:36	Vial-A Internal standard fail	SY/MD	ReRun
19	Q1494-05	SOIL	VY021468.D	07 Mar 2025 17:59	Vial-A Internal standard fail	SY/MD	ReRun
20	Q1494-03	ASPHALT-SOIL	VY021469.D	07 Mar 2025 18:23	Vial-A Internal standard fail	SY/MD	ReRun
21	Q1494-07	SOIL-COMP	VY021470.D	07 Mar 2025 18:46	Vial-A Internal standard fail	SY/MD	ReRun
22	Q1490-01	CLAY-SLUDGE-DRUM	VY021471.D	07 Mar 2025 19:09	Vial-A Not purge	SY/MD	Not Ok
23	VSTDCCC050	VSTDCCC050EC	VY021472.D	07 Mar 2025 19:32		SY/MD	Ok,M

M : Manual Integration