

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: GENV01
 Lab Code: CHEM Case No.: Q1479 SAS No.: Q1479 SDG NO.: Q1479
 Lab File ID: VY021451.D Date Analyzed: 03/07/2025
 Instrument ID: MSVOA_Y Time Analyzed: 09:59
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) Y

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	212653	7.71	324643	8.62	291936	11.42
UPPER LIMIT	425306	8.207	649286	9.116	583872	11.92
LOWER LIMIT	106327	7.207	162322	8.116	145968	10.92
EPA SAMPLE NO.						
DSP2	337523	7.68	510119	8.60	399863	11.43
VY0307SBL01	247432	7.71	455944	8.62	386511	11.42
VY0307SBS01	219868	7.71	343563	8.62	301562	11.42

IS1 = Pentafluorobenzene
 IS2 = 1,4-Difluorobenzene
 IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.