

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: GENV01
 Lab Code: CHEM Case No.: Q1479 SAS No.: Q1479 SDG NO.: Q1479
 Lab File ID: VY021426.D Date Analyzed: 03/06/2025
 Instrument ID: MSVOA_Y Time Analyzed: 10:10
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) Y

	IS4 AREA #	RT #				
12 HOUR STD	168733	13.346				
UPPER LIMIT	337466	13.846				
LOWER LIMIT	84366.5	12.846				
EPA SAMPLE NO.						
P5	133949	13.35				
VY0306SBL01	156000	13.35				
VY0306SBS01	152176	13.35				
VY0306SBSD01	145541	13.35				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.