

CASE NARRATIVE

GFE LLC

Project Name: 97-36 63rd Rd, Rego Park NY

Project # N/A

Chemtech Project # Q1483

Test Name: VOCMS Group3

A. Number of Samples and Date of Receipt:

6 Air samples were received on 03/04/2025.

B. Parameters

According to the Chain of Custody document, the following analyses were requested: VOCMS Group3. This data package contains results for VOCMS Group3.

C. Analytical Techniques:

The analysis performed on instrument MSVOA_L were done using GC column RTX-1, which is 60 meters, 0.32 mm id, 1.0 um df, Restek Cat. #10157. The Trap was supplied by Entech, glass bead and Tenax , Entech 7100A Preconcentrator. The analysis of VOCMS Group3 was based on method TO-15.

D. QA/ QC Samples:

The Holding Times were met for all analysis.

The Surrogate recoveries met the acceptable criteria.

The Internal Standards Areas met the acceptable requirements.

The Retention Times were acceptable for all samples.

The RPD met criteria.

The Blank Spike met requirements for all samples.

The Blank Spike Duplicate met requirements for all samples.

The Blank analysis did not indicate the presence of lab contamination.

The Initial Calibration met the requirements.

The Continuous Calibration met the requirements.

The Tuning criteria met requirements.

Due to potential high concentration of target analytes, Samples SV1, SV2, SV3 and SV4 were initially diluted.

Samples IA2, SV3 were diluted due to high concentrations.

E. Additional Comments:

Sample #SV1, SV2, SV3DL and SV4 concentrations of some compounds exceeded calibration range and due to limitation of the Matrix and the analytical system, Further dilution of the sample was not possible therefore reported with E flag.



284 Sheffield Street, Mountainside, NJ 07092
Phone: 908 789 8900 Fax: 908 789 8922

F. Manual Integration Comments:

Please refer to the Manual integration Report included with the Run Logs for information on the manual integrations performed.

I certify that the data package is in compliance with the terms and conditions of the contract, both technically and for completeness, for other than the conditions detailed above. The laboratory manager or his designee, as verified by the following signature has authorized release of the data contained in this hard copy data package.

Signature_____