

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: GFEL01

Lab Code: CHEM Case No.: Q1483 SAS No.: Q1483 SDG No.: Q1483

Instrument ID: MSVOA_L Calibration Date/Time: 03/05/2025 08:30

Lab File ID: VL042095.D Init. Calib. Date(s): 02/25/2025 02/25/2025

Heated Purge: (Y/N) N Init. Calib. Time(s): 08:56 12:07

GC Column: RTX-1 ID: 0.32 (mm)

COMPOUND	RRF	RRF010	MIN RRF	%D	MAX%D
Vinyl Chloride	0.364	0.422		15.93	30
1,1-Dichloroethene	0.376	0.402		6.91	30
cis-1,2-Dichloroethene	1.353	1.318		-2.59	30
1,1,1-Trichloroethane	1.866	1.769		-5.2	30
Trichloroethene	0.427	0.436		2.11	30
Tetrachloroethene	0.358	0.368		2.79	30
1-Bromo-4-Fluorobenzene	0.769	0.780		1.43	30

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.