

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q1483

Client: GFE LLC

Analytical Method: SWTO-15

Datafile : VL042097.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VL0305ABS01	Vinyl Chloride	10	11.9	ppbv	119			70	130	
	1,1-Dichloroethene	10	10.9	ppbv	109			70	130	
	cis-1,2-Dichloroethene	10	9.80	ppbv	98			70	130	
	1,1,1-Trichloroethane	10	9.80	ppbv	98			70	130	
	Trichloroethene	10	10.3	ppbv	103			70	130	
	Tetrachloroethene	10	10.6	ppbv	106			70	130	