

INITIAL CALIBRATION OF MULTICOMPONENT ANALYTES

Contract: JPCL01

Lab Code: CHEM **Case No.:** Q1485 **SAS No.:** Q1485 **SDG NO.:** Q1485

Instrument ID: ECD_P **Date(s) Analyzed:** 02/24/2025 02/24/2025

GC Column: ZB-MR1 **ID:** 0.32 (mm)

COMPOUND	AMOUNT (ng)	PEAK	RT	RT WINDOW		CALIBRATION FACTOR
				FROM	TO	
Aroclor-1221	500	1	4.73	4.63	4.83	18915900
		2	4.81	4.71	4.91	14591600
		3	4.89	4.79	4.99	43158200
		4	0.00			0
		5	0.00			0
Aroclor-1232	500	1	4.89	4.79	4.99	32884000
		2	5.42	5.32	5.52	16260200
		3	5.70	5.60	5.80	34867200
		4	5.86	5.76	5.96	17799400
		5	5.95	5.85	6.05	12607200
Aroclor-1262	500	1	8.11	8.01	8.21	80729000
		2	8.43	8.33	8.53	161087000
		3	8.75	8.65	8.85	110612000
		4	8.84	8.74	8.94	83914000
		5	9.49	9.39	9.59	58804600

INITIAL CALIBRATION OF MULTICOMPONENT ANALYTES

Contract: JPCL01

Lab Code: CHEM **Case No.:** Q1485 **SAS No.:** Q1485 **SDG NO.:** Q1485

Instrument ID: ECD_P **Date(s) Analyzed:** 02/24/2025 02/24/2025

GC Column: ZB-MR2 **ID:** 0.32 (mm)

COMPOUND	AMOUNT (ng)	PEAK	RT	RT WINDOW		CALIBRATION FACTOR
				FROM	TO	
Aroclor-1221	500	1	4.04	3.94	4.14	13421200
		2	4.13	4.03	4.23	10166200
		3	4.21	4.11	4.31	30359600
		4	0.00			0
		5	0.00			0
Aroclor-1232	500	1	4.21	4.11	4.31	22742000
		2	4.94	4.84	5.04	22833200
		3	5.12	5.02	5.22	12024000
		4	5.20	5.10	5.30	10865300
		5	5.37	5.27	5.47	11305400
Aroclor-1262	500	1	6.97	6.87	7.07	83475400
		2	7.23	7.13	7.33	66291000
		3	7.75	7.65	7.85	60580400
		4	7.81	7.71	7.91	105474000
		5	8.32	8.22	8.42	51930200