

### CALIBRATION VERIFICATION SUMMARY

Contract: JPCL01

Lab Code: CHEM Case No.: Q1485 SAS No.: Q1485 SDG NO.: Q1485

Continuing Calib Date: 03/05/2025 Initial Calibration Date(s): 02/24/2025 02/24/2025

Continuing Calib Time: 18:30 Initial Calibration Time(s): 14:59 22:17

GC Column: ZB-MR1 ID: 0.32 (mm)

| COMPOUND             | CCAL<br>RT | AVG<br>RT | RT WINDOW |       | DIFF<br>RT |
|----------------------|------------|-----------|-----------|-------|------------|
|                      |            |           | FROM      | TO    |            |
| Aroclor-1242-1 (1)   | 5.68       | 5.68      | 5.58      | 5.78  | 0.00       |
| Aroclor-1242-2 (2)   | 5.70       | 5.70      | 5.60      | 5.80  | 0.00       |
| Aroclor-1242-3 (3)   | 5.76       | 5.77      | 5.67      | 5.87  | 0.01       |
| Aroclor-1242-4 (4)   | 5.86       | 5.86      | 5.76      | 5.96  | 0.00       |
| Aroclor-1242-5 (5)   | 6.59       | 6.59      | 6.49      | 6.69  | 0.00       |
| Tetrachloro-m-xylene | 4.52       | 4.53      | 4.43      | 4.63  | 0.01       |
| Decachlorobiphenyl   | 10.25      | 10.26     | 10.16     | 10.36 | 0.01       |

### CALIBRATION VERIFICATION SUMMARY

Contract: JPCL01

Lab Code: CHEM Case No.: Q1485 SAS No.: Q1485 SDG NO.: Q1485

Continuing Calib Date: 03/05/2025 Initial Calibration Date(s): 02/24/2025 02/24/2025

Continuing Calib Time: 18:30 Initial Calibration Time(s): 14:59 22:17

GC Column: ZB-MR2 ID: 0.32 (mm)

| COMPOUND             |     | CCAL<br>RT | AVG<br>RT | RT WINDOW |      | DIFF<br>RT |
|----------------------|-----|------------|-----------|-----------|------|------------|
|                      |     |            |           | FROM      | TO   |            |
| Aroclor-1242-1       | (1) | 4.92       | 4.92      | 4.82      | 5.02 | 0.00       |
| Aroclor-1242-2       | (2) | 4.94       | 4.94      | 4.84      | 5.04 | 0.00       |
| Aroclor-1242-3       | (3) | 5.11       | 5.12      | 5.02      | 5.22 | 0.01       |
| Aroclor-1242-4       | (4) | 5.20       | 5.20      | 5.10      | 5.30 | 0.00       |
| Aroclor-1242-5       | (5) | 5.72       | 5.73      | 5.63      | 5.83 | 0.01       |
| Tetrachloro-m-xylene |     | 3.83       | 3.83      | 3.73      | 3.93 | 0.00       |
| Decachlorobiphenyl   |     | 8.88       | 8.89      | 8.79      | 8.99 | 0.01       |