

CALIBRATION VERIFICATION SUMMARY

Contract: JPCL01

Lab Code: CHEM Case No.: Q1485 SAS No.: Q1485 SDG NO.: Q1485

Continuing Calib Date: 03/05/2025 Initial Calibration Date(s): 02/24/2025 02/24/2025

Continuing Calib Time: 23:23 Initial Calibration Time(s): 14:59 22:17

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
Aroclor-1248-1 (1)	5.68	5.68	5.58	5.78	0.00
Aroclor-1248-2 (2)	5.95	5.95	5.85	6.05	0.00
Aroclor-1248-3 (3)	6.15	6.15	6.05	6.25	0.00
Aroclor-1248-4 (4)	6.55	6.55	6.45	6.65	0.00
Aroclor-1248-5 (5)	6.59	6.59	6.49	6.69	0.00
Tetrachloro-m-xylene	4.53	4.53	4.43	4.63	0.00
Decachlorobiphenyl	10.26	10.26	10.16	10.36	0.00

CALIBRATION VERIFICATION SUMMARY

Contract: JPCL01

Lab Code: CHEM Case No.: Q1485 SAS No.: Q1485 SDG NO.: Q1485

Continuing Calib Date: 03/05/2025 Initial Calibration Date(s): 02/24/2025 02/24/2025

Continuing Calib Time: 23:23 Initial Calibration Time(s): 14:59 22:17

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Aroclor-1248-1	(1)	4.92	4.92	4.82	5.02	0.00
Aroclor-1248-2	(2)	5.16	5.16	5.06	5.26	0.00
Aroclor-1248-3	(3)	5.20	5.20	5.10	5.30	0.00
Aroclor-1248-4	(4)	5.37	5.37	5.27	5.47	0.00
Aroclor-1248-5	(5)	5.77	5.77	5.67	5.87	0.01
Tetrachloro-m-xylene		3.83	3.83	3.73	3.93	0.00
Decachlorobiphenyl		8.89	8.89	8.79	8.99	0.00