

Instrument ID: ECD_P

Daily Analysis Runlog For Sequence/QC Batch ID # PP022425

Review By	yogesh	Review On	2/25/2025 9:00:36 AM
Supervise By	Ankita	Supervise On	2/25/2025 9:48:29 AM
SubDirectory	PP022425	HP Acquire Method	HP Processing Method PP022425
STD. NAME	STD REF.#		
Tune/Reschk			
Initial Calibration Stds	PP23735,PP23736,PP23737,PP23738,PP23739,PP23740,PP23741,PP23742,PP23743,PP23744,PP23745,PP23747,PP23748,PP23749,PP23750,PP23751,PP23752,PP23753,PP23754,PP23755,PP23756,PP23757,PP23758,PP23759,PP23760,PP23761,PP23762,PP23763,PP23764,PP23765,PP23766,PP23767,PP23768,PP23769,PP23770,PP23771,PP23772,PP23773,PP23774,PP23775		
CCC	PP23737,PP23742,PP23749,PP23754,PP23758,PP23763,PP23768,PP23773		
Internal Standard/PEM			
ICV/I.BLK	PP23778,PP23780,PP23783,PP23784,PP23786,PP23788,PP23790,PP23947		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

Sr#	SampleId	Data File Name	Date-Time	Operator	Status
1	HEXANE	PP069994.D	24 Feb 2025 14:27	YP\AJ	Ok
2	I.BLK	PP069995.D	24 Feb 2025 14:43	YP\AJ	Ok
3	AR1660ICC1000	PP069996.D	24 Feb 2025 14:59	YP\AJ	Ok
4	AR1660ICC750	PP069997.D	24 Feb 2025 15:15	YP\AJ	Ok
5	AR1660ICC500	PP069998.D	24 Feb 2025 15:32	YP\AJ	Ok
6	AR1660ICC250	PP069999.D	24 Feb 2025 15:48	YP\AJ	Ok
7	AR1660ICC050	PP070000.D	24 Feb 2025 16:04	YP\AJ	Ok,M
8	AR1221ICC500	PP070001.D	24 Feb 2025 16:20	YP\AJ	Ok,M
9	AR1232ICC500	PP070002.D	24 Feb 2025 16:37	YP\AJ	Ok
10	AR1242ICC1000	PP070003.D	24 Feb 2025 16:53	YP\AJ	Ok
11	AR1242ICC750	PP070004.D	24 Feb 2025 17:09	YP\AJ	Ok,M
12	AR1242ICC500	PP070005.D	24 Feb 2025 17:25	YP\AJ	Ok
13	AR1242ICC250	PP070006.D	24 Feb 2025 17:42	YP\AJ	Ok,M
14	AR1242ICC050	PP070007.D	24 Feb 2025 17:58	YP\AJ	Ok,M
15	AR1248ICC1000	PP070008.D	24 Feb 2025 18:14	YP\AJ	Ok
16	AR1248ICC750	PP070009.D	24 Feb 2025 18:30	YP\AJ	Ok
17	AR1248ICC500	PP070010.D	24 Feb 2025 18:46	YP\AJ	Ok
18	AR1248ICC250	PP070011.D	24 Feb 2025 19:03	YP\AJ	Ok
19	AR1248ICC050	PP070012.D	24 Feb 2025 19:19	YP\AJ	Ok,M
20	AR1254ICC1000	PP070013.D	24 Feb 2025 19:35	YP\AJ	Ok,M
21	AR1254ICC750	PP070014.D	24 Feb 2025 19:51	YP\AJ	Ok,M

Instrument ID: ECD_P

Daily Analysis Runlog For Sequence/QC Batch ID # PP022425

Review By	yogesh	Review On	2/25/2025 9:00:36 AM
Supervise By	Ankita	Supervise On	2/25/2025 9:48:29 AM
SubDirectory	PP022425	HP Acquire Method	HP Processing Method PP022425
STD. NAME	STD REF.#		
Tune/Reschk			
Initial Calibration Stds	PP23735,PP23736,PP23737,PP23738,PP23739,PP23740,PP23741,PP23742,PP23743,PP23744,PP23745,PP23747,PP23748,PP23749,PP23750,PP23751,PP23752,PP23753,PP23754,PP23755,PP23756,PP23757,PP23758,PP23759,PP23760,PP23761,PP23762,PP23763,PP23764,PP23765,PP23766,PP23767,PP23768,PP23769,PP23770,PP23771,PP23772,PP23773,PP23774,PP23775		
CCC	PP23737,PP23742,PP23749,PP23754,PP23758,PP23763,PP23768,PP23773		
Internal Standard/PEM			
ICV/I.BLK	PP23778,PP23780,PP23783,PP23784,PP23786,PP23788,PP23790,PP23947		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

22	AR1254ICC500	PP070015.D	24 Feb 2025 20:08	YP\AJ	Ok
23	AR1254ICC250	PP070016.D	24 Feb 2025 20:24	YP\AJ	Ok
24	AR1254ICC050	PP070017.D	24 Feb 2025 20:40	YP\AJ	Ok,M
25	AR1262ICC500	PP070018.D	24 Feb 2025 20:56	YP\AJ	Ok
26	AR1268ICC1000	PP070019.D	24 Feb 2025 21:12	YP\AJ	Ok
27	AR1268ICC750	PP070020.D	24 Feb 2025 21:29	YP\AJ	Ok,M
28	AR1268ICC500	PP070021.D	24 Feb 2025 21:45	YP\AJ	Ok
29	AR1268ICC250	PP070022.D	24 Feb 2025 22:01	YP\AJ	Ok,M
30	AR1268ICC050	PP070023.D	24 Feb 2025 22:17	YP\AJ	Ok,M
31	PP022425ICV500	PP070024.D	24 Feb 2025 22:34	YP\AJ	Ok
32	AR1242ICV500	PP070025.D	24 Feb 2025 22:50	YP\AJ	Ok
33	AR1248ICV500	PP070026.D	24 Feb 2025 23:06	YP\AJ	Ok
34	AR1254ICV500	PP070027.D	24 Feb 2025 23:22	YP\AJ	Ok
35	AR1268ICV500	PP070028.D	24 Feb 2025 23:38	YP\AJ	Ok

M : Manual Integration

Instrument ID: ECD_P

Daily Analysis Runlog For Sequence/QC Batch ID # PP030525

Review By	yogesh	Review On	3/5/2025 11:09:53 AM
Supervise By		Supervise On	
SubDirectory	PP030525	HP Acquire Method	HP Processing Method PP022425
STD. NAME	STD REF.#		
Tune/Reschk			
Initial Calibration Stds	PP23735,PP23736,PP23737,PP23738,PP23739,PP23740,PP23741,PP23742,PP23743,PP23744,PP23745,PP23747,PP23748,PP23749,PP23750,PP23751,PP23752,PP23753,PP23754,PP23755,PP23756,PP23757,PP23758,PP23759,PP23760,PP23761,PP23762,PP23763,PP23764,PP23765,PP23766,PP23767,PP23768,PP23769,PP23770,PP23771,PP23772,PP23773,PP23774,PP23775		
CCC	PP23737,PP23742,PP23749,PP23754,PP23758,PP23763,PP23768,PP23773		
Internal Standard/PEM			
ICV/I.BLK	PP23778,PP23780,PP23783,PP23784,PP23786,PP23788,PP23790,PP23947		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

Sr#	SampleId	Data File Name	Date-Time	Operator	Status
1	HEXANE	PP070250.D	05 Mar 2025 08:31	YP\AJ	Ok
2	AR1660CCC500	PP070251.D	05 Mar 2025 08:47	YP\AJ	Ok
3	AR1242CCC500	PP070252.D	05 Mar 2025 09:04	YP\AJ	Ok
4	AR1248CCC500	PP070253.D	05 Mar 2025 09:20	YP\AJ	Ok
5	AR1254CCC500	PP070254.D	05 Mar 2025 09:36	YP\AJ	Ok,NS
6	I.BLK	PP070255.D	05 Mar 2025 09:52	YP\AJ	Ok
7	Q1476-01	PP070256.D	05 Mar 2025 10:09	YP\AJ	ReRun
8	Q1476-01RE	PP070257.D	05 Mar 2025 10:25	YP\AJ	Confirms
9	Q1473-02RE	PP070258.D	05 Mar 2025 11:00	YP\AJ	Confirms
10	AR1660CCC500	PP070259.D	05 Mar 2025 11:49	YP\AJ	Ok
11	AR1242CCC500	PP070260.D	05 Mar 2025 12:05	YP\AJ	Ok,NS
12	AR1248CCC500	PP070261.D	05 Mar 2025 12:21	YP\AJ	Ok,NS
13	AR1254CCC500	PP070262.D	05 Mar 2025 12:38	YP\AJ	Ok,NS
14	I.BLK	PP070263.D	05 Mar 2025 12:54	YP\AJ	Ok
15	PB166985BL	PP070264.D	05 Mar 2025 14:58	YP\AJ	Ok
16	PB166985BS	PP070265.D	05 Mar 2025 15:14	YP\AJ	Ok
17	Q1481-01	PP070266.D	05 Mar 2025 15:31	YP\AJ	Ok
18	Q1482-01	PP070267.D	05 Mar 2025 15:47	YP\AJ	Ok,NS
19	Q1484-01	PP070268.D	05 Mar 2025 16:03	YP\AJ	Ok
20	Q1485-01	PP070269.D	05 Mar 2025 16:19	YP\AJ	Ok
21	Q1486-01	PP070270.D	05 Mar 2025 16:36	YP\AJ	Ok

Instrument ID: ECD_P

Daily Analysis Runlog For Sequence/QC Batch ID # PP030525

Review By	yogesh	Review On	3/5/2025 11:09:53 AM
Supervise By		Supervise On	
SubDirectory	PP030525	HP Acquire Method	HP Processing Method PP022425
STD. NAME	STD REF.#		
Tune/Reschk			
Initial Calibration Stds	PP23735,PP23736,PP23737,PP23738,PP23739,PP23740,PP23741,PP23742,PP23743,PP23744,PP23745,PP23747,PP23748,PP23749,PP23750,PP23751,PP23752,PP23753,PP23754,PP23755,PP23756,PP23757,PP23758,PP23759,PP23760,PP23761,PP23762,PP23763,PP23764,PP23765,PP23766,PP23767,PP23768,PP23769,PP23770,PP23771,PP23772,PP23773,PP23774,PP23775		
CCC	PP23737,PP23742,PP23749,PP23754,PP23758,PP23763,PP23768,PP23773		
Internal Standard/PEM			
ICV/I.BLK	PP23778,PP23780,PP23783,PP23784,PP23786,PP23788,PP23790,PP23947		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

22	Q1487-01	PP070271.D	05 Mar 2025 16:52	YPIAJ	Ok
23	Q1488-01	PP070272.D	05 Mar 2025 17:08	YPIAJ	Ok
24	Q1488-03	PP070273.D	05 Mar 2025 17:25	YPIAJ	Ok
25	AR1660CCC500	PP070274.D	05 Mar 2025 18:13	YPIAJ	Ok
26	AR1242CCC500	PP070275.D	05 Mar 2025 18:30	YPIAJ	Ok,NS
27	AR1248CCC500	PP070276.D	05 Mar 2025 18:46	YPIAJ	Ok,NS
28	AR1254CCC500	PP070277.D	05 Mar 2025 19:02	YPIAJ	Ok,NS
29	I.BLK	PP070278.D	05 Mar 2025 19:19	YPIAJ	Ok
30	Q1488-03MS	PP070279.D	05 Mar 2025 19:35	YPIAJ	Ok
31	Q1488-03MSD	PP070280.D	05 Mar 2025 19:51	YPIAJ	Ok,NS
32	Q1488-05	PP070281.D	05 Mar 2025 20:08	YPIAJ	Ok
33	Q1488-07	PP070282.D	05 Mar 2025 20:24	YPIAJ	Ok
34	Q1488-09	PP070283.D	05 Mar 2025 20:40	YPIAJ	Ok
35	Q1488-11	PP070284.D	05 Mar 2025 20:56	YPIAJ	Ok
36	Q1489-01	PP070285.D	05 Mar 2025 21:13	YPIAJ	Ok
37	Q1489-05	PP070286.D	05 Mar 2025 21:29	YPIAJ	Ok
38	Q1489-09	PP070287.D	05 Mar 2025 21:45	YPIAJ	Ok
39	Q1489-13	PP070288.D	05 Mar 2025 22:02	YPIAJ	Ok
40	AR1660CCC500	PP070289.D	05 Mar 2025 22:50	YPIAJ	Ok
41	AR1242CCC500	PP070290.D	05 Mar 2025 23:07	YPIAJ	Ok,NS
42	AR1248CCC500	PP070291.D	05 Mar 2025 23:23	YPIAJ	Ok,NS
43	AR1254CCC500	PP070292.D	05 Mar 2025 23:39	YPIAJ	Ok,NS
44	I.BLK	PP070293.D	05 Mar 2025 23:56	YPIAJ	Ok

Instrument ID: ECD_P

Daily Analysis Runlog For Sequence/QC Batch ID # PP030525

Review By	yogesh	Review On	3/5/2025 11:09:53 AM
Supervise By		Supervise On	
SubDirectory	PP030525	HP Acquire Method	HP Processing Method PP022425
STD. NAME	STD REF.#		
Tune/Reschk			
Initial Calibration Stds	PP23735,PP23736,PP23737,PP23738,PP23739,PP23740,PP23741,PP23742,PP23743,PP23744,PP23745,PP23747,PP23748,PP23749,PP23750,PP23751,PP23752,PP23753,PP23754,PP23755,PP23756,PP23757,PP23758,PP23759,PP23760,PP23761,PP23762,PP23763,PP23764,PP23765,PP23766,PP23767,PP23768,PP23769,PP23770,PP23771,PP23772,PP23773,PP23774,PP23775		
CCC	PP23737,PP23742,PP23749,PP23754,PP23758,PP23763,PP23768,PP23773		
Internal Standard/PEM			
ICV/I.BLK	PP23778,PP23780,PP23783,PP23784,PP23786,PP23788,PP23790,PP23947		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

M : Manual Integration