

PESTICIDE CALIBRATION VERIFICATION SUMMARY

Contract: JPCL01

Lab Code: CHEM Case No.: Q1485 SAS No.: Q1485 SDG NO.: Q1485

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 02/21/2025 02/21/2025

Client Sample No. (PEM): PEM - PL094324.D Date Analyzed: 02/21/2025

Lab Sample No.(PEM): PEM Time Analyzed: 10:29

PEM COMPOUND	RT	RT WINDOW		CALC	NOM	%D
		FROM	TO	AMOUNT(ng)	AMOUNT(ng)	
Decachlorobiphenyl	9.049	8.950	9.150	19.210	20.000	-4.0
Tetrachloro-m-xylene	3.536	3.490	3.590	18.850	20.000	-5.8
alpha-BHC	3.991	3.940	4.040	10.030	10.000	0.3
beta-BHC	4.523	4.470	4.570	10.250	10.000	2.5
gamma-BHC (Lindane)	4.324	4.270	4.370	9.920	10.000	-0.8
Endrin	6.570	6.500	6.640	41.860	50.000	-16.3
4,4'-DDT	7.020	6.950	7.090	87.790	100.000	-12.2
Methoxychlor	7.497	7.430	7.570	207.340	250.000	-17.1

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 02/21/2025 02/21/2025

Client Sample No. (PEM): PEM - PL094324.D Date Analyzed: 02/21/2025

Lab Sample No.(PEM): PEM Time Analyzed: 10:29

PEM COMPOUND	RT	RT WINDOW		CALC	NOM	%D
		FROM	TO	AMOUNT(ng)	AMOUNT(ng)	
Decachlorobiphenyl	7.904	7.800	8.000	18.560	20.000	-7.2
Tetrachloro-m-xylene	2.771	2.720	2.820	18.550	20.000	-7.3
alpha-BHC	3.273	3.220	3.320	8.830	10.000	-11.7
beta-BHC	3.903	3.850	3.950	9.960	10.000	-0.4
gamma-BHC (Lindane)	3.603	3.550	3.650	8.530	10.000	-14.7
Endrin	5.632	5.560	5.700	45.740	50.000	-8.5
4,4'-DDT	6.029	5.960	6.100	98.350	100.000	-1.7
Methoxychlor	6.605	6.530	6.680	222.980	250.000	-10.8