

7C

SEMIVOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: JPCI01
 Lab Code: CHEM Case No.: Q1485 SAS No.: Q1485 SDG No.: Q1485
 Instrument ID: BNA_F Calibration Date/Time: 03/05/2025 10:09
 Lab File ID: BF141835.D Init. Calib. Date(s): 02/27/2025 02/27/2025
 EPA Sample No.: SSTDCCC040 Init. Calib. Time(s): 15:17 18:45
 GC Column: DB-UI ID: 0.18 (mm)

COMPOUND	RRF	RRF040	MIN RRF	%D	MAX%D
2-Fluorophenol	1.252	1.263		0.9	
Benzaldehyde	0.948	0.834		-12.0	
Phenol-d6	1.630	1.613		-1.0	
Phenol	1.742	1.734		-0.5	20.0
bis(2-Chloroethyl)ether	1.328	1.232		-7.2	
2-Chlorophenol	1.355	1.354		-0.1	
2-Methylphenol	1.118	1.094		-2.1	
2,2-oxybis(1-Chloropropane)	2.016	1.771		-12.2	
Acetophenone	0.487	0.480		-1.4	
3+4-Methylphenols	1.392	1.384		-0.6	
n-Nitroso-di-n-propylamine	1.072	0.955	0.050	-10.9	
Nitrobenzene-d5	0.313	0.371		18.5	
Hexachloroethane	0.547	0.522		-4.6	
Nitrobenzene	0.335	0.375		11.9	
Isophorone	0.672	0.633		-5.8	
2-Nitrophenol	0.127	0.179		40.9	20.0
2,4-Dimethylphenol	0.247	0.248		0.4	
bis(2-Chloroethoxy)methane	0.429	0.395		-7.9	
2,4-Dichlorophenol	0.285	0.291		2.1	20.0
Naphthalene	1.028	1.017		-1.1	
4-Chloroaniline	0.377	0.378		0.3	
Hexachlorobutadiene	0.194	0.187		-3.6	20.0
Caprolactam	0.088	0.098		11.4	
4-Chloro-3-methylphenol	0.317	0.324		2.2	20.0
2-Methylnaphthalene	0.674	0.657		-2.5	
Hexachlorocyclopentadiene	0.241	0.235	0.050	-2.5	
2,4,6-Trichlorophenol	0.373	0.395		5.9	20.0
2-Fluorobiphenyl	1.247	1.230		-1.4	
2,4,5-Trichlorophenol	0.369	0.412		11.7	
1,1-Biphenyl	1.522	1.505		-1.1	
2-Chloronaphthalene	1.124	1.144		1.8	
2-Nitroaniline	0.260	0.374		43.8	
Dimethylphthalate	1.339	1.326		-1.0	
Acenaphthylene	1.685	1.726		2.4	
2,6-Dinitrotoluene	0.226	0.297		31.4	
3-Nitroaniline	0.241	0.328		36.1	
Acenaphthene	1.137	1.188		4.5	20.0
2,4-Dinitrophenol	0.089	0.149	0.050	67.4	
4-Nitrophenol	0.198	0.283	0.050	42.9	

7C

SEMIVOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: JPCI01
 Lab Code: CHEM Case No.: Q1485 SAS No.: Q1485 SDG No.: Q1485
 Instrument ID: BNA_F Calibration Date/Time: 03/05/2025 10:09
 Lab File ID: BF141835.D Init. Calib. Date(s): 02/27/2025 02/27/2025
 EPA Sample No.: SSTDCCC040 Init. Calib. Time(s): 15:17 18:45
 GC Column: DB-UI ID: 0.18 (mm)

COMPOUND	RRF	RRF040	MIN RRF	%D	MAX%D
Dibenzofuran	1.654	1.670		1.0	
2,4-Dinitrotoluene	0.269	0.392		45.7	
Diethylphthalate	1.343	1.320		-1.7	
4-Chlorophenyl-phenylether	0.621	0.606		-2.4	
Fluorene	1.239	1.253		1.1	
4-Nitroaniline	0.229	0.330		44.1	
4,6-Dinitro-2-methylphenol	0.081	0.122		50.6	
n-Nitrosodiphenylamine	0.657	0.638		-2.9	20.0
2,4,6-Tribromophenol	0.188	0.210		11.7	
4-Bromophenyl-phenylether	0.234	0.223		-4.7	
Hexachlorobenzene	0.247	0.245		-0.8	
Atrazine	0.181	0.156		-13.8	
Pentachlorophenol	0.156	0.176		12.8	20.0
Phenanthrene	1.070	1.082		1.1	
Anthracene	1.072	1.088		1.5	
Carbazole	0.947	0.990		4.5	
Di-n-butylphthalate	1.200	1.131		-5.8	
Fluoranthene	1.096	1.157		5.6	20.0
Pyrene	1.732	1.722		-0.6	
Terphenyl-d14	1.236	1.190		-3.7	
Butylbenzylphthalate	0.629	0.646		2.7	
3,3-Dichlorobenzidine	0.379	0.382		0.8	
Benzo (a) anthracene	1.321	1.308		-1.0	
Chrysene	1.218	1.190		-2.3	
Bis (2-ethylhexyl) phthalate	0.783	0.807		3.1	
Di-n-octyl phthalate	1.173	1.132		-3.5	20.0
Benzo (b) fluoranthene	1.362	1.368		0.4	
Benzo (k) fluoranthene	1.158	1.057		-8.7	
Benzo (a) pyrene	1.093	1.074		-1.7	20.0
Indeno (1,2,3-cd) pyrene	1.274	1.310		2.8	
Dibenzo (a,h) anthracene	1.040	1.067		2.6	
Benzo (g,h,i) perylene	1.065	1.105		3.8	
1,2,4,5-Tetrachlorobenzene	0.578	0.577		-0.2	
1,4-Dioxane	0.568	0.571		0.5	20.0
2,3,4,6-Tetrachlorophenol	0.321	0.351		9.3	

All other compounds must meet a minimum RRF of 0.010.