

7C

SEMIVOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: JPC101

Lab Code: CHEM Case No.: Q1485 SAS No.: Q1485 SDG No.: Q1485

Instrument ID: BNA_F Calibration Date/Time: 03/06/2025 09:52

Lab File ID: BF141858.D Init. Calib. Date(s): 02/27/2025 02/27/2025

EPA Sample No.: SSTDCCC040 Init. Calib. Time(s): 15:17 18:45

GC Column: DB-UI ID: 0.18 (mm)

COMPOUND	RRF	RRF040	MIN RRF	%D	MAX%D
2-Fluorophenol	1.252	1.263		0.9	
Benzaldehyde	0.948	0.847		-10.7	
Phenol-d6	1.630	1.584		-2.8	
Phenol	1.742	1.692		-2.9	20.0
bis(2-Chloroethyl)ether	1.328	1.246		-6.2	
2-Chlorophenol	1.355	1.366		0.8	
2-Methylphenol	1.118	1.076		-3.8	
2,2-oxybis(1-Chloropropane)	2.016	1.664		-17.5	
Acetophenone	0.487	0.490		0.6	
3+4-Methylphenols	1.392	1.366		-1.9	
n-Nitroso-di-n-propylamine	1.072	0.946	0.050	-11.8	
Nitrobenzene-d5	0.313	0.382		22.0	
Hexachloroethane	0.547	0.554		1.3	
Nitrobenzene	0.335	0.382		14.0	
Isophorone	0.672	0.635		-5.5	
2-Nitrophenol	0.127	0.188		48.0	20.0
2,4-Dimethylphenol	0.247	0.253		2.4	
bis(2-Chloroethoxy)methane	0.429	0.412		-4.0	
2,4-Dichlorophenol	0.285	0.305		7.0	20.0
Naphthalene	1.028	1.054		2.5	
4-Chloroaniline	0.377	0.374		-0.8	
Hexachlorobutadiene	0.194	0.208		7.2	20.0
Caprolactam	0.088	0.093		5.7	
4-Chloro-3-methylphenol	0.317	0.330		4.1	20.0
2-Methylnaphthalene	0.674	0.684		1.5	
Hexachlorocyclopentadiene	0.241	0.266	0.050	10.4	
2,4,6-Trichlorophenol	0.373	0.399		7.0	20.0
2-Fluorobiphenyl	1.247	1.292		3.6	
2,4,5-Trichlorophenol	0.369	0.413		11.9	
1,1-Biphenyl	1.522	1.560		2.5	
2-Chloronaphthalene	1.124	1.157		2.9	
2-Nitroaniline	0.260	0.356		36.9	
Dimethylphthalate	1.339	1.363		1.8	
Acenaphthylene	1.685	1.735		3.0	
2,6-Dinitrotoluene	0.226	0.300		32.7	
3-Nitroaniline	0.241	0.314		30.3	
Acenaphthene	1.137	1.209		6.3	20.0
2,4-Dinitrophenol	0.089	0.159	0.050	78.7	
4-Nitrophenol	0.198	0.262	0.050	32.3	

7C

SEMIVOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: JPCL01

Lab Code: CHEM Case No.: Q1485 SAS No.: Q1485 SDG No.: Q1485

Instrument ID: BNA_F Calibration Date/Time: 03/06/2025 09:52

Lab File ID: BF141858.D Init. Calib. Date(s): 02/27/2025 02/27/2025

EPA Sample No.: SSTDCCC040 Init. Calib. Time(s): 15:17 18:45

GC Column: DB-UI ID: 0.18 (mm)

COMPOUND	RRF	RRF040	MIN RRF	%D	MAX%D
Dibenzofuran	1.654	1.682		1.7	
2,4-Dinitrotoluene	0.269	0.396		47.2	
Diethylphthalate	1.343	1.344		0.1	
4-Chlorophenyl-phenylether	0.621	0.640		3.1	
Fluorene	1.239	1.270		2.5	
4-Nitroaniline	0.229	0.307		34.1	
4,6-Dinitro-2-methylphenol	0.081	0.132		63.0	
n-Nitrosodiphenylamine	0.657	0.670		2.0	20.0
2,4,6-Tribromophenol	0.188	0.224		19.1	
4-Bromophenyl-phenylether	0.234	0.243		3.8	
Hexachlorobenzene	0.247	0.266		7.7	
Atrazine	0.181	0.155		-14.4	
Pentachlorophenol	0.156	0.186		19.2	20.0
Phenanthrene	1.070	1.115		4.2	
Anthracene	1.072	1.115		4.0	
Carbazole	0.947	0.988		4.3	
Di-n-butylphthalate	1.200	1.188		-1.0	
Fluoranthene	1.096	1.142		4.2	20.0
Pyrene	1.732	1.820		5.1	
Terphenyl-d14	1.236	1.301		5.3	
Butylbenzylphthalate	0.629	0.650		3.3	
3,3-Dichlorobenzidine	0.379	0.405		6.9	
Benzo (a) anthracene	1.321	1.360		3.0	
Chrysene	1.218	1.218		0.0	
Bis(2-ethylhexyl)phthalate	0.783	0.799		2.0	
Di-n-octyl phthalate	1.173	1.163		-0.9	20.0
Benzo (b) fluoranthene	1.362	1.388		1.9	
Benzo (k) fluoranthene	1.158	1.073		-7.3	
Benzo (a) pyrene	1.093	1.113		1.8	20.0
Indeno (1,2,3-cd) pyrene	1.274	1.507		18.3	
Dibenzo (a,h) anthracene	1.040	1.220		17.3	
Benzo (g,h,i) perylene	1.065	1.282		20.4	
1,2,4,5-Tetrachlorobenzene	0.578	0.607		5.0	
1,4-Dioxane	0.568	0.556		-2.1	20.0
2,3,4,6-Tetrachlorophenol	0.321	0.362		12.8	

All other compounds must meet a minimum RRF of 0.010.