


LB 134903

7471B

INSTRUMENT ID: CV2

Linear



A= 0.0000e+000

B= 1.8483e-004

C= 1.8785e-003

Rho= 0.9995544

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	9/10
0.0	0.000	0.006	0.006	23	0.000	23					0/10
0.2	0.200	0.215	0.015	1154	0.0 %	1154					1
2.5	2.500	2.389	-0.111	12914	0.0 %	12914					2
5.0	5.000	5.043	0.043	27275	0.0 %	27275					4
7.5	7.500	7.692	0.192	41609	0.0 %	41609					1
10.0	10.000	9.854	-0.146	53305	0.0 %	53305					3

LB134903 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 05 Mar 2025 13:38:27

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	23	-	0.0000	7471B	PPB	05 Mar 2025 13:42:10	S	Std
0.2 - 1	50.2	1154	-	0.2000	7471B	PPB	05 Mar 2025 13:44:27	S	Std
2.5 - 1	52.5	12914	-	2.5000	7471B	PPB	05 Mar 2025 13:46:44	S	Std
5.0 - 1	55	27275	-	5.0000	7471B	PPB	05 Mar 2025 13:49:01	S	Std
7.5 - 1	57.5	41609	-	7.5000	7471B	PPB	05 Mar 2025 13:51:20	S	Std
10.0 - 1	57.5	53305	-	10.0000	7471B	PPB	05 Mar 2025 13:56:25	S	Std
ICV32 - 1	ICV32	19726	3.6478	-	7471B	PPB	05 Mar 2025 13:59:13	U	SMPL
ICB32 - 1	ICB32	-366	-0.0658	-	7471B	PPB	05 Mar 2025 14:01:29	U	SMPL
CCV65 - 1	CCV65	27129	5.0162	-	7471B	PPB	05 Mar 2025 14:03:47	U	SMPL
CCB65 - 1	CCB65	-396	-0.0713	-	7471B	PPB	05 Mar 2025 14:06:02	U	SMPL
CRA - 1	CRA	1104	0.2059	-	7471B	PPB	05 Mar 2025 14:08:21	U	SMPL
HighStd - 1	HighStd	53632	9.9147	-	7471B	PPB	05 Mar 2025 14:10:36	U	SMPL
ChkStd - 1	ChkStd	38520	7.1216	-	7471B	PPB	05 Mar 2025 14:12:51	U	SMPL
PB166988BL - 1	PBS	-458	-0.0828	-	7471B	PPB	05 Mar 2025 14:19:38	U	SMPL
PB166988BS - 1	LCSS	21257	3.9308	-	7471B	PPB	05 Mar 2025 14:21:56	U	SMPL
Q1474-01 - 1	BU-03-02282025	1910	0.3549	-	7471B	PPB	05 Mar 2025 14:29:32	U	SMPL
Q1474-01DUP - 1	BU-03-02282025DUP	1759	0.3270	-	7471B	PPB	05 Mar 2025 14:31:48	U	SMPL
Q1474-01MS - 1	BU-03-02282025MS	22644	4.1872	-	7471B	PPB	05 Mar 2025 14:34:05	U	SMPL
Q1474-01MSD - 1	BU-03-02282025MSD	25538	4.7221	-	7471B	PPB	05 Mar 2025 14:36:22	U	SMPL
Q1475-01 - 1	TR-04-02282025	649	0.1218	-	7471B	PPB	05 Mar 2025 14:38:40	U	SMPL
CCV66 - 1	CCV66	26624	4.9228	-	7471B	PPB	05 Mar 2025 14:40:58	U	SMPL
CCB66 - 1	CCB66	-379	-0.0682	-	7471B	PPB	05 Mar 2025 14:43:14	U	SMPL
Q1480-01 - 1	TP-1	6358	1.1770	-	7471B	PPB	05 Mar 2025 14:45:32	U	SMPL
Q1480-09 - 1	TP-2	3099	0.5747	-	7471B	PPB	05 Mar 2025 14:47:48	U	SMPL
Q1480-17 - 1	TP-3	9255	1.7125	-	7471B	PPB	05 Mar 2025 14:50:05	U	SMPL
Q1480-25 - 1	TP-4	2949	0.5469	-	7471B	PPB	05 Mar 2025 14:52:22	U	SMPL
Q1481-01 - 1	WC-K1311	3083	0.5717	-	7471B	PPB	05 Mar 2025 14:54:40	U	SMPL
Q1482-01 - 1	OR-03-030425	1514	0.2817	-	7471B	PPB	05 Mar 2025 14:56:57	U	SMPL
Q1484-01 - 1	TR-OTR-COMP-01	2175	0.4039	-	7471B	PPB	05 Mar 2025 14:59:15	U	SMPL
Q1485-01 - 1	DN-B-41	2573	0.4774	-	7471B	PPB	05 Mar 2025 15:01:31	U	SMPL
Q1486-01 - 1	DN-B-40	1222	0.2277	-	7471B	PPB	05 Mar 2025 15:03:48	U	SMPL
Q1487-01 - 1	DN-B-42	2458	0.4562	-	7471B	PPB	05 Mar 2025 15:06:03	U	SMPL
CCV67 - 1	CCV67	26873	4.9688	-	7471B	PPB	05 Mar 2025 15:08:19	U	SMPL
CCB67 - 1	CCB67	-313	-0.0560	-	7471B	PPB	05 Mar 2025 15:10:37	U	SMPL
Q1489-01 - 1	TP-1	5790	1.0720	-	7471B	PPB	05 Mar 2025 15:12:56	U	SMPL
Q1489-05 - 1	TP-2	80529	14.8861	-	7471B	PPB	05 Mar 2025 15:15:12	U	SMPL
Q1489-09 - 1	TP-3	14240	2.6339	-	7471B	PPB	05 Mar 2025 15:17:29	U	SMPL
Q1489-13 - 1	TP-4	1470	0.2736	-	7471B	PPB	05 Mar 2025 15:20:00	U	SMPL
Q1474-01LX5 - 1		246	0.0473	-	7471B	PPB	05 Mar 2025 15:22:19	U	SMPL
Q1474-01A - 1		24065	4.4498	-	7471B	PPB	05 Mar 2025 15:24:35	U	SMPL
Q1489-05DLX5 - 1		17217	3.1841	-	7471B	PPB	05 Mar 2025 15:26:52	U	SMPL
CCV68 - 1	CCV68	26644	4.9265	-	7471B	PPB	05 Mar 2025 15:29:11	U	SMPL
CCB68 - 1	CCB68	-278	-0.0495	-	7471B	PPB	05 Mar 2025 15:31:33	U	SMPL