

CALIBRATION VERIFICATION SUMMARY

Contract: JPCL01

Lab Code: CHEM Case No.: Q1486 SAS No.: Q1486 SDG NO.: Q1486

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 02/24/2025 02/24/2025

Client Sample No.: CCAL01 Date Analyzed: 03/05/2025

Lab Sample No.: AR1660CCC500 Data File : PP070259.D Time Analyzed: 11:49

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	5.679	5.581	5.781	474.860	500.000	-5.0
Aroclor-1016-2	5.700	5.602	5.802	491.780	500.000	-1.6
Aroclor-1016-3	5.763	5.665	5.865	489.540	500.000	-2.1
Aroclor-1016-4	5.860	5.762	5.962	496.730	500.000	-0.7
Aroclor-1016-5	6.153	6.056	6.256	483.820	500.000	-3.2
Aroclor-1260-1	7.273	7.175	7.375	482.320	500.000	-3.5
Aroclor-1260-2	7.527	7.429	7.629	455.430	500.000	-8.9
Aroclor-1260-3	7.885	7.788	7.988	488.800	500.000	-2.2
Aroclor-1260-4	8.110	8.012	8.212	479.200	500.000	-4.2
Aroclor-1260-5	8.430	8.333	8.533	484.960	500.000	-3.0
Decachlorobiphenyl	10.253	10.156	10.356	47.630	50.000	-4.7
Tetrachloro-m-xylene	4.526	4.427	4.627	52.370	50.000	4.7

CALIBRATION VERIFICATION SUMMARY

Contract: JPCL01

Lab Code: CHEM Case No.: Q1486 SAS No.: Q1486 SDG NO.: Q1486

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 02/24/2025 02/24/2025

Client Sample No.: CCAL01 Date Analyzed: 03/05/2025

Lab Sample No.: AR1660CCC500 Data File : PP070259.D Time Analyzed: 11:49

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.917	4.820	5.020	513.670	500.000	2.7
Aroclor-1016-2	4.936	4.839	5.039	509.760	500.000	2.0
Aroclor-1016-3	5.113	5.017	5.217	518.780	500.000	3.8
Aroclor-1016-4	5.154	5.058	5.258	507.120	500.000	1.4
Aroclor-1016-5	5.370	5.273	5.473	534.650	500.000	6.9
Aroclor-1260-1	6.407	6.312	6.512	493.320	500.000	-1.3
Aroclor-1260-2	6.596	6.500	6.700	502.660	500.000	0.5
Aroclor-1260-3	6.750	6.654	6.854	468.200	500.000	-6.4
Aroclor-1260-4	7.222	7.126	7.326	524.080	500.000	4.8
Aroclor-1260-5	7.463	7.367	7.567	533.970	500.000	6.8
Decachlorobiphenyl	8.882	8.788	8.988	48.820	50.000	-2.4
Tetrachloro-m-xylene	3.828	3.730	3.930	51.480	50.000	3.0