

LAB CHRONICLE

OrderID:	Q1486	OrderDate:	3/4/2025 3:32:00 PM
Client:	JPCL Engineering	Project:	NYSDOT Two Bronx River Parkway Bridges
Contact:	Paul Rotondi	Location:	I11

LabID	ClientID	Matrix	Test	Method	Sample Date	Prep Date	Anal Date	Received
Q1486-01	DN-B-40	SOIL	PCB	8082A	03/04/25	03/05/25	03/05/25	03/04/25



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Hit Summary Sheet
SW-846

SDG No.: Q1486

Order ID: Q1486

Client: JPCL Engineering

Project ID: NYSDOT Two Bronx River Parkway I

Sample ID	Client ID	Matrix	Parameter	Concentration	C	MDL	RDL	Units
-----------	-----------	--------	-----------	---------------	---	-----	-----	-------

Client ID :

Total Concentration: 0.000

A
B
C
D
E
F
G



SAMPLE DATA

Report of Analysis

Client:	JPCL Engineering	Date Collected:	03/04/25
Project:	NYSDOT Two Bronx River Parkway Bridges	Date Received:	03/04/25
Client Sample ID:	DN-B-40	SDG No.:	Q1486
Lab Sample ID:	Q1486-01	Matrix:	SOIL
Analytical Method:	SW8082A	% Solid:	93.6
Sample Wt/Vol:	30.05	Units:	g
Soil Aliquot Vol:			uL
Extraction Type:		Test:	PCB
GPC Factor :	1.0	PH :	
Prep Method :	SW3541B	Injection Volume :	

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PP070270.D	1	03/05/25 09:10	03/05/25 16:36	PB166985

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
12674-11-2	Aroclor-1016	18.1	U	3.60	18.1	ug/kg
11104-28-2	Aroclor-1221	18.1	U	6.80	18.1	ug/kg
11141-16-5	Aroclor-1232	18.1	U	3.60	18.1	ug/kg
53469-21-9	Aroclor-1242	18.1	U	3.60	18.1	ug/kg
12672-29-6	Aroclor-1248	18.1	U	8.40	18.1	ug/kg
11097-69-1	Aroclor-1254	18.1	U	2.90	18.1	ug/kg
37324-23-5	Aroclor-1262	18.1	U	4.90	18.1	ug/kg
11100-14-4	Aroclor-1268	18.1	U	3.70	18.1	ug/kg
11096-82-5	Aroclor-1260	18.1	U	3.10	18.1	ug/kg
SURROGATES						
877-09-8	Tetrachloro-m-xylene	25.1		32 - 144	126%	SPK: 20
2051-24-3	Decachlorobiphenyl	22.7		32 - 175	113%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit



QC SUMMARY

Surrogate Summary

SDG No.: Q1486

Client: JPCL Engineering

Analytical Method: 8082A

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Rec	Qual	Limits	
								Low	High
I.BLK-PP069995.D	PIBLK-PP069995.D	Tetrachloro-m-xylene	1	20	21.9	109		60	140
		Decachlorobiphenyl	1	20	21.8	109		60	140
		Tetrachloro-m-xylene	2	20	22.1	110		60	140
		Decachlorobiphenyl	2	20	21.6	108		60	140
I.BLK-PP070263.D	PIBLK-PP070263.D	Tetrachloro-m-xylene	1	20	20.1	100		60	140
		Decachlorobiphenyl	1	20	18.8	94		60	140
		Tetrachloro-m-xylene	2	20	20.9	105		60	140
		Decachlorobiphenyl	2	20	20.3	102		60	140
PB166985BL	PB166985BL	Tetrachloro-m-xylene	1	20	24.9	125		32	144
		Decachlorobiphenyl	1	20	22.2	111		32	175
		Tetrachloro-m-xylene	2	20	24.1	121		32	144
		Decachlorobiphenyl	2	20	23.8	119		32	175
PB166985BS	PB166985BS	Tetrachloro-m-xylene	1	20	25.4	127		32	144
		Decachlorobiphenyl	1	20	22.5	113		32	175
		Tetrachloro-m-xylene	2	20	24.1	121		32	144
		Decachlorobiphenyl	2	20	23.7	119		32	175
Q1486-01	DN-B-40	Tetrachloro-m-xylene	1	20	24.0	120		32	144
		Decachlorobiphenyl	1	20	22.7	113		32	175
		Tetrachloro-m-xylene	2	20	25.1	126		32	144
		Decachlorobiphenyl	2	20	22.2	111		32	175
I.BLK-PP070278.D	PIBLK-PP070278.D	Tetrachloro-m-xylene	1	20	21.2	106		60	140
		Decachlorobiphenyl	1	20	19.6	98		60	140
		Tetrachloro-m-xylene	2	20	20.7	104		60	140
		Decachlorobiphenyl	2	20	20.6	103		60	140
Q1488-03MS	ENV-101-SB02MS	Tetrachloro-m-xylene	1	20	24.9	125		32	144
		Decachlorobiphenyl	1	20	19.6	98		32	175
		Tetrachloro-m-xylene	2	20	23.8	119		32	144
		Decachlorobiphenyl	2	20	21.1	106		32	175
Q1488-03MSD	ENV-101-SB02MSD	Tetrachloro-m-xylene	1	20	26.5	132		32	144
		Decachlorobiphenyl	1	20	22.3	112		32	175
		Tetrachloro-m-xylene	2	20	24.6	123		32	144
		Decachlorobiphenyl	2	20	24.1	121		32	175
I.BLK-PP070293.D	PIBLK-PP070293.D	Tetrachloro-m-xylene	1	20	21.6	108		60	140
		Decachlorobiphenyl	1	20	20.2	101		60	140
		Tetrachloro-m-xylene	2	20	21.7	109		60	140
		Decachlorobiphenyl	2	20	21.4	107		60	140

Matrix Spike/Matrix Spike Duplicate Summary

SW-846

SDG No.: Q1486

Client: JPCL Engineering

Analytical Method: 8082A

DataFile : PP070279.D

Lab Sample ID:	Parameter	Spike	Sample Result	Result	Units	Rec	Rec Qual	RPD	RPD Qual	Low	Limits High	RPD
Client Sample ID:	ENV-101-SB02MS											
Q1488-03MS	AR1016	183.5	0	188	ug/kg	102				55	146	
	AR1260	183.5	0	174	ug/kg	95				31	146	

Matrix Spike/Matrix Spike Duplicate Summary

SW-846

SDG No.: Q1486

Client: JPCL Engineering

Analytical Method: 8082A

DataFile : PP070280.D

Lab Sample ID:		Parameter	Spike	Sample Result	Result	Units	Rec Qual	RPD Qual	Low	Limits High	RPD
Client Sample ID:		ENV-101-SB02MSD									
Q1488-03MSD		AR1016	183.7	0	193	ug/kg	105	3	55	146	20
		AR1260	183.7	0	183	ug/kg	100	5	31	146	20



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q1486

Client: JPCL Engineering

Analytical Method: 8082A Datafile : PP070265.D

Lab Sample ID	Parameter	Spike	Result	Units	Rec	RPD	Qual	RPD		Limits	
								Qual	Low	High	RPD
PB166985BS	AR1016	166.6	165	ug/kg	99				71	120	
	AR1260	166.6	161	ug/kg	97				65	130	

4C

PESTICIDE METHOD BLANK SUMMARY

EPA SAMPLE NO.

PB166985BL

Lab Name: CHEMTECH

Contract: JPCL01

Lab Code: CHEM Case No.: Q1486

SAS No.: Q1486 SDG NO.: Q1486

Lab Sample ID: PB166985BL

Lab File ID: PP070264.D

Matrix: (soil/water) Solid

Extraction: (Type) SOXH

Sulfur Cleanup: (Y/N) N

Date Extracted: 03/05/2025

Date Analyzed (1): 03/05/2025

Date Analyzed (2): 03/05/2025

Time Analyzed (1): 14:58

Time Analyzed (2): 14:58

Instrument ID (1): ECD_P

Instrument ID (2): ECD_P

GC Column (1): ZB-MR1 ID: 0.32 (mm)

GC Column (2): ZB-MR2 ID: 0.32 (mm)

THIS METHOD BLANK APPLIES TO THE FOLLOWING SAMPLES, MS AND MSD:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED 1	DATE ANALYZED 2
PB166985BS	PB166985BS	PP070265.D	03/05/2025	03/05/2025
DN-B-40	Q1486-01	PP070270.D	03/05/2025	03/05/2025
ENV-101-SB02MS	Q1488-03MS	PP070279.D	03/05/2025	03/05/2025
ENV-101-SB02MSD	Q1488-03MSD	PP070280.D	03/05/2025	03/05/2025

COMMENTS: _____



CALIBRATION SUMMARY

A
B
C
D
E
F
G

GC Column: ZB-MR1 ID: 0.32 (mm)

[illegible]

RETENTION TIMES OF INITIAL CALIBRATION

Decachlorobiphenyl	10.25	10.25	10.26	10.25	10.26	10.25	10.15	10.35
Tetrachloro-m-xylene	4.52	4.53	4.53	4.52	4.53	4.53	4.43	4.63

A
B
C
D
E
F
G

A
B
C
D
E
F
G

GC Column: ZB-MR2 ID: 0.32 (mm)

[illegible]

RETENTION TIMES OF INITIAL CALIBRATION

Decachlorobiphenyl	8.89	8.89	8.89	8.89	8.89	8.89	8.79	8.99
Tetrachloro-m-xylene	3.83	3.83	3.83	3.83	3.83	3.83	3.73	3.93

A

B

C

D

E

F

G

CALIBRATION FACTOR OF INITIAL CALIBRATION

Contract: JPCL01

Lab Code: CHEM **Case No.:** Q1486 **SAS No.:** Q1486 **SDG NO.:** Q1486

Instrument ID: ECD_P **Calibration Date(s):** 02/24/2025 02/24/2025
Calibration Times: 14:59 22:17

GC Column: ZB-MR1 **ID:** 0.32 (mm)

LAB FILE ID:		CF 1000 = <u>PP069996.D</u>		CF 750 = <u>PP069997.D</u>			
CF 500 = <u>PP069998.D</u>		CF 250 = <u>PP069999.D</u>		CF 050 = <u>PP070000.D</u>			
COMPOUND		CF 1000	CF 750	CF 500	CF 250	CF 050	% RSD
Aroclor-1016-1	(1)	45557625	47855512	49968516	54800208	50948980	49826168
Aroclor-1016-2	(2)	66706952	67781012	72991154	76640772	69769920	70777962
Aroclor-1016-3	(3)	40794671	44579197	45146958	47844192	41247180	43922440
Aroclor-1016-4	(4)	33978080	35043389	37559532	39668600	35057220	36261364
Aroclor-1016-5	(5)	31514906	31680175	34068042	35364724	35064900	33538549
Aroclor-1260-1	(1)	54823994	55486535	59388846	62695304	59405260	58359988
Aroclor-1260-2	(2)	73383444	75987972	80712800	85961116	92581060	81725278
Aroclor-1260-3	(3)	59489338	60560020	64011918	67792592	61951460	62761066
Aroclor-1260-4	(4)	58704582	59596489	63141360	67201512	68361020	63400993
Aroclor-1260-5	(5)	124764910	128178828	134805260	140592412	127581700	131184622
Decachlorobiphenyl		1071742430	1139629213	1165555660	1227723520	1090515400	1139033245
Tetrachloro-m-xylene		1433727050	1435859520	1533424440	1590932200	1344174400	1467623522
Aroclor-1242-1	(1)	38359202	40766156	41566668	47963164	43620760	42455190
Aroclor-1242-2	(2)	56757612	55666953	60328714	62664932	52294440	57542530
Aroclor-1242-3	(3)	34692482	36785992	38528310	46539712	33080140	37925327
Aroclor-1242-4	(4)	28738709	28291829	32018112	32372976	27652280	29814781
Aroclor-1242-5	(5)	32588629	32785485	35429746	39255208	36117200	35235254
Decachlorobiphenyl		1021050090	1049520227	1099648880	1125224680	979176800	1054924135
Tetrachloro-m-xylene		1392047690	1350066600	1467421460	1456806320	1242599600	1381788334
Aroclor-1248-1	(1)	29585124	31433301	32559076	36305496	33667780	32710155
Aroclor-1248-2	(2)	39473771	42013749	42864276	46958912	39450360	42152214
Aroclor-1248-3	(3)	43926772	44170904	46948106	50004072	40811360	45172243
Aroclor-1248-4	(4)	53918087	55655911	58640506	66304332	60297700	58963307
Aroclor-1248-5	(5)	51833391	52919351	56894804	67563228	56719440	57186043
Decachlorobiphenyl		1044729320	1067835560	1119982260	1186291160	1031748600	1090117380
Tetrachloro-m-xylene		1393266240	1426265507	1481945040	1527838200	1310175600	1427898117
Aroclor-1254-1	(1)	51156577	57570420	60622162	67008468	47144000	56700325
Aroclor-1254-2	(2)	78872042	81254612	86985854	93340172	97974740	87685484
Aroclor-1254-3	(3)	79973406	82322503	87641586	92940192	102754200	89126377
Aroclor-1254-4	(4)	66856800	68050653	73453394	76814612	87380160	74511124
Aroclor-1254-5	(5)	65315189	68448207	69620926	72356948	60070200	67162294
Decachlorobiphenyl		1075591380	1094765307	1221801800	1207489520	1097348200	1139399241
Tetrachloro-m-xylene		1443804100	1475957520	1555264320	1605313440	1392239200	1494515716
Aroclor-1268-1	(1)	188433469	184749517	193628348	206851508	186804180	192093404

CALIBRATION FACTOR OF INITIAL CALIBRATION

Aroclor-1268-2	(2)	162697797	159627413	168307400	177293676	160000200	165585297	4
Aroclor-1268-3	(3)	141621356	137143849	145386650	153273656	140094780	143504058	4
Aroclor-1268-4	(4)	63392961	60284985	64357574	66970196	60240320	63049207	5
Aroclor-1268-5	(5)	414016384	401287087	418333378	444189832	401775920	415920520	4
Decachlorobiphenyl		1839528480	1828279133	1916502300	2031355480	1847859400	1892704959	4
Tetrachloro-m-xylene		1501238100	1376017987	1532711540	1621137720	1417123400	1489645749	7

CALIBRATION FACTOR OF INITIAL CALIBRATION

Contract: JPCL01

Lab Code: CHEM **Case No.:** Q1486 **SAS No.:** Q1486 **SDG NO.:** Q1486

Instrument ID: ECD_P **Calibration Date(s):** 02/24/2025 02/24/2025
Calibration Times: 14:59 22:17

GC Column: ZB-MR2 **ID:** 0.32 (mm)

LAB FILE ID: <u>PP069996.D</u> CF 750 = <u>PP069997.D</u> CF 500 = <u>PP069998.D</u> CF 250 = <u>PP069999.D</u> CF 050 = <u>PP070000.D</u>							
COMPOUND		CF 1000	CF 750	CF 500	CF 250	CF 050	% RSD
Aroclor-1016-1	(1)	29885830	30191223	33586918	36186236	37142680	33398577
Aroclor-1016-2	(2)	42247324	42370143	46984138	50510676	50858960	46594248
Aroclor-1016-3	(3)	23325129	22666977	25816242	27757500	25596220	25032414
Aroclor-1016-4	(4)	18421261	17991491	20585360	22152124	21203180	20070683
Aroclor-1016-5	(5)	23822938	23028484	26536406	28616852	27728040	25946544
Aroclor-1260-1	(1)	44977992	45171227	49550340	56779968	51216800	49539265
Aroclor-1260-2	(2)	59688698	59926561	64720792	71640216	71124240	65420101
Aroclor-1260-3	(3)	52251359	51461445	57261004	62988144	77648420	60322074
Aroclor-1260-4	(4)	44361395	45109507	48615116	51524344	54726600	48867392
Aroclor-1260-5	(5)	111484150	113477796	119592622	125243744	126075940	119174850
Decachlorobiphenyl		984571270	1029317907	1118262220	1135932120	1152437600	1084104223
Tetrachloro-m-xylene		879784240	907640400	952717360	1058760600	981262000	956032920
Aroclor-1242-1	(1)	25489035	24475017	27848416	27117444	28407820	26667546
Aroclor-1242-2	(2)	35977820	35199769	38539928	39076044	37212540	37201220
Aroclor-1242-3	(3)	19863858	18273261	20997152	20082888	19211500	19685732
Aroclor-1242-4	(4)	18777771	17922656	19968364	21100536	20505160	19654897
Aroclor-1242-5	(5)	24360224	22407967	26366090	27368000	23855560	24871568
Decachlorobiphenyl		899284040	984619373	1017502000	1027377360	962336000	978223755
Tetrachloro-m-xylene		920809800	854904253	959106780	929536920	894966400	911864831
Aroclor-1248-1	(1)	19339318	20312839	22158310	22596092	21441940	21169700
Aroclor-1248-2	(2)	25841628	26874321	28982946	30984640	28578600	28252427
Aroclor-1248-3	(3)	26866525	28025000	30088164	32810700	30987640	29755606
Aroclor-1248-4	(4)	31739244	33065895	35285822	36808472	38827540	35145395
Aroclor-1248-5	(5)	32923317	34190145	36090288	38291344	39345240	36168067
Decachlorobiphenyl		999826530	1002616133	1003875200	1054961080	1165261200	1045308029
Tetrachloro-m-xylene		903111320	931293867	957405100	986258320	972075400	950028801
Aroclor-1254-1	(1)	50279933	51640439	55752488	60013056	55160920	54569367
Aroclor-1254-2	(2)	44271421	45567105	49578352	53549416	50761380	48745535
Aroclor-1254-3	(3)	70979950	72598121	78881282	84287020	72336540	75816583
Aroclor-1254-4	(4)	50771725	50149405	54249580	57298592	44830820	51460024
Aroclor-1254-5	(5)	64408865	66088107	72696700	75630908	59021560	67569228
Decachlorobiphenyl		963124600	1013018133	1145558760	1245512000	1148271200	1103096939
Tetrachloro-m-xylene		957176240	952284093	991535260	1081903280	978837400	992347255
Aroclor-1268-1	(1)	160284327	157759444	159569126	170516088	180615940	165748985

CALIBRATION FACTOR OF INITIAL CALIBRATION

Aroclor-1268-2	(2)	139528496	137224495	137304354	146542344	164608100	145041558	8
Aroclor-1268-3	(3)	118669811	115934707	120470364	129761704	127485360	122464389	5
Aroclor-1268-4	(4)	50956768	50155827	52914114	57814536	54084040	53185057	6
Aroclor-1268-5	(5)	345492428	353062379	344150492	366619864	366772500	355219533	3
Decachlorobiphenyl		1574839450	1670989493	1686779340	1791292760	1882315600	1721243329	7
Tetrachloro-m-xylene		969499550	855552893	995885280	1076293640	1034726400	986391553	8

INITIAL CALIBRATION OF MULTICOMPONENT ANALYTES

Contract: JPCL01

Lab Code: CHEM Case No.: Q1486 SAS No.: Q1486 SDG NO.: Q1486

Instrument ID: ECD_P Date(s) Analyzed: 02/24/2025 02/24/2025

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND	AMOUNT (ng)	PEAK	RT	RT WINDOW		CALIBRATION FACTOR
				FROM	TO	
Aroclor-1221	500	1	4.73	4.63	4.83	18915900
		2	4.81	4.71	4.91	14591600
		3	4.89	4.79	4.99	43158200
		4	0.00			0
		5	0.00			0
Aroclor-1232	500	1	4.89	4.79	4.99	32884000
		2	5.42	5.32	5.52	16260200
		3	5.70	5.60	5.80	34867200
		4	5.86	5.76	5.96	17799400
		5	5.95	5.85	6.05	12607200
Aroclor-1262	500	1	8.11	8.01	8.21	80729000
		2	8.43	8.33	8.53	161087000
		3	8.75	8.65	8.85	110612000
		4	8.84	8.74	8.94	83914000
		5	9.49	9.39	9.59	58804600

INITIAL CALIBRATION OF MULTICOMPONENT ANALYTES

Contract: JPCL01

Lab Code: CHEM Case No.: Q1486 SAS No.: Q1486 SDG NO.: Q1486

Instrument ID: ECD_P Date(s) Analyzed: 02/24/2025 02/24/2025

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND	AMOUNT (ng)	PEAK	RT	RT WINDOW		CALIBRATION FACTOR
				FROM	TO	
Aroclor-1221	500	1	4.04	3.94	4.14	13421200
		2	4.13	4.03	4.23	10166200
		3	4.21	4.11	4.31	30359600
		4	0.00			0
		5	0.00			0
Aroclor-1232	500	1	4.21	4.11	4.31	22742000
		2	4.94	4.84	5.04	22833200
		3	5.12	5.02	5.22	12024000
		4	5.20	5.10	5.30	10865300
		5	5.37	5.27	5.47	11305400
Aroclor-1262	500	1	6.97	6.87	7.07	83475400
		2	7.23	7.13	7.33	66291000
		3	7.75	7.65	7.85	60580400
		4	7.81	7.71	7.91	105474000
		5	8.32	8.22	8.42	51930200

CALIBRATION VERIFICATION SUMMARY

Contract: JPCL01

Lab Code: CHEM Case No.: Q1486 SAS No.: Q1486 SDG NO.: Q1486

Continuing Calib Date: 03/05/2025 Initial Calibration Date(s): 02/24/2025 02/24/2025

Continuing Calib Time: 11:49 Initial Calibration Time(s): 14:59 22:17

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Aroclor-1016-1	(1)	5.68	5.68	5.58	5.78	0.00
Aroclor-1016-2	(2)	5.70	5.70	5.60	5.80	0.00
Aroclor-1016-3	(3)	5.76	5.77	5.67	5.87	0.01
Aroclor-1016-4	(4)	5.86	5.86	5.76	5.96	0.00
Aroclor-1016-5	(5)	6.15	6.16	6.06	6.26	0.01
Aroclor-1260-1	(1)	7.27	7.28	7.18	7.38	0.01
Aroclor-1260-2	(2)	7.53	7.53	7.43	7.63	0.00
Aroclor-1260-3	(3)	7.89	7.89	7.79	7.99	0.00
Aroclor-1260-4	(4)	8.11	8.11	8.01	8.21	0.00
Aroclor-1260-5	(5)	8.43	8.43	8.33	8.53	0.00
Tetrachloro-m-xylene		4.53	4.53	4.43	4.63	0.00
Decachlorobiphenyl		10.25	10.26	10.16	10.36	0.01

CALIBRATION VERIFICATION SUMMARY

Contract: JPCL01

Lab Code: CHEM Case No.: Q1486 SAS No.: Q1486 SDG NO.: Q1486

Continuing Calib Date: 03/05/2025 Initial Calibration Date(s): 02/24/2025 02/24/2025

Continuing Calib Time: 11:49 Initial Calibration Time(s): 14:59 22:17

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Aroclor-1016-1	(1)	4.92	4.92	4.82	5.02	0.00
Aroclor-1016-2	(2)	4.94	4.94	4.84	5.04	0.00
Aroclor-1016-3	(3)	5.11	5.12	5.02	5.22	0.01
Aroclor-1016-4	(4)	5.15	5.16	5.06	5.26	0.01
Aroclor-1016-5	(5)	5.37	5.37	5.27	5.47	0.00
Aroclor-1260-1	(1)	6.41	6.41	6.31	6.51	0.00
Aroclor-1260-2	(2)	6.60	6.60	6.50	6.70	0.00
Aroclor-1260-3	(3)	6.75	6.75	6.65	6.85	0.00
Aroclor-1260-4	(4)	7.22	7.23	7.13	7.33	0.01
Aroclor-1260-5	(5)	7.46	7.47	7.37	7.57	0.01
Tetrachloro-m-xylene		3.83	3.83	3.73	3.93	0.00
Decachlorobiphenyl		8.88	8.89	8.79	8.99	0.01

CALIBRATION VERIFICATION SUMMARY

Contract: JPCL01

Lab Code: CHEM Case No.: Q1486 SAS No.: Q1486 SDG NO.: Q1486

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 02/24/2025 02/24/2025

Client Sample No.: CCAL01 Date Analyzed: 03/05/2025

Lab Sample No.: AR1660CCC500 Data File : PP070259.D Time Analyzed: 11:49

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.679	5.581	5.781	474.860	500.000	-5.0
Aroclor-1016-2	5.700	5.602	5.802	491.780	500.000	-1.6
Aroclor-1016-3	5.763	5.665	5.865	489.540	500.000	-2.1
Aroclor-1016-4	5.860	5.762	5.962	496.730	500.000	-0.7
Aroclor-1016-5	6.153	6.056	6.256	483.820	500.000	-3.2
Aroclor-1260-1	7.273	7.175	7.375	482.320	500.000	-3.5
Aroclor-1260-2	7.527	7.429	7.629	455.430	500.000	-8.9
Aroclor-1260-3	7.885	7.788	7.988	488.800	500.000	-2.2
Aroclor-1260-4	8.110	8.012	8.212	479.200	500.000	-4.2
Aroclor-1260-5	8.430	8.333	8.533	484.960	500.000	-3.0
Decachlorobiphenyl	10.253	10.156	10.356	47.630	50.000	-4.7
Tetrachloro-m-xylene	4.526	4.427	4.627	52.370	50.000	4.7

CALIBRATION VERIFICATION SUMMARY

Contract: JPCL01

Lab Code: CHEM Case No.: Q1486 SAS No.: Q1486 SDG NO.: Q1486

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 02/24/2025 02/24/2025

Client Sample No.: CCAL01 Date Analyzed: 03/05/2025

Lab Sample No.: AR1660CCC500 Data File : PP070259.D Time Analyzed: 11:49

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.917	4.820	5.020	513.670	500.000	2.7
Aroclor-1016-2	4.936	4.839	5.039	509.760	500.000	2.0
Aroclor-1016-3	5.113	5.017	5.217	518.780	500.000	3.8
Aroclor-1016-4	5.154	5.058	5.258	507.120	500.000	1.4
Aroclor-1016-5	5.370	5.273	5.473	534.650	500.000	6.9
Aroclor-1260-1	6.407	6.312	6.512	493.320	500.000	-1.3
Aroclor-1260-2	6.596	6.500	6.700	502.660	500.000	0.5
Aroclor-1260-3	6.750	6.654	6.854	468.200	500.000	-6.4
Aroclor-1260-4	7.222	7.126	7.326	524.080	500.000	4.8
Aroclor-1260-5	7.463	7.367	7.567	533.970	500.000	6.8
Decachlorobiphenyl	8.882	8.788	8.988	48.820	50.000	-2.4
Tetrachloro-m-xylene	3.828	3.730	3.930	51.480	50.000	3.0

CALIBRATION VERIFICATION SUMMARY

Contract: JPCL01

Lab Code: CHEM Case No.: Q1486 SAS No.: Q1486 SDG NO.: Q1486

Continuing Calib Date: 03/05/2025 Initial Calibration Date(s): 02/24/2025 02/24/2025

Continuing Calib Time: 18:13 Initial Calibration Time(s): 14:59 22:17

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW FROM TO		DIFF RT
Aroclor-1016-1	(1)	5.68	5.68	5.58	5.78	0.00
Aroclor-1016-2	(2)	5.70	5.70	5.60	5.80	0.00
Aroclor-1016-3	(3)	5.76	5.77	5.67	5.87	0.01
Aroclor-1016-4	(4)	5.86	5.86	5.76	5.96	0.00
Aroclor-1016-5	(5)	6.15	6.16	6.06	6.26	0.01
Aroclor-1260-1	(1)	7.27	7.28	7.18	7.38	0.01
Aroclor-1260-2	(2)	7.53	7.53	7.43	7.63	0.00
Aroclor-1260-3	(3)	7.89	7.89	7.79	7.99	0.00
Aroclor-1260-4	(4)	8.11	8.11	8.01	8.21	0.00
Aroclor-1260-5	(5)	8.43	8.43	8.33	8.53	0.00
Tetrachloro-m-xylene		4.53	4.53	4.43	4.63	0.00
Decachlorobiphenyl		10.25	10.26	10.16	10.36	0.01

CALIBRATION VERIFICATION SUMMARY

Contract: JPCL01

Lab Code: CHEM Case No.: Q1486 SAS No.: Q1486 SDG NO.: Q1486

Continuing Calib Date: 03/05/2025 Initial Calibration Date(s): 02/24/2025 02/24/2025

Continuing Calib Time: 18:13 Initial Calibration Time(s): 14:59 22:17

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW FROM TO		DIFF RT
Aroclor-1016-1	(1)	4.92	4.92	4.82	5.02	0.00
Aroclor-1016-2	(2)	4.94	4.94	4.84	5.04	0.00
Aroclor-1016-3	(3)	5.11	5.12	5.02	5.22	0.01
Aroclor-1016-4	(4)	5.16	5.16	5.06	5.26	0.00
Aroclor-1016-5	(5)	5.37	5.37	5.27	5.47	0.00
Aroclor-1260-1	(1)	6.41	6.41	6.31	6.51	0.00
Aroclor-1260-2	(2)	6.60	6.60	6.50	6.70	0.00
Aroclor-1260-3	(3)	6.75	6.75	6.65	6.85	0.00
Aroclor-1260-4	(4)	7.22	7.23	7.13	7.33	0.01
Aroclor-1260-5	(5)	7.46	7.47	7.37	7.57	0.01
Tetrachloro-m-xylene		3.83	3.83	3.73	3.93	0.00
Decachlorobiphenyl		8.88	8.89	8.79	8.99	0.01

CALIBRATION VERIFICATION SUMMARY

Contract: JPCL01

Lab Code: CHEM Case No.: Q1486 SAS No.: Q1486 SDG NO.: Q1486

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 02/24/2025 02/24/2025

Client Sample No.: CCAL02 Date Analyzed: 03/05/2025

Lab Sample No.: AR1660CCC500 Data File : PP070274.D Time Analyzed: 18:13

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.679	5.581	5.781	519.700	500.000	3.9
Aroclor-1016-2	5.701	5.602	5.802	539.100	500.000	7.8
Aroclor-1016-3	5.763	5.665	5.865	538.430	500.000	7.7
Aroclor-1016-4	5.860	5.762	5.962	548.080	500.000	9.6
Aroclor-1016-5	6.154	6.056	6.256	519.480	500.000	3.9
Aroclor-1260-1	7.274	7.175	7.375	514.400	500.000	2.9
Aroclor-1260-2	7.527	7.429	7.629	499.540	500.000	-0.1
Aroclor-1260-3	7.886	7.788	7.988	529.380	500.000	5.9
Aroclor-1260-4	8.110	8.012	8.212	505.520	500.000	1.1
Aroclor-1260-5	8.431	8.333	8.533	525.530	500.000	5.1
Decachlorobiphenyl	10.252	10.156	10.356	50.370	50.000	0.7
Tetrachloro-m-xylene	4.526	4.427	4.627	55.610	50.000	11.2

CALIBRATION VERIFICATION SUMMARY

Contract: JPCL01

Lab Code: CHEM Case No.: Q1486 SAS No.: Q1486 SDG NO.: Q1486

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 02/24/2025 02/24/2025

Client Sample No.: CCAL02 Date Analyzed: 03/05/2025

Lab Sample No.: AR1660CCC500 Data File : PP070274.D Time Analyzed: 18:13

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.917	4.820	5.020	528.080	500.000	5.6
Aroclor-1016-2	4.936	4.839	5.039	524.790	500.000	5.0
Aroclor-1016-3	5.113	5.017	5.217	536.640	500.000	7.3
Aroclor-1016-4	5.155	5.058	5.258	534.000	500.000	6.8
Aroclor-1016-5	5.370	5.273	5.473	553.470	500.000	10.7
Aroclor-1260-1	6.408	6.312	6.512	510.260	500.000	2.1
Aroclor-1260-2	6.596	6.500	6.700	514.490	500.000	2.9
Aroclor-1260-3	6.750	6.654	6.854	485.590	500.000	-2.9
Aroclor-1260-4	7.223	7.126	7.326	541.710	500.000	8.3
Aroclor-1260-5	7.464	7.367	7.567	564.500	500.000	12.9
Decachlorobiphenyl	8.883	8.788	8.988	50.840	50.000	1.7
Tetrachloro-m-xylene	3.828	3.730	3.930	52.730	50.000	5.5

CALIBRATION VERIFICATION SUMMARY

Contract: JPCL01

Lab Code: CHEM Case No.: Q1486 SAS No.: Q1486 SDG NO.: Q1486

Continuing Calib Date: 03/05/2025 Initial Calibration Date(s): 02/24/2025 02/24/2025

Continuing Calib Time: 22:50 Initial Calibration Time(s): 14:59 22:17

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Aroclor-1016-1	(1)	5.68	5.68	5.58	5.78	0.00
Aroclor-1016-2	(2)	5.70	5.70	5.60	5.80	0.00
Aroclor-1016-3	(3)	5.76	5.77	5.67	5.87	0.01
Aroclor-1016-4	(4)	5.86	5.86	5.76	5.96	0.00
Aroclor-1016-5	(5)	6.16	6.16	6.06	6.26	0.00
Aroclor-1260-1	(1)	7.28	7.28	7.18	7.38	0.00
Aroclor-1260-2	(2)	7.53	7.53	7.43	7.63	0.00
Aroclor-1260-3	(3)	7.89	7.89	7.79	7.99	0.00
Aroclor-1260-4	(4)	8.11	8.11	8.01	8.21	0.00
Aroclor-1260-5	(5)	8.43	8.43	8.33	8.53	0.00
Tetrachloro-m-xylene		4.53	4.53	4.43	4.63	0.00
Decachlorobiphenyl		10.26	10.26	10.16	10.36	0.00

CALIBRATION VERIFICATION SUMMARY

Contract: JPCL01

Lab Code: CHEM Case No.: Q1486 SAS No.: Q1486 SDG NO.: Q1486

Continuing Calib Date: 03/05/2025 Initial Calibration Date(s): 02/24/2025 02/24/2025

Continuing Calib Time: 22:50 Initial Calibration Time(s): 14:59 22:17

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Aroclor-1016-1	(1)	4.92	4.92	4.82	5.02	0.00
Aroclor-1016-2	(2)	4.94	4.94	4.84	5.04	0.00
Aroclor-1016-3	(3)	5.12	5.12	5.02	5.22	0.01
Aroclor-1016-4	(4)	5.16	5.16	5.06	5.26	0.00
Aroclor-1016-5	(5)	5.37	5.37	5.27	5.47	0.00
Aroclor-1260-1	(1)	6.41	6.41	6.31	6.51	0.00
Aroclor-1260-2	(2)	6.60	6.60	6.50	6.70	0.00
Aroclor-1260-3	(3)	6.75	6.75	6.65	6.85	0.00
Aroclor-1260-4	(4)	7.23	7.23	7.13	7.33	0.01
Aroclor-1260-5	(5)	7.47	7.47	7.37	7.57	0.00
Tetrachloro-m-xylene		3.83	3.83	3.73	3.93	0.00
Decachlorobiphenyl		8.89	8.89	8.79	8.99	0.00

CALIBRATION VERIFICATION SUMMARY

Contract: JPCL01

Lab Code: CHEM Case No.: Q1486 SAS No.: Q1486 SDG NO.: Q1486

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 02/24/2025 02/24/2025

Client Sample No.: CCAL03 Date Analyzed: 03/05/2025

Lab Sample No.: AR1660CCC500 Data File : PP070289.D Time Analyzed: 22:50

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	5.681	5.581	5.781	507.650	500.000	1.5
Aroclor-1016-2	5.702	5.602	5.802	514.280	500.000	2.9
Aroclor-1016-3	5.764	5.665	5.865	501.590	500.000	0.3
Aroclor-1016-4	5.862	5.762	5.962	507.010	500.000	1.4
Aroclor-1016-5	6.155	6.056	6.256	518.280	500.000	3.7
Aroclor-1260-1	7.275	7.175	7.375	532.180	500.000	6.4
Aroclor-1260-2	7.529	7.429	7.629	505.610	500.000	1.1
Aroclor-1260-3	7.888	7.788	7.988	546.130	500.000	9.2
Aroclor-1260-4	8.113	8.012	8.212	525.010	500.000	5.0
Aroclor-1260-5	8.433	8.333	8.533	541.570	500.000	8.3
Decachlorobiphenyl	10.258	10.156	10.356	51.680	50.000	3.4
Tetrachloro-m-xylene	4.527	4.427	4.627	55.950	50.000	11.9

CALIBRATION VERIFICATION SUMMARY

Contract: JPCL01

Lab Code: CHEM Case No.: Q1486 SAS No.: Q1486 SDG NO.: Q1486

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 02/24/2025 02/24/2025

Client Sample No.: CCAL03 Date Analyzed: 03/05/2025

Lab Sample No.: AR1660CCC500 Data File : PP070289.D Time Analyzed: 22:50

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.919	4.820	5.020	536.170	500.000	7.2
Aroclor-1016-2	4.938	4.839	5.039	532.730	500.000	6.5
Aroclor-1016-3	5.115	5.017	5.217	542.770	500.000	8.6
Aroclor-1016-4	5.157	5.058	5.258	529.850	500.000	6.0
Aroclor-1016-5	5.372	5.273	5.473	573.280	500.000	14.7
Aroclor-1260-1	6.410	6.312	6.512	516.300	500.000	3.3
Aroclor-1260-2	6.598	6.500	6.700	527.370	500.000	5.5
Aroclor-1260-3	6.752	6.654	6.854	498.350	500.000	-0.3
Aroclor-1260-4	7.225	7.126	7.326	567.590	500.000	13.5
Aroclor-1260-5	7.466	7.367	7.567	585.700	500.000	17.1
Decachlorobiphenyl	8.886	8.788	8.988	52.490	50.000	5.0
Tetrachloro-m-xylene	3.830	3.730	3.930	52.590	50.000	5.2

Analytical Sequence

Client: JPCL Engineering	SDG No.: Q1486
Project: NYSDOT Two Bronx River Parkway Bridge	Instrument ID: ECD_P
GC Column: ZB-MR1	ID: 0.32 (mm) Inst. Calib. Date(s): 02/24/2025 02/24/2025

THE ANALYTICAL SEQUENCE OF PERFORMANCE EVALUATION MIXTURES, BLANKS, SAMPLES,
AND STANDARDS IS GIVEN BELOW:

EPA SAMPLE NO.	LAB SAMPLE ID	DATE ANALYZED	TIME ANALYZED	DATAFILE	DCB RT #	TCX RT #
IBLK	IBLK	02/24/2025	14:43	PP069995.D	10.25	4.53
AR1660ICC1000	AR1660ICC1000	02/24/2025	14:59	PP069996.D	10.25	4.52
AR1660ICC750	AR1660ICC750	02/24/2025	15:15	PP069997.D	10.26	4.53
AR1660ICC500	AR1660ICC500	02/24/2025	15:32	PP069998.D	10.26	4.53
AR1660ICC250	AR1660ICC250	02/24/2025	15:48	PP069999.D	10.25	4.52
AR1660ICC050	AR1660ICC050	02/24/2025	16:04	PP070000.D	10.26	4.53
AR1221ICC500	AR1221ICC500	02/24/2025	16:20	PP070001.D	10.25	4.52
AR1232ICC500	AR1232ICC500	02/24/2025	16:37	PP070002.D	10.26	4.53
AR1242ICC1000	AR1242ICC1000	02/24/2025	16:53	PP070003.D	10.26	4.53
AR1242ICC750	AR1242ICC750	02/24/2025	17:09	PP070004.D	10.25	4.52
AR1242ICC500	AR1242ICC500	02/24/2025	17:25	PP070005.D	10.26	4.53
AR1242ICC250	AR1242ICC250	02/24/2025	17:42	PP070006.D	10.25	4.53
AR1242ICC050	AR1242ICC050	02/24/2025	17:58	PP070007.D	10.26	4.53
AR1248ICC1000	AR1248ICC1000	02/24/2025	18:14	PP070008.D	10.26	4.53
AR1248ICC750	AR1248ICC750	02/24/2025	18:30	PP070009.D	10.26	4.53
AR1248ICC500	AR1248ICC500	02/24/2025	18:46	PP070010.D	10.26	4.53
AR1248ICC250	AR1248ICC250	02/24/2025	19:03	PP070011.D	10.26	4.53
AR1248ICC050	AR1248ICC050	02/24/2025	19:19	PP070012.D	10.25	4.52
AR1254ICC1000	AR1254ICC1000	02/24/2025	19:35	PP070013.D	10.26	4.53
AR1254ICC750	AR1254ICC750	02/24/2025	19:51	PP070014.D	10.25	4.52
AR1254ICC500	AR1254ICC500	02/24/2025	20:08	PP070015.D	10.26	4.53
AR1254ICC250	AR1254ICC250	02/24/2025	20:24	PP070016.D	10.26	4.53
AR1254ICC050	AR1254ICC050	02/24/2025	20:40	PP070017.D	10.26	4.52
AR1262ICC500	AR1262ICC500	02/24/2025	20:56	PP070018.D	10.25	4.53
AR1268ICC1000	AR1268ICC1000	02/24/2025	21:12	PP070019.D	10.25	4.52
AR1268ICC750	AR1268ICC750	02/24/2025	21:29	PP070020.D	10.25	4.53
AR1268ICC500	AR1268ICC500	02/24/2025	21:45	PP070021.D	10.26	4.53
AR1268ICC250	AR1268ICC250	02/24/2025	22:01	PP070022.D	10.25	4.52
AR1268ICC050	AR1268ICC050	02/24/2025	22:17	PP070023.D	10.26	4.53
AR1660CCC500	AR1660CCC500	03/05/2025	11:49	PP070259.D	10.25	4.53
IBLK	IBLK	03/05/2025	12:54	PP070263.D	10.25	4.52
PB166985BL	PB166985BL	03/05/2025	14:58	PP070264.D	10.26	4.53
PB166985BS	PB166985BS	03/05/2025	15:14	PP070265.D	10.26	4.53
DN-B-40	Q1486-01	03/05/2025	16:36	PP070270.D	10.26	4.53
AR1660CCC500	AR1660CCC500	03/05/2025	18:13	PP070274.D	10.25	4.53
IBLK	IBLK	03/05/2025	19:19	PP070278.D	10.26	4.53
ENV-101-SB02MS	Q1488-03MS	03/05/2025	19:35	PP070279.D	10.26	4.53
ENV-101-SB02MSD	Q1488-03MSD	03/05/2025	19:51	PP070280.D	10.26	4.53
AR1660CCC500	AR1660CCC500	03/05/2025	22:50	PP070289.D	10.26	4.53
IBLK	IBLK	03/05/2025	23:56	PP070293.D	10.26	4.52

Analytical Sequence

Client: JPCL Engineering	SDG No.: Q1486	
Project: NYSDOT Two Bronx River Parkway Bridge	Instrument ID: ECD_P	
GC Column: ZB-MR2	ID: 0.32 (mm)	Inst. Calib. Date(s): 02/24/2025 02/24/2025

THE ANALYTICAL SEQUENCE OF PERFORMANCE EVALUATION MIXTURES, BLANKS, SAMPLES,
AND STANDARDS IS GIVEN BELOW:

EPA SAMPLE NO.	LAB SAMPLE ID	DATE ANALYZED	TIME ANALYZED	DATAFILE	DCB RT #	TCX RT #
IBLK	IBLK	02/24/2025	14:43	PP069995.D	8.89	3.83
AR1660ICC1000	AR1660ICC1000	02/24/2025	14:59	PP069996.D	8.89	3.83
AR1660ICC750	AR1660ICC750	02/24/2025	15:15	PP069997.D	8.89	3.83
AR1660ICC500	AR1660ICC500	02/24/2025	15:32	PP069998.D	8.89	3.83
AR1660ICC250	AR1660ICC250	02/24/2025	15:48	PP069999.D	8.89	3.83
AR1660ICC050	AR1660ICC050	02/24/2025	16:04	PP070000.D	8.89	3.83
AR1221ICC500	AR1221ICC500	02/24/2025	16:20	PP070001.D	8.89	3.83
AR1232ICC500	AR1232ICC500	02/24/2025	16:37	PP070002.D	8.89	3.83
AR1242ICC1000	AR1242ICC1000	02/24/2025	16:53	PP070003.D	8.89	3.83
AR1242ICC750	AR1242ICC750	02/24/2025	17:09	PP070004.D	8.89	3.83
AR1242ICC500	AR1242ICC500	02/24/2025	17:25	PP070005.D	8.89	3.83
AR1242ICC250	AR1242ICC250	02/24/2025	17:42	PP070006.D	8.89	3.83
AR1242ICC050	AR1242ICC050	02/24/2025	17:58	PP070007.D	8.89	3.83
AR1248ICC1000	AR1248ICC1000	02/24/2025	18:14	PP070008.D	8.89	3.83
AR1248ICC750	AR1248ICC750	02/24/2025	18:30	PP070009.D	8.89	3.83
AR1248ICC500	AR1248ICC500	02/24/2025	18:46	PP070010.D	8.89	3.83
AR1248ICC250	AR1248ICC250	02/24/2025	19:03	PP070011.D	8.89	3.83
AR1248ICC050	AR1248ICC050	02/24/2025	19:19	PP070012.D	8.89	3.83
AR1254ICC1000	AR1254ICC1000	02/24/2025	19:35	PP070013.D	8.89	3.83
AR1254ICC750	AR1254ICC750	02/24/2025	19:51	PP070014.D	8.89	3.83
AR1254ICC500	AR1254ICC500	02/24/2025	20:08	PP070015.D	8.89	3.83
AR1254ICC250	AR1254ICC250	02/24/2025	20:24	PP070016.D	8.89	3.83
AR1254ICC050	AR1254ICC050	02/24/2025	20:40	PP070017.D	8.89	3.83
AR1262ICC500	AR1262ICC500	02/24/2025	20:56	PP070018.D	8.89	3.83
AR1268ICC1000	AR1268ICC1000	02/24/2025	21:12	PP070019.D	8.89	3.83
AR1268ICC750	AR1268ICC750	02/24/2025	21:29	PP070020.D	8.89	3.83
AR1268ICC500	AR1268ICC500	02/24/2025	21:45	PP070021.D	8.89	3.83
AR1268ICC250	AR1268ICC250	02/24/2025	22:01	PP070022.D	8.89	3.83
AR1268ICC050	AR1268ICC050	02/24/2025	22:17	PP070023.D	8.89	3.83
AR1660CCC500	AR1660CCC500	03/05/2025	11:49	PP070259.D	8.88	3.83
IBLK	IBLK	03/05/2025	12:54	PP070263.D	8.88	3.83
PB166985BL	PB166985BL	03/05/2025	14:58	PP070264.D	8.89	3.83
PB166985BS	PB166985BS	03/05/2025	15:14	PP070265.D	8.88	3.83
DN-B-40	Q1486-01	03/05/2025	16:36	PP070270.D	8.89	3.83
AR1660CCC500	AR1660CCC500	03/05/2025	18:13	PP070274.D	8.88	3.83
IBLK	IBLK	03/05/2025	19:19	PP070278.D	8.88	3.83
ENV-101-SB02MS	Q1488-03MS	03/05/2025	19:35	PP070279.D	8.89	3.83
ENV-101-SB02MSD	Q1488-03MSD	03/05/2025	19:51	PP070280.D	8.88	3.83
AR1660CCC500	AR1660CCC500	03/05/2025	22:50	PP070289.D	8.89	3.83
IBLK	IBLK	03/05/2025	23:56	PP070293.D	8.89	3.83



QC SAMPLE DATA

Report of Analysis

Client:	JPCL Engineering	Date Collected:	
Project:	NYSDOT Two Bronx River Parkway Bridges	Date Received:	
Client Sample ID:	PB166985BL	SDG No.:	Q1486
Lab Sample ID:	PB166985BL	Matrix:	SOIL
Analytical Method:	SW8082A	% Solid:	100
Sample Wt/Vol:	30.01 Units: g	Final Vol:	10000 uL
Soil Aliquot Vol:	uL	Test:	PCB
Extraction Type:		Injection Volume :	
GPC Factor :	1.0	PH :	
Prep Method :	SW3541B		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PP070264.D	1	03/05/25 09:10	03/05/25 14:58	PB166985

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
12674-11-2	Aroclor-1016	17.0	U	3.40	17.0	ug/kg
11104-28-2	Aroclor-1221	17.0	U	6.40	17.0	ug/kg
11141-16-5	Aroclor-1232	17.0	U	3.40	17.0	ug/kg
53469-21-9	Aroclor-1242	17.0	U	3.40	17.0	ug/kg
12672-29-6	Aroclor-1248	17.0	U	7.90	17.0	ug/kg
11097-69-1	Aroclor-1254	17.0	U	2.70	17.0	ug/kg
37324-23-5	Aroclor-1262	17.0	U	4.60	17.0	ug/kg
11100-14-4	Aroclor-1268	17.0	U	3.40	17.0	ug/kg
11096-82-5	Aroclor-1260	17.0	U	2.90	17.0	ug/kg
SURROGATES						
877-09-8	Tetrachloro-m-xylene	24.9		32 - 144	125%	SPK: 20
2051-24-3	Decachlorobiphenyl	23.8		32 - 175	119%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Report of Analysis

Client:	JPCL Engineering	Date Collected:	02/24/25
Project:	NYSDOT Two Bronx River Parkway Bridges	Date Received:	02/24/25
Client Sample ID:	PIBLK-PP069995.D	SDG No.:	Q1486
Lab Sample ID:	I.BLK-PP069995.D	Matrix:	WATER
Analytical Method:	SW8082A	% Solid:	0
Sample Wt/Vol:	1000 Units: mL	Final Vol:	10000 uL
Soil Aliquot Vol:	uL	Test:	PCB
Extraction Type:		Injection Volume :	
GPC Factor :	1.0	PH :	
Prep Method :	5030		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PP069995.D	1		02/24/25	PP022425

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
12674-11-2	Aroclor-1016	0.50	U	0.15	0.50	ug/L
11104-28-2	Aroclor-1221	0.50	U	0.23	0.50	ug/L
11141-16-5	Aroclor-1232	0.50	U	0.37	0.50	ug/L
53469-21-9	Aroclor-1242	0.50	U	0.16	0.50	ug/L
12672-29-6	Aroclor-1248	0.50	U	0.12	0.50	ug/L
11097-69-1	Aroclor-1254	0.50	U	0.11	0.50	ug/L
11096-82-5	Aroclor-1260	0.50	U	0.15	0.50	ug/L
37324-23-5	Aroclor-1262	0.50	U	0.14	0.50	ug/L
11100-14-4	Aroclor-1268	0.50	U	0.12	0.50	ug/L
SURROGATES						
877-09-8	Tetrachloro-m-xylene	21.9		60 - 140	109%	SPK: 20
2051-24-3	Decachlorobiphenyl	21.6		60 - 140	108%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Report of Analysis

Client:	JPCL Engineering	Date Collected:	03/05/25
Project:	NYSDOT Two Bronx River Parkway Bridges	Date Received:	03/05/25
Client Sample ID:	PIBLK-PP070263.D	SDG No.:	Q1486
Lab Sample ID:	I.BLK-PP070263.D	Matrix:	WATER
Analytical Method:	SW8082A	% Solid:	0
Sample Wt/Vol:	1000 Units: mL	Final Vol:	10000 uL
Soil Aliquot Vol:	uL	Test:	PCB
Extraction Type:		Injection Volume :	
GPC Factor :	1.0	PH :	
Prep Method :	5030		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PP070263.D	1		03/05/25	PP030525

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
12674-11-2	Aroclor-1016	0.50	U	0.15	0.50	ug/L
11104-28-2	Aroclor-1221	0.50	U	0.23	0.50	ug/L
11141-16-5	Aroclor-1232	0.50	U	0.37	0.50	ug/L
53469-21-9	Aroclor-1242	0.50	U	0.16	0.50	ug/L
12672-29-6	Aroclor-1248	0.50	U	0.12	0.50	ug/L
11097-69-1	Aroclor-1254	0.50	U	0.11	0.50	ug/L
11096-82-5	Aroclor-1260	0.50	U	0.15	0.50	ug/L
37324-23-5	Aroclor-1262	0.50	U	0.14	0.50	ug/L
11100-14-4	Aroclor-1268	0.50	U	0.12	0.50	ug/L
SURROGATES						
877-09-8	Tetrachloro-m-xylene	20.1		60 - 140	100%	SPK: 20
2051-24-3	Decachlorobiphenyl	18.8		60 - 140	94%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Report of Analysis

Client:	JPCL Engineering	Date Collected:	03/05/25
Project:	NYSDOT Two Bronx River Parkway Bridges	Date Received:	03/05/25
Client Sample ID:	PIBLK-PP070278.D	SDG No.:	Q1486
Lab Sample ID:	I.BLK-PP070278.D	Matrix:	WATER
Analytical Method:	SW8082A	% Solid:	0
Sample Wt/Vol:	1000 Units: mL	Final Vol:	10000 uL
Soil Aliquot Vol:	uL	Test:	PCB
Extraction Type:		Injection Volume :	
GPC Factor :	1.0	PH :	
Prep Method :	5030		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PP070278.D	1		03/05/25	PP030525

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
12674-11-2	Aroclor-1016	0.50	U	0.15	0.50	ug/L
11104-28-2	Aroclor-1221	0.50	U	0.23	0.50	ug/L
11141-16-5	Aroclor-1232	0.50	U	0.37	0.50	ug/L
53469-21-9	Aroclor-1242	0.50	U	0.16	0.50	ug/L
12672-29-6	Aroclor-1248	0.50	U	0.12	0.50	ug/L
11097-69-1	Aroclor-1254	0.50	U	0.11	0.50	ug/L
11096-82-5	Aroclor-1260	0.50	U	0.15	0.50	ug/L
37324-23-5	Aroclor-1262	0.50	U	0.14	0.50	ug/L
11100-14-4	Aroclor-1268	0.50	U	0.12	0.50	ug/L
SURROGATES						
877-09-8	Tetrachloro-m-xylene	20.7		60 - 140	104%	SPK: 20
2051-24-3	Decachlorobiphenyl	19.6		60 - 140	98%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Report of Analysis

Client:	JPCL Engineering	Date Collected:	03/05/25
Project:	NYSDOT Two Bronx River Parkway Bridges	Date Received:	03/05/25
Client Sample ID:	PIBLK-PP070293.D	SDG No.:	Q1486
Lab Sample ID:	I.BLK-PP070293.D	Matrix:	WATER
Analytical Method:	SW8082A	% Solid:	0
Sample Wt/Vol:	1000 Units: mL	Decanted:	
Soil Aliquot Vol:	uL	Final Vol:	10000 uL
Extraction Type:		Test:	PCB
GPC Factor :	1.0	Injection Volume :	
Prep Method :	5030		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PP070293.D	1		03/05/25	PP030525

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
12674-11-2	Aroclor-1016	0.50	U	0.15	0.50	ug/L
11104-28-2	Aroclor-1221	0.50	U	0.23	0.50	ug/L
11141-16-5	Aroclor-1232	0.50	U	0.37	0.50	ug/L
53469-21-9	Aroclor-1242	0.50	U	0.16	0.50	ug/L
12672-29-6	Aroclor-1248	0.50	U	0.12	0.50	ug/L
11097-69-1	Aroclor-1254	0.50	U	0.11	0.50	ug/L
11096-82-5	Aroclor-1260	0.50	U	0.15	0.50	ug/L
37324-23-5	Aroclor-1262	0.50	U	0.14	0.50	ug/L
11100-14-4	Aroclor-1268	0.50	U	0.12	0.50	ug/L
SURROGATES						
877-09-8	Tetrachloro-m-xylene	21.6		60 - 140	108%	SPK: 20
2051-24-3	Decachlorobiphenyl	20.2		60 - 140	101%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Report of Analysis

Client:	JPCL Engineering	Date Collected:	
Project:	NYSDOT Two Bronx River Parkway Bridges	Date Received:	
Client Sample ID:	PB166985BS	SDG No.:	Q1486
Lab Sample ID:	PB166985BS	Matrix:	SOIL
Analytical Method:	SW8082A	% Solid:	100
Sample Wt/Vol:	30.02 Units: g	Final Vol:	10000 uL
Soil Aliquot Vol:	uL	Test:	PCB
Extraction Type:		Injection Volume :	
GPC Factor :	1.0	PH :	
Prep Method :	SW3541B		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PP070265.D	1	03/05/25 09:10	03/05/25 15:14	PB166985

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
12674-11-2	Aroclor-1016	165		3.40	17.0	ug/kg
11104-28-2	Aroclor-1221	17.0	U	6.40	17.0	ug/kg
11141-16-5	Aroclor-1232	17.0	U	3.40	17.0	ug/kg
53469-21-9	Aroclor-1242	17.0	U	3.40	17.0	ug/kg
12672-29-6	Aroclor-1248	17.0	U	7.90	17.0	ug/kg
11097-69-1	Aroclor-1254	17.0	U	2.70	17.0	ug/kg
37324-23-5	Aroclor-1262	17.0	U	4.60	17.0	ug/kg
11100-14-4	Aroclor-1268	17.0	U	3.40	17.0	ug/kg
11096-82-5	Aroclor-1260	161		2.90	17.0	ug/kg
SURROGATES						
877-09-8	Tetrachloro-m-xylene	25.4		32 - 144	127%	SPK: 20
2051-24-3	Decachlorobiphenyl	23.7		32 - 175	119%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Report of Analysis

Client:	JPCL Engineering	Date Collected:	03/04/25
Project:	NYSDOT Two Bronx River Parkway Bridges	Date Received:	03/04/25
Client Sample ID:	ENV-101-SB02MS	SDG No.:	Q1486
Lab Sample ID:	Q1488-03MS	Matrix:	SOIL
Analytical Method:	SW8082A	% Solid:	90.6
Sample Wt/Vol:	30.07 Units: g	Final Vol:	10000 uL
Soil Aliquot Vol:	uL	Test:	PCB
Extraction Type:		Injection Volume :	
GPC Factor :	1.0	PH :	
Prep Method :	SW3541B		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PP070279.D	1	03/05/25 09:10	03/05/25 19:35	PB166985

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
12674-11-2	Aroclor-1016	188		3.70	18.7	ug/kg
11104-28-2	Aroclor-1221	18.7	U	7.10	18.7	ug/kg
11141-16-5	Aroclor-1232	18.7	U	3.70	18.7	ug/kg
53469-21-9	Aroclor-1242	18.7	U	3.70	18.7	ug/kg
12672-29-6	Aroclor-1248	18.7	U	8.70	18.7	ug/kg
11097-69-1	Aroclor-1254	18.7	U	3.00	18.7	ug/kg
37324-23-5	Aroclor-1262	18.7	U	5.00	18.7	ug/kg
11100-14-4	Aroclor-1268	18.7	U	3.80	18.7	ug/kg
11096-82-5	Aroclor-1260	174		3.20	18.7	ug/kg
SURROGATES						
877-09-8	Tetrachloro-m-xylene	24.9		32 - 144	125%	SPK: 20
2051-24-3	Decachlorobiphenyl	21.1		32 - 175	106%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Report of Analysis

Client:	JPCL Engineering	Date Collected:	03/04/25
Project:	NYSDOT Two Bronx River Parkway Bridges	Date Received:	03/04/25
Client Sample ID:	ENV-101-SB02MSD	SDG No.:	Q1486
Lab Sample ID:	Q1488-03MSD	Matrix:	SOIL
Analytical Method:	SW8082A	% Solid:	90.6
Sample Wt/Vol:	30.05 Units: g	Final Vol:	10000 uL
Soil Aliquot Vol:	uL	Test:	PCB
Extraction Type:		Injection Volume :	
GPC Factor :	1.0	PH :	
Prep Method :	SW3541B		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PP070280.D	1	03/05/25 09:10	03/05/25 19:51	PB166985

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
12674-11-2	Aroclor-1016	193		3.70	18.7	ug/kg
11104-28-2	Aroclor-1221	18.7	U	7.10	18.7	ug/kg
11141-16-5	Aroclor-1232	18.7	U	3.70	18.7	ug/kg
53469-21-9	Aroclor-1242	18.7	U	3.70	18.7	ug/kg
12672-29-6	Aroclor-1248	18.7	U	8.70	18.7	ug/kg
11097-69-1	Aroclor-1254	18.7	U	3.00	18.7	ug/kg
37324-23-5	Aroclor-1262	18.7	U	5.00	18.7	ug/kg
11100-14-4	Aroclor-1268	18.7	U	3.80	18.7	ug/kg
11096-82-5	Aroclor-1260	183		3.20	18.7	ug/kg
SURROGATES						
877-09-8	Tetrachloro-m-xylene	26.5		32 - 144	132%	SPK: 20
2051-24-3	Decachlorobiphenyl	24.1		32 - 175	121%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit