

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL030525\
 Data File : PL094499.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 05 Mar 2025 09:25
 Operator : AR\AJ
 Sample : PSTDCCC050
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 PSTDCCC050

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 06 00:51:30 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL022125.M
 Quant Title : GC Extractables
 QLast Update : Fri Feb 21 14:52:19 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.538	2.771	130.2E6	163.5E6	51.297	51.236
28)	SA Decachlor...	9.054	7.905	102.4E6	182.1E6	48.114	47.362
Target Compounds							
2)	A alpha-BHC	3.994	3.274	189.5E6	248.7E6	50.089	50.605
3)	MA gamma-BHC...	4.327	3.604	184.8E6	242.2E6	50.007	51.426
4)	MA Heptachlor	4.914	3.941	171.5E6	239.8E6	50.587	50.810
5)	MB Aldrin	5.255	4.221	172.0E6	238.6E6	50.113	51.635
6)	B beta-BHC	4.525	3.904	79211578	101.0E6	49.603	50.799
7)	B delta-BHC	4.772	4.132	180.3E6	242.7E6	50.945	51.240
8)	B Heptachlo...	5.682	4.723	156.4E6	214.8E6	51.913	50.468
9)	A Endosulfan I	6.068	5.093	137.6E6	160.6E6	50.057	40.290
10)	B gamma-Chl...	5.939	4.973	146.5E6	218.4E6	49.897	50.331
11)	B alpha-Chl...	6.017	5.037	146.2E6	215.8E6	50.460	49.993
12)	B 4,4'-DDE	6.191	5.224	136.2E6	218.6E6	52.156	52.048m
13)	MA Dieldrin	6.343	5.357	143.0E6	224.7E6	50.200	51.290
14)	MA Endrin	6.571	5.632	118.6E6	199.8E6	46.476m	58.679 #
15)	B Endosulfa...	6.793	5.927	119.8E6	198.0E6	47.611	52.950m
16)	A 4,4'-DDD	6.709	5.781	103.8E6	176.4E6	56.102	55.376
17)	MA 4,4'-DDT	7.022	6.030	109.0E6	188.7E6	51.890	53.059
18)	B Endrin al...	6.923	6.107	98259289	160.6E6	50.472	50.805
19)	B Endosulfa...	7.158	6.330	110.0E6	187.4E6	48.589	50.743
20)	A Methoxychlor	7.499	6.606	56794127	101.6E6	51.118	52.052
21)	B Endrin ke...	7.643	6.834	122.9E6	229.6E6	49.014	53.801m
22)	Mirex	8.115	7.015	97095530	179.5E6	47.058	51.481

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL030525\
Data File : PL094499.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 05 Mar 2025 09:25
Operator : AR\AJ
Sample : PSTDCCC050
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 06 00:51:30 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL022125.M
Quant Title : GC Extractables
QLast Update : Fri Feb 21 14:52:19 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm

