

Matrix Spike/Matrix Spike Duplicate Summary

SW-846

SDG No.: Q1486

Client: JPCL Engineering

Analytical Method: 8081B

DataFile : PL094517.D

Lab Sample ID:	Parameter	Spike	Sample Result	Result	Units	Rec	Rec Qual	RPD	RPD Qual	Low	Limits High	RPD
Client Sample ID:	DN-B-41MSD											
Q1485-01MSD	alpha-BHC	20	0	20.2	ug/kg	101		1		60	144	20
	beta-BHC	20	0	21.0	ug/kg	105		1		54	143	20
	delta-BHC	20	0	20.6	ug/kg	103		1		29	151	20
	gamma-BHC (Lindane)	20	0	20.1	ug/kg	101		1		61	140	20
	Heptachlor	20	0	21.0	ug/kg	105		2		63	135	20
	Aldrin	20	0	20.2	ug/kg	101		0		49	139	20
	Heptachlor epoxide	20	0	20.7	ug/kg	104		1		41	156	20
	Endosulfan I	20	0	20.9	ug/kg	104		2		56	142	20
	Dieldrin	20	0	21.2	ug/kg	106		0		47	161	20
	4,4'-DDE	20	0	21.1	ug/kg	106		1		55	136	20
	Endrin	20	0	24.4	ug/kg	122		2		57	139	20
	Endosulfan II	20	0	21.4	ug/kg	107		1		40	163	20
	4,4'-DDD	20	0	24.0	ug/kg	120		2		47	163	20
	Endosulfan sulfate	20	0	22.1	ug/kg	111		0		62	139	20
	4,4'-DDT	20	0	21.6	ug/kg	108		1		51	146	20
	Methoxychlor	20	0	23.5	ug/kg	117		1		54	136	20
	Endrin ketone	20	0	20.3	ug/kg	102		2		60	129	20
	Endrin aldehyde	20	0	20.6	ug/kg	103		3		59	132	20
	alpha-Chlordane	20	0	21.1	ug/kg	106		0		39	166	20
	gamma-Chlordane	20	0	21.2	ug/kg	106		0		44	175	20