

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: Q1486 SAS No.: Q1486 SDG NO.: Q1486

EPA Sample No.: SSTDCCC040 Date Analyzed: 03/06/2025

Lab File ID: BF141858.D Time Analyzed: 09:52

Instrument ID: BNA_F GC Column: DB-UI ID: 0.18 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	207866	6.887	779725	8.17	422269	9.92
UPPER LIMIT	415732	7.387	1559450	8.669	844538	10.422
LOWER LIMIT	103933	6.387	389863	7.669	211135	9.422
EPA SAMPLE NO.						
01 PB166984BL	218862	6.89	871720	8.16	477596	9.92
02 PB166984BS	222677	6.89	860348	8.17	467040	9.92

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.