



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 2b -

CRDL STANDARD FOR AA & ICP

Client: JPCL Engineering SDG No.: Q1487
Contract: JPCL01 Lab Code: CHEM Case No.: Q1487 SAS No.: Q1487
Initial Calibration Source: _____
Continuing Calibration Source: _____

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CRA	Mercury	0.21	0.2	103	40 - 160	CV	03/05/2025	14:08	LB134903
CRI01	Arsenic	22.6	20.0	113	40 - 160	P	03/06/2025	11:47	LB134928
	Barium	92.7	100	93	40 - 160	P	03/06/2025	11:47	LB134928
	Cadmium	5.78	6.0	96	40 - 160	P	03/06/2025	11:47	LB134928
	Chromium	10.1	10.0	101	40 - 160	P	03/06/2025	11:47	LB134928
	Lead	12.6	12.0	105	40 - 160	P	03/06/2025	11:47	LB134928
	Selenium	22.0	20.0	110	40 - 160	P	03/06/2025	11:47	LB134928
	Silver	10.2	10.0	102	40 - 160	P	03/06/2025	11:47	LB134928
CRI01	Arsenic	20.1	20.0	100	40 - 160	P	03/12/2025	11:05	LB135011
	Barium	88.8	100	89	40 - 160	P	03/12/2025	11:05	LB135011
	Cadmium	5.85	6.0	98	40 - 160	P	03/12/2025	11:05	LB135011
	Chromium	10.3	10.0	103	40 - 160	P	03/12/2025	11:05	LB135011
	Lead	10.9	12.0	91	40 - 160	P	03/12/2025	11:05	LB135011
	Selenium	18.3	20.0	92	40 - 160	P	03/12/2025	11:05	LB135011
	Silver	11.1	10.0	111	40 - 160	P	03/12/2025	11:05	LB135011
CRI01	Arsenic	21.2	20.0	106	40 - 160	P	03/13/2025	12:21	LB135035
	Barium	94.0	100	94	40 - 160	P	03/13/2025	12:21	LB135035
	Cadmium	5.90	6.0	98	40 - 160	P	03/13/2025	12:21	LB135035
	Chromium	10.4	10.0	104	40 - 160	P	03/13/2025	12:21	LB135035
	Lead	11.7	12.0	98	40 - 160	P	03/13/2025	12:21	LB135035
	Selenium	21.3	20.0	107	40 - 160	P	03/13/2025	12:21	LB135035
	Silver	10.6	10.0	106	40 - 160	P	03/13/2025	12:21	LB135035