

CALIBRATION VERIFICATION SUMMARY

Contract: JPCL01

Lab Code: CHEM Case No.: Q1487 SAS No.: Q1487 SDG NO.: Q1487

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 02/24/2025 02/24/2025

Client Sample No.: CCAL02 Date Analyzed: 03/05/2025

Lab Sample No.: AR1660CCC500 Data File : PP070274.D Time Analyzed: 18:13

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.679	5.581	5.781	519.700	500.000	3.9
Aroclor-1016-2	5.701	5.602	5.802	539.100	500.000	7.8
Aroclor-1016-3	5.763	5.665	5.865	538.430	500.000	7.7
Aroclor-1016-4	5.860	5.762	5.962	548.080	500.000	9.6
Aroclor-1016-5	6.154	6.056	6.256	519.480	500.000	3.9
Aroclor-1260-1	7.274	7.175	7.375	514.400	500.000	2.9
Aroclor-1260-2	7.527	7.429	7.629	499.540	500.000	-0.1
Aroclor-1260-3	7.886	7.788	7.988	529.380	500.000	5.9
Aroclor-1260-4	8.110	8.012	8.212	505.520	500.000	1.1
Aroclor-1260-5	8.431	8.333	8.533	525.530	500.000	5.1
Decachlorobiphenyl	10.252	10.156	10.356	50.370	50.000	0.7
Tetrachloro-m-xylene	4.526	4.427	4.627	55.610	50.000	11.2

CALIBRATION VERIFICATION SUMMARY

Contract: JPCL01

Lab Code: CHEM Case No.: Q1487 SAS No.: Q1487 SDG NO.: Q1487

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 02/24/2025 02/24/2025

Client Sample No.: CCAL02 Date Analyzed: 03/05/2025

Lab Sample No.: AR1660CCC500 Data File : PP070274.D Time Analyzed: 18:13

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.917	4.820	5.020	528.080	500.000	5.6
Aroclor-1016-2	4.936	4.839	5.039	524.790	500.000	5.0
Aroclor-1016-3	5.113	5.017	5.217	536.640	500.000	7.3
Aroclor-1016-4	5.155	5.058	5.258	534.000	500.000	6.8
Aroclor-1016-5	5.370	5.273	5.473	553.470	500.000	10.7
Aroclor-1260-1	6.408	6.312	6.512	510.260	500.000	2.1
Aroclor-1260-2	6.596	6.500	6.700	514.490	500.000	2.9
Aroclor-1260-3	6.750	6.654	6.854	485.590	500.000	-2.9
Aroclor-1260-4	7.223	7.126	7.326	541.710	500.000	8.3
Aroclor-1260-5	7.464	7.367	7.567	564.500	500.000	12.9
Decachlorobiphenyl	8.883	8.788	8.988	50.840	50.000	1.7
Tetrachloro-m-xylene	3.828	3.730	3.930	52.730	50.000	5.5