

CALIBRATION VERIFICATION SUMMARY

Contract: JPCL01

Lab Code: CHEM Case No.: Q1487 SAS No.: Q1487 SDG NO.: Q1487

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 02/24/2025 02/24/2025

Client Sample No.: CCAL03 Date Analyzed: 03/05/2025

Lab Sample No.: AR1660CCC500 Data File : PP070289.D Time Analyzed: 22:50

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	5.681	5.581	5.781	507.650	500.000	1.5
Aroclor-1016-2	5.702	5.602	5.802	514.280	500.000	2.9
Aroclor-1016-3	5.764	5.665	5.865	501.590	500.000	0.3
Aroclor-1016-4	5.862	5.762	5.962	507.010	500.000	1.4
Aroclor-1016-5	6.155	6.056	6.256	518.280	500.000	3.7
Aroclor-1260-1	7.275	7.175	7.375	532.180	500.000	6.4
Aroclor-1260-2	7.529	7.429	7.629	505.610	500.000	1.1
Aroclor-1260-3	7.888	7.788	7.988	546.130	500.000	9.2
Aroclor-1260-4	8.113	8.012	8.212	525.010	500.000	5.0
Aroclor-1260-5	8.433	8.333	8.533	541.570	500.000	8.3
Decachlorobiphenyl	10.258	10.156	10.356	51.680	50.000	3.4
Tetrachloro-m-xylene	4.527	4.427	4.627	55.950	50.000	11.9

CALIBRATION VERIFICATION SUMMARY

Contract: JPCL01

Lab Code: CHEM Case No.: Q1487 SAS No.: Q1487 SDG NO.: Q1487

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 02/24/2025 02/24/2025

Client Sample No.: CCAL03 Date Analyzed: 03/05/2025

Lab Sample No.: AR1660CCC500 Data File : PP070289.D Time Analyzed: 22:50

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.919	4.820	5.020	536.170	500.000	7.2
Aroclor-1016-2	4.938	4.839	5.039	532.730	500.000	6.5
Aroclor-1016-3	5.115	5.017	5.217	542.770	500.000	8.6
Aroclor-1016-4	5.157	5.058	5.258	529.850	500.000	6.0
Aroclor-1016-5	5.372	5.273	5.473	573.280	500.000	14.7
Aroclor-1260-1	6.410	6.312	6.512	516.300	500.000	3.3
Aroclor-1260-2	6.598	6.500	6.700	527.370	500.000	5.5
Aroclor-1260-3	6.752	6.654	6.854	498.350	500.000	-0.3
Aroclor-1260-4	7.225	7.126	7.326	567.590	500.000	13.5
Aroclor-1260-5	7.466	7.367	7.567	585.700	500.000	17.1
Decachlorobiphenyl	8.886	8.788	8.988	52.490	50.000	5.0
Tetrachloro-m-xylene	3.830	3.730	3.930	52.590	50.000	5.2