

Matrix Spike/Matrix Spike Duplicate Summary

SW-846

SDG No.: Q1487

Client: JPCL Engineering

Analytical Method: 8081B

DataFile : PL094516.D

Lab Sample ID:	Parameter	Spike	Sample Result	Result	Units	Rec	Rec Qual	RPD	RPD Qual	Low	Limits High	RPD
Client Sample ID:	DN-B-41MS											
Q1485-01MS	alpha-BHC	20.01	0	20.5	ug/kg	102				60	144	
	beta-BHC	20.01	0	21.2	ug/kg	106				54	143	
	delta-BHC	20.01	0	20.8	ug/kg	104				29	151	
	gamma-BHC (Lindane)	20.01	0	20.4	ug/kg	102				61	140	
	Heptachlor	20.01	0	21.4	ug/kg	107				63	135	
	Aldrin	20.01	0	20.3	ug/kg	101				49	139	
	Heptachlor epoxide	20.01	0	20.7	ug/kg	103				41	156	
	Endosulfan I	20.01	0	21.2	ug/kg	106				56	142	
	Dieldrin	20.01	0	21.3	ug/kg	106				47	161	
	4,4'-DDE	20.01	0	21.1	ug/kg	105				55	136	
	Endrin	20.01	0	24.9	ug/kg	124				57	139	
	Endosulfan II	20.01	0	21.6	ug/kg	108				40	163	
	4,4'-DDD	20.01	0	23.6	ug/kg	118				47	163	
	Endosulfan sulfate	20.01	0	22.2	ug/kg	111				62	139	
	4,4'-DDT	20.01	0	21.4	ug/kg	107				51	146	
	Methoxychlor	20.01	0	23.3	ug/kg	116				54	136	
	Endrin ketone	20.01	0	20.0	ug/kg	100				60	129	
	Endrin aldehyde	20.01	0	20.1	ug/kg	100				59	132	
	alpha-Chlordane	20.01	0	21.2	ug/kg	106				39	166	
	gamma-Chlordane	20.01	0	21.2	ug/kg	106				44	175	