

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q1487

Client: JPCL Engineering

Analytical Method: **8081B** Datafile : PL094501.D

Lab Sample ID	Parameter	Spike	Result	Units	Rec	RPD	Qual	RPD		Limits	
								Qual	Low	High	RPD
PB166986BS	alpha-BHC	16.66	17.1	ug/kg	103				84	123	
	beta-BHC	16.66	17.6	ug/kg	106				82	123	
	delta-BHC	16.66	17.4	ug/kg	104				83	126	
	gamma-BHC (Lindane)	16.66	16.9	ug/kg	101				83	125	
	Heptachlor	16.66	18.1	ug/kg	109				83	122	
	Aldrin	16.66	17.2	ug/kg	103				82	124	
	Heptachlor epoxide	16.66	17.9	ug/kg	107				83	120	
	Endosulfan I	16.66	18.0	ug/kg	108				81	124	
	Dieldrin	16.66	18.1	ug/kg	109				85	121	
	4,4'-DDE	16.66	19.0	ug/kg	114				81	123	
	Endrin	16.66	20.8	ug/kg	125				76	130	
	Endosulfan II	16.66	19.1	ug/kg	115				80	125	
	4,4'-DDD	16.66	20.3	ug/kg	122				80	131	
	Endosulfan sulfate	16.66	18.5	ug/kg	111				81	122	
	4,4'-DDT	16.66	19.4	ug/kg	116				70	129	
	Methoxychlor	16.66	18.6	ug/kg	112				60	119	
	Endrin ketone	16.66	19.2	ug/kg	115				77	132	
	Endrin aldehyde	16.66	17.6	ug/kg	106				79	124	
	alpha-Chlordane	16.66	18.3	ug/kg	110				84	120	
	gamma-Chlordane	16.66	18.2	ug/kg	109				83	122	