

Method Path : Z:\svoasrv\HPCHEM1\BNA_F\Methods\
 Method File : 8270-BF022725.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Fri Feb 28 01:48:16 2025
 Response Via : Initial Calibration

Calibration Files

2.5 =BF141793.D 5 =BF141794.D 10 =BF141795.D 20 =BF141796.D 40 =BF141797.D 50 =BF141798.D 60 =BF141799.D 80 =BF141800.D

	Compound	2.5	5	10	20	40	50	60	80	Avg	%RSD
1) I	1,4-Dichlorobenzen...	-----ISTD-----									
2)	1,4-Dioxane	0.613	0.554	0.576	0.596	0.545	0.554	0.541	0.568	4.81	
3)	Pyridine	1.403	1.412	1.486	1.476	1.380	1.389	1.350	1.414	3.54	
4)	n-Nitrosodimet...	0.634	0.666	0.689	0.730	0.674	0.686	0.683	0.680	4.21	
5) S	2-Fluorophenol	1.348	1.315	1.326	1.286	1.185	1.176	1.131	1.252	6.91	
6)	Aniline	1.746	1.727	1.805	1.809	1.642	1.625	1.509	1.695	6.42	
7) S	Phenol-d6	1.750	1.707	1.728	1.684	1.541	1.523	1.477	1.630	6.90	
8)	2-Chlorophenol	1.442	1.406	1.439	1.412	1.299	1.266	1.220	1.355	6.72	
9)	Benzaldehyde	1.151	1.088	1.024	0.982	0.874	0.824	0.694	0.948	16.84	
10) C	Phenol	1.854	1.823	1.866	1.819	1.641	1.625	1.564	1.742	7.26	
11)	bis(2-Chloroet...	1.377	1.376	1.392	1.339	1.279	1.265	1.265	1.328	4.26	
12)	1,3-Dichlorobe...	1.544	1.520	1.541	1.504	1.378	1.361	1.308	1.451	6.80	
13) C	1,4-Dichlorobe...	1.554	1.533	1.549	1.512	1.390	1.377	1.317	1.462	6.66	
14)	1,2-Dichlorobe...	1.493	1.444	1.476	1.401	1.286	1.270	1.186	1.365	8.63	
15)	Benzyl Alcohol	1.347	1.353	1.374	1.366	1.258	1.241	1.180	1.303	5.82	
16)	2,2'-oxybis(1-...	2.139	2.094	2.149	2.079	1.930	1.910	1.814	2.016	6.47	
17)	2-Methylphenol	1.155	1.132	1.165	1.164	1.079	1.084	1.049	1.118	4.22	
18)	Hexachloroethane	0.559	0.542	0.575	0.578	0.533	0.530	0.513	0.547	4.39	
19) P	n-Nitroso-di-n...	1.133	1.136	1.105	1.120	1.087	1.009	1.007	0.975	1.072	6.01
20)	3+4-Methylphenols	1.507	1.510	1.506	1.442	1.304	1.281	1.198	1.392	9.29	
21) I	Naphthalene-d8	-----ISTD-----									
22)	Acetophenone	0.539	0.517	0.512	0.497	0.456	0.455	0.434	0.487	8.04	
23) S	Nitrobenzene-d5	0.248	0.271	0.318	0.341	0.332	0.340	0.339	0.313	12.10	
24)	Nitrobenzene	0.273	0.303	0.346	0.363	0.349	0.357	0.354	0.335	10.07	
25)	Isophorone	0.702	0.682	0.690	0.682	0.642	0.659	0.649	0.672	3.35	
26) C	2-Nitrophenol	0.079	0.090	0.116	0.144	0.145	0.155	0.158	0.127	25.21	
27)	2,4-Dimethylph...	0.258	0.250	0.256	0.254	0.238	0.238	0.236	0.247	3.87	
28)	bis(2-Chloroet...	0.463	0.448	0.450	0.435	0.404	0.408	0.394	0.429	6.19	
29) C	2,4-Dichloroph...	0.280	0.288	0.296	0.298	0.277	0.281	0.275	0.285	3.20	
30)	1,2,4-Trichlor...	0.328	0.321	0.327	0.319	0.298	0.301	0.292	0.312	4.71	
31)	Naphthalene	1.140	1.097	1.100	1.043	0.960	0.950	0.906	1.028	8.74	
32)	Benzoic acid	0.099	0.145	0.191	0.200	0.215	0.226	0.179	0.179	26.89	
33)	4-Chloroaniline	0.402	0.392	0.398	0.379	0.350	0.356	0.360	0.377	5.66	
34) C	Hexachlorobuta...	0.199	0.193	0.202	0.199	0.188	0.191	0.187	0.194	2.98	
35)	Caprolactam	0.087	0.086	0.090	0.091	0.085	0.089	0.086	0.088	2.77	
36) C	4-Chloro-3-met...	0.333	0.316	0.329	0.326	0.300	0.310	0.303	0.317	4.16	
37)	2-Methylnaphth...	0.762	0.725	0.723	0.677	0.621	0.619	0.591	0.674	9.66	
38)	1-Methylnaphth...	0.734	0.688	0.697	0.652	0.594	0.594	0.570	0.647	9.67	

Method Path : Z:\svoasrv\HPCHEM1\BNA_F\Methods\
 Method File : 8270-BF022725.M


39) I	Acenaphthene-d10	-----ISTD-----									
40)	1,2,4,5-Tetrac...	0.609	0.600	0.602	0.589	0.549	0.553	0.540	0.578	5.03	
41) P	Hexachlorocycl...	0.217	0.230	0.252	0.258	0.246	0.247	0.236	0.241	5.87	
42) S	2,4,6-Tribromo...	0.175	0.179	0.188	0.199	0.183	0.196	0.194	0.188	4.80	
43) C	2,4,6-Trichlor...	0.355	0.368	0.381	0.398	0.366	0.368	0.377	0.373	3.65	
44)	2,4,5-Trichlor...	0.348	0.364	0.388	0.383	0.364	0.380	0.360	0.369	3.89	
45) S	2-Fluorobiphenyl	1.443	1.387	1.346	1.237	1.128	1.141	1.049	1.247	11.93	
46)	1,1'-Biphenyl	1.694	1.659	1.630	1.519	1.413	1.419	1.323	1.522	9.38	
47)	2-Chloronaphth...	1.211	1.184	1.201	1.135	1.058	1.071	1.006	1.124	7.10	
48)	2-Nitroaniline	0.152	0.187	0.246	0.300	0.297	0.319	0.323	0.260	26.00	
49)	Acenaphthylene	1.847	1.809	1.798	1.711	1.560	1.596	1.472	1.685	8.54	
50)	Dimethylphthalate	1.431	1.405	1.405	1.363	1.262	1.285	1.224	1.339	6.09	
51)	2,6-Dinitrotol...	0.136	0.184	0.231	0.257	0.251	0.265	0.261	0.226	21.50	
52) C	Acenaphthene	1.217	1.201	1.197	1.150	1.068	1.084	1.039	1.137	6.37	
53)	3-Nitroaniline	0.146	0.193	0.243	0.277	0.264	0.282	0.280	0.241	21.68	
54) P	2,4-Dinitrophenol		0.049	0.066	0.093	0.096	0.112	0.116	0.089	29.73	
55)	Dibenzofuran	1.850	1.794	1.765	1.663	1.532	1.542	1.436	1.654	9.40	
56) P	4-Nitrophenol	0.129	0.159	0.197	0.228	0.215	0.229	0.227	0.198	19.87	
57)	2,4-Dinitrotol...	0.151	0.198	0.265	0.313	0.310	0.325	0.320	0.269	25.57	
58)	Fluorene	1.433	1.367	1.313	1.228	1.128	1.134	1.072	1.239	10.97	
59)	2,3,4,6-Tetrac...	0.313	0.314	0.336	0.339	0.309	0.322	0.315	0.321	3.66	
60)	Diethylphthalate	1.429	1.413	1.428	1.383	1.259	1.275	1.212	1.343	6.78	
61)	4-Chlorophenyl...	0.691	0.665	0.654	0.620	0.572	0.583	0.562	0.621	8.09	
62)	4-Nitroaniline	0.149	0.187	0.229	0.267	0.248	0.265	0.256	0.229	19.54	
63)	Azobenzene	1.544	1.503	1.505	1.437	1.320	1.325	1.262	1.414	7.81	
64) I	Phenanthrene-d10	-----ISTD-----									
65)	4,6-Dinitro-2-...		0.043	0.063	0.087	0.090	0.100	0.103	0.081	28.86	
66) c	n-Nitrosodiphe...	0.702	0.681	0.689	0.665	0.631	0.623	0.607	0.657	5.56	
67)	4-Bromophenyl-...	0.241	0.237	0.239	0.238	0.227	0.230	0.224	0.234	2.86	
68)	Hexachlorobenzene	0.253	0.249	0.255	0.251	0.240	0.242	0.237	0.247	2.79	
69)	Atrazine	0.213	0.208	0.196	0.179	0.149	0.139		0.181	17.11	
70) C	Pentachlorophenol	0.137	0.146	0.161	0.169	0.159	0.164	0.160	0.156	7.07	
71)	Phenanthrene	1.186	1.152	1.134	1.073	1.001	1.001	0.940	1.070	8.58	
72)	Anthracene	1.203	1.156	1.155	1.068	1.002	0.990	0.931	1.072	9.54	
73)	Carbazole	1.074	1.010	1.014	0.955	0.882	0.871	0.821	0.947	9.69	
74)	Di-n-butylphth...	1.335	1.291	1.291	1.208	1.128	1.114	1.035	1.200	9.30	
75) C	Fluoranthene	1.285	1.216	1.192	1.089	0.992	0.981	0.917	1.096	12.66	
76) I	Chrysene-d12	-----ISTD-----									
77)	Benzidine		0.342	0.198	0.259	0.214	0.181	0.316	0.252	26.09	
78)	Pyrene	1.698	1.687	1.815	1.796	1.738	1.734	1.656	1.732	3.32	
79) S	Terphenyl-d14	1.250	1.233	1.297	1.259	1.230	1.227	1.157	1.236	3.43	
80)	Butylbenzylpht...	0.546	0.579	0.643	0.668	0.650	0.667	0.652	0.629	7.54	
81)	Benzo(a)anthra...	1.367	1.387	1.384	1.312	1.244	1.262	1.292	1.321	4.46	
82)	3,3'-Dichlorob...	0.406	0.380	0.382	0.393	0.364	0.364	0.365	0.379	4.27	
83)	Chrysene	1.303	1.184	1.235	1.265	1.185	1.209	1.144	1.218	4.42	
84)	Bis(2-ethylhex...	0.702	0.696	0.788	0.832	0.811	0.828	0.827	0.783	7.59	
85) c	Di-n-octyl pht...	0.903	0.922	1.109	1.307	1.304	1.318	1.348	1.173	16.59	

Method Path : Z:\svoasrv\HPCHEM1\BNA_F\Methods\
Method File : 8270-BF022725.M

86) I	Perylene-d12	-----ISTD-----									
87)	Indeno(1,2,3-c...	1.097	1.126	1.240	1.346	1.337	1.371	1.397	1.274	9.52	
88)	Benzo(b)fluora...	1.442	1.396	1.533	1.306	1.232	1.370	1.256	1.362	7.81	
89)	Benzo(k)fluora...	1.274	1.240	1.098	1.214	1.165	1.040	1.074	1.158	7.70	
90) C	Benzo(a)pyrene	1.113	1.083	1.133	1.119	1.060	1.080	1.060	1.093	2.68	
91)	Dibenzo(a,h)an...	0.902	0.931	1.025	1.109	1.094	1.101	1.115	1.040	8.63	
92)	Benzo(g,h,i)pe...	0.898	0.929	1.034	1.130	1.129	1.156	1.179	1.065	10.60	

(#) = Out of Range


2-Nitrophenol



Response = $1.654 \times 10^{-1} \times \text{Amt} - 3.762 \times 10^{-2}$
 Coef of Det (r^2) = 0.997727 Curve Fit: Linear
 Method Name: Z:\svoasrv\HPCHEM1\BNA F\Methods\8270-BF022725.M
 Calibration Table Last Updated: Fri Feb 28 01:48:16 2025

Benzoic acid

Response Ratio



$$\text{Response} = 2.456\text{e-}001 * \text{Amt} - 9.444\text{e-}002$$


Coef of Det (r^2) = 0.997308 Curve Fit: Linear

Method Name: Z:\svoasrv\HPCHEM1\BNA F\Methods\8270-BF022725.M

Calibration Table Last Updated: Fri Feb 28 01:48:16 2025

2-Nitroaniline

Response Ratio



$$\text{Response} = 3.381\text{e-}001 * \text{Amt} - 7.336\text{e-}002$$


Coef of Det (r^2) = 0.998178 Curve Fit: Linear

Method Name: Z:\svoasrv\HPCHEM1\BNA F\Methods\8270-BF022725.M

Calibration Table Last Updated: Fri Feb 28 01:48:16 2025

2,6-Dinitrotoluene

Response Ratio



$$\text{Response} = 2.723\text{e-}001 * \text{Amt} - 3.850\text{e-}002$$


Coef of Det (r^2) = 0.999171 Curve Fit: Linear

Method Name: Z:\svoasrv\HPCHEM1\BNA F\Methods\8270-BF022725.M

Calibration Table Last Updated: Fri Feb 28 01:48:16 2025

3-Nitroaniline

Response Ratio




$$\text{Response} = 2.909\text{e-}001 * \text{Amt} - 4.315\text{e-}002$$

Coef of Det (r^2) = 0.998769 Curve Fit: Linear

Method Name: Z:\svoasrv\HPCHEM1\BNA F\Methods\8270-BF022725.M

Calibration Table Last Updated: Fri Feb 28 01:48:16 2025


2,4-Dinitrophenol



$R = 8.979e-003 A^2 + 8.829e-002 A - 2.630e-002$
 Coef of Det (r^2) = 0.996714 Curve Fit: Quadratic
 Method Name: Z:\svoasrv\HPCHEM1\BNA F\Methods\8270-BF022725.M
 Calibration Table Last Updated: Fri Feb 28 01:48:16 2025

2,4-Dinitrotoluene

Response Ratio




$$\text{Response} = 3.378\text{e-}001 * \text{Amt} - 5.953\text{e-}002$$

Coef of Det (r^2) = 0.999067 Curve Fit: Linear

Method Name: Z:\svoasrv\HPCHEM1\BNA F\Methods\8270-BF022725.M

Calibration Table Last Updated: Fri Feb 28 01:48:16 2025

4,6-Dinitro-2-methylphenol



Response = $1.134 \times 10^{-1} \times \text{Amt} - 4.606 \times 10^{-2}$
Coef of Det (r^2) = 0.996201 Curve Fit: Linear
Method Name: Z:\svoasrv\HPCHEM1\BNA F\Methods\8270-BF022725.M
Calibration Table Last Updated: Fri Feb 28 01:48:16 2025