

Report of Analysis

Client:	JPCL Engineering	Date Collected:	
Project:	NYSDOT Two Bronx River Parkway Bridges	Date Received:	
Client Sample ID:	PB166984BS	SDG No.:	Q1487
Lab Sample ID:	PB166984BS	Matrix:	SOIL
Analytical Method:	SW8270	% Solid:	100
Sample Wt/Vol:	30.03 Units: g	Final Vol:	1000 uL
Soil Aliquot Vol:	uL	Test:	SVOC-TCL BNA -20
Extraction Type :	Decanted : N	Level :	LOW
Injection Volume :	GPC Factor : 1.0	GPC Cleanup :	N PH :
Prep Method :	SW3541		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BF141860.D	1	03/05/25 09:40	03/06/25 10:52	PB166984

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
100-52-7	Benzaldehyde	710		180	330	ug/Kg
108-95-2	Phenol	1300		82.8	170	ug/Kg
111-44-4	bis(2-Chloroethyl)ether	1300		83.6	170	ug/Kg
95-57-8	2-Chlorophenol	1400		83.4	170	ug/Kg
95-48-7	2-Methylphenol	1300		80.5	170	ug/Kg
108-60-1	2,2-oxybis(1-Chloropropane)	1100		90.8	170	ug/Kg
98-86-2	Acetophenone	1500		86.8	170	ug/Kg
65794-96-9	3+4-Methylphenols	1300		79.7	330	ug/Kg
621-64-7	n-Nitroso-di-n-propylamine	1200		40.3	79.9	ug/Kg
67-72-1	Hexachloroethane	1400		82.9	170	ug/Kg
98-95-3	Nitrobenzene	1500		90.7	170	ug/Kg
78-59-1	Isophorone	1400		84.5	170	ug/Kg
88-75-5	2-Nitrophenol	1600		94.4	170	ug/Kg
105-67-9	2,4-Dimethylphenol	1700		93.1	170	ug/Kg
111-91-1	bis(2-Chloroethoxy)methane	1300		85.7	170	ug/Kg
120-83-2	2,4-Dichlorophenol	1400		75.4	170	ug/Kg
91-20-3	Naphthalene	1400		82.5	170	ug/Kg
106-47-8	4-Chloroaniline	400		82.5	170	ug/Kg
87-68-3	Hexachlorobutadiene	1400		83.2	170	ug/Kg
105-60-2	Caprolactam	1600		86.7	330	ug/Kg
59-50-7	4-Chloro-3-methylphenol	1400		77.4	170	ug/Kg
91-57-6	2-Methylnaphthalene	1300		82.4	170	ug/Kg
77-47-4	Hexachlorocyclopentadiene	4900	E	160	330	ug/Kg
88-06-2	2,4,6-Trichlorophenol	1500		71.3	170	ug/Kg
95-95-4	2,4,5-Trichlorophenol	1500		73.9	170	ug/Kg
92-52-4	1,1-Biphenyl	1500		87.3	170	ug/Kg
91-58-7	2-Chloronaphthalene	1400		83.2	170	ug/Kg
88-74-4	2-Nitroaniline	1600		94.9	170	ug/Kg
131-11-3	Dimethylphthalate	1400		81.6	170	ug/Kg

Report of Analysis

Client:	JPCL Engineering	Date Collected:	
Project:	NYSDOT Two Bronx River Parkway Bridges	Date Received:	
Client Sample ID:	PB166984BS	SDG No.:	Q1487
Lab Sample ID:	PB166984BS	Matrix:	SOIL
Analytical Method:	SW8270	% Solid:	100
Sample Wt/Vol:	30.03 Units: g	Final Vol:	1000 uL
Soil Aliquot Vol:	uL	Test:	SVOC-TCL BNA -20
Extraction Type :	Decanted : N	Level :	LOW
Injection Volume :	GPC Factor : 1.0	GPC Cleanup :	N PH :
Prep Method :	SW3541		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BF141860.D	1	03/05/25 09:40	03/06/25 10:52	PB166984

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
208-96-8	Acenaphthylene	1500		86.4	170	ug/Kg
606-20-2	2,6-Dinitrotoluene	1600		83.1	170	ug/Kg
99-09-2	3-Nitroaniline	850		89.1	170	ug/Kg
83-32-9	Acenaphthene	1600		81.0	170	ug/Kg
51-28-5	2,4-Dinitrophenol	3700	E	240	330	ug/Kg
100-02-7	4-Nitrophenol	3800	E	120	330	ug/Kg
132-64-9	Dibenzofuran	1300		84.3	170	ug/Kg
121-14-2	2,4-Dinitrotoluene	1800		86.1	170	ug/Kg
84-66-2	Diethylphthalate	1400		80.0	170	ug/Kg
7005-72-3	4-Chlorophenyl-phenylether	1400		85.5	170	ug/Kg
86-73-7	Fluorene	1400		85.4	170	ug/Kg
100-01-6	4-Nitroaniline	1800		110	170	ug/Kg
534-52-1	4,6-Dinitro-2-methylphenol	1900		120	330	ug/Kg
86-30-6	n-Nitrosodiphenylamine	1400		81.5	170	ug/Kg
101-55-3	4-Bromophenyl-phenylether	1400		78.8	170	ug/Kg
118-74-1	Hexachlorobenzene	1500		84.9	170	ug/Kg
1912-24-9	Atrazine	2000		91.3	170	ug/Kg
87-86-5	Pentachlorophenol	3000	E	77.2	330	ug/Kg
85-01-8	Phenanthrene	1400		83.9	170	ug/Kg
120-12-7	Anthracene	1400		84.3	170	ug/Kg
86-74-8	Carbazole	1400		80.2	170	ug/Kg
84-74-2	Di-n-butylphthalate	1400		84.2	170	ug/Kg
206-44-0	Fluoranthene	1500		81.6	170	ug/Kg
129-00-0	Pyrene	1400		82.9	170	ug/Kg
85-68-7	Butylbenzylphthalate	1400		96.7	170	ug/Kg
91-94-1	3,3-Dichlorobenzidine	720		98.5	330	ug/Kg
56-55-3	Benzo(a)anthracene	1400		80.6	170	ug/Kg
218-01-9	Chrysene	1400		79.4	170	ug/Kg
117-81-7	Bis(2-ethylhexyl)phthalate	1500		90.9	170	ug/Kg
117-84-0	Di-n-octyl phthalate	1400		110	330	ug/Kg
205-99-2	Benzo(b)fluoranthene	1300		81.0	170	ug/Kg

Report of Analysis

Client:	JPCL Engineering	Date Collected:	
Project:	NYSDOT Two Bronx River Parkway Bridges	Date Received:	
Client Sample ID:	PB166984BS	SDG No.:	Q1487
Lab Sample ID:	PB166984BS	Matrix:	SOIL
Analytical Method:	SW8270	% Solid:	100
Sample Wt/Vol:	30.03 Units: g	Final Vol:	1000 uL
Soil Aliquot Vol:	uL	Test:	SVOC-TCL BNA -20
Extraction Type :	Decanted : N	Level :	LOW
Injection Volume :	GPC Factor : 1.0	GPC Cleanup :	N PH :
Prep Method :	SW3541		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BF141860.D	1	03/05/25 09:40	03/06/25 10:52	PB166984

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
207-08-9	Benzo(k)fluoranthene	1300		82.5	170	ug/Kg
50-32-8	Benzo(a)pyrene	1500		92.9	170	ug/Kg
193-39-5	Indeno(1,2,3-cd)pyrene	1600		78.0	170	ug/Kg
53-70-3	Dibenzo(a,h)anthracene	1500		81.1	170	ug/Kg
191-24-2	Benzo(g,h,i)perylene	1500		80.0	170	ug/Kg
95-94-3	1,2,4,5-Tetrachlorobenzene	1500		86.7	170	ug/Kg
123-91-1	1,4-Dioxane	1200		110	170	ug/Kg
58-90-2	2,3,4,6-Tetrachlorophenol	1500		74.6	170	ug/Kg
SURROGATES						
367-12-4	2-Fluorophenol	122		18 - 112	81%	SPK: 150
13127-88-3	Phenol-d6	116		15 - 107	77%	SPK: 150
4165-60-0	Nitrobenzene-d5	105		18 - 107	105%	SPK: 100
321-60-8	2-Fluorobiphenyl	91.3		20 - 109	91%	SPK: 100
118-79-6	2,4,6-Tribromophenol	166		10 - 116	110%	SPK: 150
1718-51-0	Terphenyl-d14	94.9		10 - 105	95%	SPK: 100
INTERNAL STANDARDS						
3855-82-1	1,4-Dichlorobenzene-d4	223000	6.887			
1146-65-2	Naphthalene-d8	860000	8.169			
15067-26-2	Acenaphthene-d10	467000	9.922			
1517-22-2	Phenanthrene-d10	802000	11.41			
1719-03-5	Chrysene-d12	537000	14.045			
1520-96-3	Perylene-d12	470000	15.515			

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products